



Houston Forensic Science Center
Comparative and Analytical Division - Toxicology

Reagent and Quality Control (Volatiles) Preparation Log

Reagent/Quality Control Name		10% Ethanol Control (v/v)		Lab Lot #		180820-Q-10	
Pipette(s)	0413	Balance		Analytical (SN:1123241782) Top Loading (SN:1123253484)		N/A	
Prepared By	CLR	Preparation Date	08/20/18	Expiration Date	08/20/19		
Total Volume	10mL	Verification Batch Name		ALC-20180822-AAJ 08/20/18			
Chemical 1	Ethyl alcohol	Manufacturer and Lot #	Acros B0523296	Expiration Date	N/A	Amount Added	1000ul
Chemical 2	DI water	Manufacturer and Lot #	Millipore N/A	Expiration Date	N/A	Amount Added	~9ml
Chemical 3		Manufacturer and Lot #		Expiration Date		Amount Added	evc 08/20/18
Chemical 4		Manufacturer and Lot #		Expiration Date		Amount Added	
Chemical 5		Manufacturer and Lot #		Expiration Date		Amount Added	
Comments:	Added 1000ul ethyl alcohol to DI water in a 10mL volumetric flask and QS. evc 08/20/18						