



# Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

## Reagent and Quality Control (Volatiles) Preparation Log

Reagent/Quality Control Name		0.01% n-propanol internal standard working solution (v/v)		Lab Lot #	201222-15B		
Pipette(s)	8932		Balance	Analytical (SN:1123241782) Top Loading (SN:1123253484) (N/A)			
Prepared By	HM	Preparation Date	12/22/2020		Expiration Date	06/22/2021	
Total Volume	2L	Verification Batch Name		ALC_20210105_EP -EP 01/11/2021			
Chemical 1	n-propyl alcohol	Manufacturer and Lot #	EMD Millipore 57335	Expiration Date	N/A	Amount Added	200 µL
Chemical 2	DI Water	Manufacturer and Lot #	Millipore N/A	Expiration Date	N/A	Amount Added	2L
Chemical 3		Manufacturer and Lot #		Expiration Date		Amount Added	
Chemical 4		Manufacturer and Lot #		Expiration Date		Amount Added	
Chemical 5		Manufacturer and Lot #		Expiration Date		Amount Added	-HM 12/22/2020
Comments:	Added <del>20 µL</del> 200 µL of n-propyl alcohol to DI water in a 2L volumetric flask. -HM 12/22/2020						