



**Houston Forensic Science Center**  
Comparative and Analytical Division - Toxicology

**Reagent and Quality Control (Volatiles) Preparation Log**

Reagent/Quality Control Name		0.01% n-propanol internal standard working solution (v/v)		Lab Lot #	210423- ISB		
Pipette(s)	8932		Balance	Analytical (SN:1123241782) Top Loading (SN:1123253484) <span style="border: 1px solid black; border-radius: 50%; padding: 2px;">N/A</span>			
Prepared By	KMY	Preparation Date	4/23/2021		Expiration Date	10/23/2021	
Total Volume	2L	Verification Batch Name		HLC-20210423-AAJ 004 4/23/2021			
Chemical 1	n-propyl alcohol	Manufacturer and Lot #	EMD Millipore 57335	Expiration Date	N/A	Amount Added	200uL
Chemical 2	DI water	Manufacturer and Lot #	Millipore N/A	Expiration Date	N/A	Amount Added	2L
Chemical 3		Manufacturer and Lot #		Expiration Date		Amount Added	
Chemical 4		Manufacturer and Lot #	KMY 4/23/2021	Expiration Date		Amount Added	
Chemical 5		Manufacturer and Lot #		Expiration Date		Amount Added	
Comments:	Added 200uL of n-propyl alcohol to DI water in a 2L volumetric flask -KMY 4/23/2021						