

Meyer Instruments
1304 Langham Creek Dr. Suite 235
Houston, TX 77084

MICROSCOPE SERVICE RECORD

CLIENT INFORMATION

Client Name: Houston Forensic Science Center
Street Address: 1200 Travis Street
City/State/Zip: Houston, Texas 77002
Primary Contact: Robin Guidry Phone/Fax: (713) 308-2620

MICROSCOPE DATA

Instrument Type: Comparison Scope Unit ID No: N/A
Mfr/Model: Reichert / K2700F Serial No: 109
Location/Room: Firearms Date of Service: April -2015
Previous Service: 2014 Next Service Due: April - 2016

MICROSCOPE SERVICES

The following services are performed as applicable for each instrument:

- Instrument is disassembled (field serviceable items) and all components inspected.
- All components cleaned, lubricated as needed and reassembled.
- Minor repairs and adjustments performed to factory standards.
- Illumination pathway(s) inspected and aligned.
- Optical pathway(s) inspected and aligned.
- Detailed list provided for any repairs as needed.

SPECIAL COMMENTS: _____

Service Technician: *Bob King*
Bob King

Meyer Instruments
1304 Langham Creek Dr. Suite 235
Houston, TX 77084



2500 Bayport Boulevard, Seabrook, Texas 77586, Phone: (281) 474-3329, Fax: (281) 474-5482

