

**To:** Callan Hundl

Houston Forensic Science Center  
Quality Associate  
1301 Fannin St, Suite 170  
Houston, Texas 77002

**Date:** May 5, 2016

**Phone:** 713-929-6748


**Fax:**

**E-Mail:** [chundl@houstonforensicscience.org](mailto:chundl@houstonforensicscience.org)

---

### **Microscope Certifications, Firearms Section**

I hereby certify the Leica K2700, # 4 comparison microscope I evaluated and cleaned, in the possession of the Houston Forensic Science Center, 1301 Fannin St., Houston, TX 77002, is technically sound and mechanically and optically correct. It has all parts in working order and can be adjusted and focused without effort. *However, the left hand stage has come loose and needs repair from a certified technician.*



Kirk Meyer  
Regional Sales  
Phone = 281-579-0342  
E-Mail =kvm@meyerinst.com

**Microscope Certification, Firearms**