

Meyer Instruments  
1304 Langham Creek Dr. Suite 235  
Houston, TX 77084

### MICROSCOPE SERVICE RECORD

#### CLIENT INFORMATION

Client Name: Houston Forensic Science Center  
Street Address: 1200 Travis Street  
City/State/Zip: Houston, Texas 77002  
Primary Contact: Robin Guidry Phone/Fax: (713) 308-2620

#### MICROSCOPE DATA

Instrument Type: Comparison Scope Unit ID No: N/A  
Mfr/Model: Leica / FS C Serial No: 353371  
Location/Room: Firearms Date of Service: April -2015  
Previous Service: 2014 Next Service Due: April - 2016

#### MICROSCOPE SERVICES

The following services are performed as applicable for each instrument:

- Instrument is disassembled (field serviceable items) and all components inspected.
- All components cleaned, lubricated as needed and reassembled.
- Minor repairs and adjustments performed to factory standards.
- Illumination pathway(s) inspected and aligned.
- Optical pathway(s) inspected and aligned.
- Detailed list provided for any repairs as needed.

SPECIAL COMMENTS: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Service Technician: *Bob King*  
Bob King



Meyer Instruments  
1304 Langham Creek Dr. Suite 235  
Houston, TX 77084

*2500 Bayport Boulevard, Seabrook, Texas 77586, Phone: (281) 474-3329, Fax: (281) 474-5482*