



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

GC/MS Maintenance Log

Instrument: GCMS-2

Date	He Tank Pressure	H ₂ Tank Pressure*	Air Tank Pressure*	Auto Tune	Septum Change	Wash Vials Filled	Syringe Washed	Rough Pump Oil Level Checked	Liner Change	Gold Seal Replaced	Column Cut	Output*	Bead Voltage*	Comments	Signature
9/25/18	70 1900	X	X	✓	X	✓	✓	✓	X	X	X	n/a	n/a		TS
9/28/18															

*For GC/MS-NPD instrument (GCM5-5).

Signature: Melvin Henry

Date Completed: 9/28/18

