



Houston Forensic Science Center
Comparative and Analytical Division - Toxicology

GC/MS Maintenance Log

Instrument: GCMS-2

Date	He Tank Pressure	H ₂ Tank Pressure*	Air Tank Pressure*	Auto Tune	Septum Change	Wash Vials Filled	Syringe Washed	Rough Pump Oil Level Checked	Liner Change	Gold Seal Replaced	Column Cut	Output*	Bead Voltage*	Comments	Signature
11/13/18	70 700	n/a	n/a	✓	X	✓	✓	✓	✓	X	X	n/a	n/a		mh
11/15/18	70 2500	n/a	n/a	✓	X	✓	✓	✓	✓	X	X	n/a	n/a		TS
11/16/18 mh															

*For GC/MS-NPD instrument (GCMS-5).

Signature: Melina Henry

Date Completed: 11/16/18

