



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

GC/MS Maintenance Log

Instrument: GCMS -3

Date	He Tank Pressure	Air Tank Pressure*	H ₂ Water Level*	Auto Tune	Septum Change	Wash Vials Filled	Syringe Washed	Rough Pump Oil Level Checked	Liner Change	Gold Seal Replaced	Column Cut	Output*	Comments	Analyst
			H ₂ Pressure*									Bead Voltage*		
8/25/2020								VC 8/25/2020					Shutting down in preparation for hurricane. VC 8/25/2020	VC
8/27/2020								8/27/2020 mh					turned back on. foreline pressure too high. FSA called!	mh
								8/28/2020 mh						

*For GC/MS-NPD instrument (GCMS-5).

Signature: Melina Henry

Date Completed: 8/28/2020