



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

GC/MS Maintenance Log

Instrument: GCMS-5

Date	He Tank Pressure	Air Tank Pressure*	H ₂ Water Level*	Auto Tune	Septum Change	Wash Vials Filled	Syringe Washed	Rough Pump Oil Level Checked	Liner Change	Gold Seal Replaced	Column Cut	Output*	Comments	Analyst
			H ₂ Pressure*									Bead Voltage*		
9/2/2020	80 1300	90 500	ybs 80	✓	X	✓	✓	✓	X	X	X	0.4 0.673		mh
9/3/2020	80 1300	90 400	ybs 80	✓	X	✓	✓	✓	X	X	X	0.4 0.673		mh
9/3/2020 mh														

*For GC/MS-NPD instrument (GCMS-5).

Signature: Melina Henry

Date Completed: 9/3/2020

