



Houston Forensic Science Center
Comparative and Analytical Division - Toxicology

Headspace GC Maintenance Log

Instrument: Headspace 2

Date	N ₂ Tank Pressure	H ₂ Pressure	Air Tank Pressure	He Tank Pressure	Air Blank	SS	Comments	Initials
		H ₂ Water Level			Pass/Fail	Pass/Fail		
09/07/2022	70 1200	70 Yes	70 750	70 800	pass	pass		EP
09/08/2022	70 1150	70 yes	70 600	70 750	pass	pass		HAI

BAM
9/12/2022

Signature: Bruce Manduhall

Date Completed: 9/12/2022