



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Headspace GC Maintenance Log

Instrument: Headspace 2

Date	N ₂ Tank Pressure	H ₂ Pressure		Air Tank Pressure	He Tank Pressure	Air Blank	SS	Comments	Initials
		H ₂ Water Level	Pass/Fail			Pass/Fail			
10/24/22	70 1900	70 yes	70 500	70 400	Pass	Pass		JR	
10/25/22	70 1900	70 yes	70 450	70 400	Pass	Pass	Agilent Stress Test	DMR	
10/26/22	70 1900	70 yes	70 2100	70 2000	PASS	PASS	REPLACED AIR & HE CYLINDERS.	DM	
10/28/2022	70 1800	70 yes	70 2050	70 1850	Pass	Pass	Agilent stress Test	BAM	
							BAM 10/31/2022		

Signature:

Brooke Mendenhall

Date Completed:

10/31/2022

LAB-039

Issued by: Manager - Toxicology

Issue Date: 2022-08-08

BAM
10/31/2022