



# Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

## Headspace GC Maintenance Log

Instrument: Headspace 2

Date	N <sub>2</sub> Tank Pressure	H <sub>2</sub> Pressure		Air Tank Pressure	He Tank Pressure	Air Blank	SS	Comments	Initials
		H <sub>2</sub> Water Level	Pass/Fail			Pass/Fail			
12/13/22	70 400	80 yes	70 <del>70</del> TT 1500	70 1300	<del>pass</del>	<del>pass</del>	Stress test on Agilent PC	MR	
12/14/2022	70 2500	70 yes	70 1300	70 1250	<del>pass</del>	<del>pass</del>	Changed N <sub>2</sub> cylinder Agilent stress test on Test PC - MR 12/14/22	HM	
12/15/2022	70 2450	70 yes	70 1100	70 1200	pass	pass		HM	
<del>BA m 12/19/2022</del>									

Signature: Brooke Mandinkell

Date Completed: 12/19/2022