



Houston Forensic Science Center
Comparative and Analytical Division - Toxicology

Headspace GC Maintenance Log

Instrument: Headspace 3

Date	N ₂ Tank Pressure	H ₂ Pressure		Air Tank Pressure	He Tank Pressure	Air Blank	SS	Comments	Initials
		H ₂ Water Level	Pass/Fail			Pass/Fail			
11/14/2022	70 900	70 YES	70 900	70 2000	BAM	11/14/2022	Replaced the Helium gas cylinder	BAM	
11/15/2022	70 800	70 YES	70 900	70 1950	PASS	PASS		EC	
11/16/2022	70 800	70 YES	70 500	70 1900	PASS	PASS		BP	
11/17/2022	70 800	70 YES	70 2200	70 1900	PASS	PASS	changed air tank	ROY	

			BAM						
			11/21/2022						

Signature: Brooke Mendenhall

Date Completed: 11/21/2022