



# Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

## Headspace GC Maintenance Log

Instrument: Headspace 3

Date	N <sub>2</sub> Tank Pressure	H <sub>2</sub> Pressure	Air Tank Pressure	He Tank Pressure	Air Blank	SS	Comments	Initials
		H <sub>2</sub> Water Level			Pass/Fail	Pass/Fail		
12/19/2022	70 1850	70 yes	70 2275	70 550	<del>PASS 12/19/2022</del>		replaced air cylinder	BAM
12/20/2022	70 1700	70 yes	70 2250	70 500	PASS	PASS		EP
12/21/2022	70 1700	70 yes	70 2100	70 475	PASS	PASS		
								EP 12/27/2022

Signature:

Date Completed: 12/27/2022

LAB-039  
 Issued by: Manager - Toxicology  
 Issue Date: 2022-08-08