



Houston Forensic Science Center  
Comparative and Analytical Division - Toxicology

Headspace GC Maintenance Log

Instrument: Headspace 3

Date	N <sub>2</sub> Tank Pressure	H <sub>2</sub> Pressure	Air Tank Pressure	He Tank Pressure	Air Blank	SS	Comments	Initials
		H <sub>2</sub> Water Level			Pass/Fail	Pass/Fail		
1/9/23	70 1100	70 yes	70 900	70 1700	pass	pass		JWR
1/10/2023	70 1100	70 yes	70 750	70 1650	pass	pass		EP
1/11/2023	70 1100	70 yes	70 600	70 1600	pass	pass		PM
EP 01/17/2023								

Signature: \_\_\_\_\_

Date Completed: \_\_\_\_\_

01/17/2023