



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Headspace GC Maintenance Log

Instrument: Headspace 3

Date	N ₂ Tank Pressure	H ₂ Pressure		Air Tank Pressure	He Tank Pressure	Air Blank	SS	Comments	Initials
		H ₂ Water Level	Pass/Fail			Pass/Fail			
1/17/2023	70 900	70 YES	70 2200	70 1350	PASS	PASS	changed air tank -EP 01/17/2023	DM	
01/18/2023	70 850	70 yes	70 2050	70 1350	pass	pass	R&D	HM	
01/19/2023	70 850	70 yes	70 1950	70 1300	pass	pass		HM	

Signature: Brooke Mendenhall

Date Completed: 1/23/2023