

Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Tecan Maintenance Log

Freedom EVO 75/Hydroflex Plate Washer


Instrument: Tecan-2 (SN 1010008386)

Date	Rinse and Fill Wash Bottles	Prime Tecan & Plate Washer	Empty Waste Containers	Shutdown System	Monthly Maintenance		Comments	Initials
					Acid/Base	Hydroflex		
8/29/22	✓	✓	✓	✓	✓	✓	Urine verification reanalysis + Blood Standards comparison run	JH
8/30/22	✓	✓	✓	✓	✓	✓	re-injection for urine verification	JH
8/30/22	✓	✓	✓	✓	✓	✓	monthly maintenance - JP 8/30/22	JP
8/31/22	✓	✓	✓	✓	✓	✓	Blood comparison urine comparison - AM 8/31/22	AM/OM
09/01/2022	✓	✓	✓	✓	✓	✓	URINE REINJECT - AM 9/1/22	AM/OM

Signature:

Date Completed:

9/2/2022



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Tecan Maintenance Log

Freedom EVO 75/Hydroflex Plate Washer

Instrument: Tecan-2 (SN 1010008386)

Date	Rinse and Fill Wash Bottles	Prime Tecan & Plate Washer	Empty Waste Containers	Shutdown System	Monthly Maintenance		Comments	Initials
					Acid/Base	Hydroflex		
09/02/2022	✓	✓	✓	✓	 	 		YHM

008912/2022

Signature:

Date Completed:

9/2/2022