



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology


Pipette Performance Check

Pipette Number: 1742

Balance Number: Analytical: 1123241782 Top Loading: 1123253484

Volume Testing (µL)	Mass 1 (g)	Mass 2 (g)	Mass 3 (g)	Mass 4 (g)	Mass 5 (g)	Average Mass (g)	Average Volume (µL)
1000 µL	0.9975 g	0.9981 g	0.9984 g	0.9979 g	0.9978 g	0.9979 g 0.9978 g gv 08/29/2023	997.94 µL 997.90 µL gv 08/29/2023
100 µL	0.0989 g	0.0993 g	0.0992 g	0.0988 g 0.4 gv 08/29/2023	0.0982 g	0.0988 g 0.0989 g gv 08/31/2023	98.88 µL 98.90 µL gv 08/31/2023
gv 08/29/2023							

Passed Failed

Signature:  Date: 08/29/2023

Comments: Performance Check following return from external clean and lube service - gv 08/29/2023
