



# Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

## Pipette Performance Check


Pipette Number: 1742

Balance Number: Analytical: ~~1123241782~~ <sup>B801240154</sup> Top Loading: 1123253484

Volume Testing (µL)	Mass 1 (g)	Mass 2 (g)	Mass 3 (g)	Mass 4 (g)	Mass 5 (g)	Average Mass (g)	Average Volume (µL)
1000 µL	0.998 g	0.998 g	0.999 g	0.999 g	1.000 g	0.9988 g	998.80 µL
100 µL	0.098 g	0.099 g	0.098 g	0.100 g	0.100 g	0.0990 g	99.00 µL
					gV 05/30/2024		

Passed

Failed

Signature:  Date: 05/30/2024

Comments: Performance check performed following the return from external calibration and PM service. The balance (S/N: B801240154) only measured up to three decimal places. Refer to Performance Check - Balance - 2024 - 05 - 30.

Hamilton 1742 is now in service and can be used in casework - gV 5/30/2024