



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Pipette Performance Check

Pipette Number:

7903

Balance Number:

Analytical: 1123241782

Top Loading: 1123253484

Volume Testing (µL)	Mass 1 (g)	Mass 2 (g)	Mass 3 (g)	Mass 4 (g)	Mass 5 (g)	Average Mass (g)	Average Volume (µL)
100µL	0.0997g	0.0999g	0.0997g	0.0998g	0.0997g	0.09976g	99.76µL
1000µL	1.0006g	1.0008g	1.0009g	1.0009g	1.0008g	1.00080g	1000.80µL
			BAM 11/23/2022				

Passed

Failed

Signature:

Brooke Mendelhall

Date:

11/23/2022

Comments: Performance check performed following the return from external calibration & PM service. Hamilton 7903 is back in-service & can be used in casework. -BAM 11/23/2022