



Houston Forensic Science Center  
Comparative and Analytical Division - Toxicology

Pipette Performance Check

Pipette Number: 7903  
 Balance Number: Analytical: 1123241782 Top Loading: 1123253484

Volume Testing (µL)	Mass 1 (g)	Mass 2 (g)	Mass 3 (g)	Mass 4 (g)	Mass 5 (g)	Average Mass (g)	Average Volume (µL)
1000 µL	0.9970g	0.9974g	0.9975g	0.9978g	0.9978g	0.9975g	997.5 µL gv 10/14/2023
gv 10/31/2023							

Passed  Failed

Signature: [Signature] Date: 10/31/2023

Comments: Performing performance check on the 1 mL syringe verification. Hamilton <sup>after 2.5 mL syringe gv 11/14/2023</sup> ~~7903~~ <sup>gv 10/31/2023</sup> MLW008E1742 is <sup>EM7903</sup> ~~is~~ <sup>gv 10/31/2023 gv 11/01/2023</sup> back in service and ready to be used in casework.  
 -gv 10/31/2023