



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Pipette Performance Check

Pipette Number: 7903

Balance Number: Analytical: C421158344 Top Loading: 1123253484

Volume Testing (µL)	Mass 1 (g)	Mass 2 (g)	Mass 3 (g)	Mass 4 (g)	Mass 5 (g)	Average Mass (g)	Average Volume (µL)
1000 µL	0.9973 g	0.9976 g	0.9978 g	0.9980 g	0.9969 g	0.99752 g	997.52 µL
100 µL	0.0997 g	0.0997 g	0.0993 g	0.0996 g	0.0997 g	0.09960 g	99.60 µL
					gv 7/31/2024		

Passed

Failed

Signature: _____

Date: _____

7/31/2024

Comments: check performed after replacing tubing in Hamilton 7903 - gv 7/31/2024