



Houston Forensic Science Center
Comparative and Analytical Division - Toxicology

Pipette Performance Check

Pipette Number: 7903

Balance Number: Analytical: C421158344 Top Loading: 1123253484

Volume Testing (µL)	Mass 1 (g)	Mass 2 (g)	Mass 3 (g)	Mass 4 (g)	Mass 5 (g)	Average Mass (g)	Average Volume (µL)
1000 µL	0.9981g	0.9991g	0.9985g	0.9985g	0.9986g	0.99856g	998.56 µL
100 µL	0.0996g	0.0996g	0.1001g	0.0992g	0.0996g	0.09962g	99.62 µL
				gv 10/18/2024			

Passed

Failed

Signature: _____

Date: _____

10/18/2024

Comments: Performance Check following return from external clean, lube, part replacement (PIN: 7875-01, Belt Timing), and calibration - gv 10/18/2024