



# Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

## Pipette Performance Check

Pipette Number: 848K

Balance Number: Analytical: 1123241782 N/A Top Loading: 1123253484

Volume Testing (µL)	Mass 1 (g)	Mass 2 (g)	Mass 3 (g)	Mass 4 (g)	Mass 5 (g)	Average Mass (g)	Average Volume (µL)
<i>BAM 11/9/2023</i>							

Passed N/A  Failed

Signature: Brooke Mendenhall Date: 11/9/2023

Comments: After attempting to use pipette 848K, it was determined that the pipette was not repaired after maintenance and calibration by Rainin on 10/30/2023. The pipette was taken out of service and discarded on 11/9/2023 and a replacement will be ordered. -BAM 11/9/2023