



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology


Balance Performance Check

Analytical Balance (Serial Number: 1123241782)

2 mg = 0.002 g (0.0015 – 0.0025 g)	1 g (0.95 – 1.05 g)	5 g (4.95 – 5.05 g)
SN: 1000150532 B615327969-2	SN: 1000111378 B605067300	SN: B526104505 B546713066
gv 02/28/2024		
<p>Comments: The onsite calibration by Mettler-Toledo failed as found on 02/27/2024 (N/AI822-009-022724 - ACC - USL) gv 02/28/2024 On 02/27/2024 the analytical balance was placed out of service. gv 02/28/2024 - gv 02/28/2024</p>		
Pass <input type="checkbox"/> Fail <input type="checkbox"/>		

Top Loading Balance (Serial Number: 1123253484)

1 g (0.95 – 1.05 g)	5 g (4.95 – 5.05 g)	100 g (99.5 – 100.5 g)
SN: 1000111378 B605067300	SN: B526104505 B546713066	SN: B526104505 B546713066
gv 02/28/2024		
<p>Comments:</p>		
Pass <input type="checkbox"/> Fail <input type="checkbox"/>		

Signature: _____ 

Date: _____ 02/28/2024