



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Heat Block Thermometer Performance Check

Heat Block 1 (Serial Number: 091230001)

Thermometer Temperature (°C) :	NIST Thermometer Temperature (°C)
63.0 °C	62.4 °C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Heat Block 2 (Serial Number: 091103001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
6/15/18 am 68.5 °C 70.0 °C	72.3 °C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Signature: _____

Jana R. Kube

Date: _____

6/15/18