



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Heat Block Thermometer Performance Check

Heat Block 1 (Serial Number: 091230001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
61 °C	61.5 °C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Heat Block 2 (Serial Number: 091103001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
58 °C	58.2 °C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Signature:  11/13/2023

Date: 11/01/2023