



# Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

## Heat Block Thermometer Performance Check

Heat Block 1 (Serial Number: 091230001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
61°C	60.7°C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Heat Block 2 (Serial Number: 091103001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
58°C	58.6°C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Signature: \_\_\_\_\_

Date: 03/04/2024