



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Heat Block Thermometer Performance Check

Heat Block 1 (Serial Number: 091230001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
59°C	59.8°C
Comments: Performing check for on-board for training purposes only - gv 6/10/2024	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Heat Block 2 (Serial Number: 091103001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
58°C	58.5°C
Comments: Performing check for on-board training purposes only - gv 6/10/2024	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Signature: _____

Date: _____