



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Heat Block Thermometer Performance Check

Heat Block 1 (Serial Number: 091230001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
62.0°C	63.7°C
Comments: Performed check for on-board training purposes only - gv 6/24/2024	
Pass <input checked="" type="checkbox"/>	Fail <input type="checkbox"/>

Heat Block 2 (Serial Number: 091103001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
60.0°C	62.3°C
Comments: Performed check for on-board training purposes only - gv 6/24/2024	
Pass <input type="checkbox"/>	Fail <input type="checkbox"/>

Signature: _____

Date: _____

6/24/2024