



# Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

## Heat Block Thermometer Performance Check

Heat Block 1 (Serial Number: 091230001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
62.0 °C	63.2 °C
Comments: Performed check for on-board training purposes only - gv 7/15/2024	
Pass <input checked="" type="checkbox"/>	Fail <input type="checkbox"/>

Heat Block 2 (Serial Number: 091103001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
58.0 °C	58 °C
Comments: Performed check for on-board training purposes only - gv 7/15/2024	
Pass <input type="checkbox"/>	Fail <input type="checkbox"/>

Signature: \_\_\_\_\_

Date: \_\_\_\_\_

7/15/2024