



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Heat Block Thermometer Performance Check

Heat Block 1 (Serial Number: 091230001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
66°C	65.0°C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Heat Block 2 (Serial Number: 091103001)

Thermometer Temperature (°C)	NIST Thermometer Temperature (°C)
60°C	60.3°C
Comments:	
Pass <input checked="" type="checkbox"/> Fail <input type="checkbox"/>	

Signature: _____

Date: 10/11/2024