



Houston Forensic Science Center

Forensic Analysis Division - Toxicology

pH Meter Performance Check

Manufacturer: Thermo Scientific

Model: STARA2140

Serial Number: X29761

Weekly Maintenance:

Refill electrode and verify expiration date

Rinse electrode

Verify electrode storage solution and expiration date

Verify pH buffer solution expiration dates

Date:

12/13/2021

Signature:

[Handwritten Signature]

Comments: _____

Date	Signature	pH 4.01*	pH 7.00*	pH 10.01*	Slope (%) (90-110%)	Pass/Fail
		Temperature (°C)	Temperature (°C)	Temperature (°C)		
<p>*Target pH at 25°C. The acceptable pH at the reported temperature is target pH ± 0.022. Refer to pH buffer solution bottle label for target pH at various temperatures.</p>						

EP 12/20/2021

Form Complete Date/Signature:

[Handwritten Signature] 12/20/2021