Houston Forensic Science Center

Forensic Analysis Division - Toxicology

pH Meter Performance Check

Manufacturer: Mettler-Toledo
Model: SevenDirect™ SD50
Serial Number: C202370166

Weekly Maintenance:

Rinse electrode ☑ Date: 03/18/2024

Verify electrode storage solution and expiration date ☑

Verify pH buffer solution expiration dates ☑ Initials: GV

Comments: _____

Date	Initials	pH 4.0 / Temperature (°C)	pH 7.0 / Temperature (°C)	pH 10.0 / Temperature (°C)	Slope (%) (95-105%)	Calibration ✓/X	Verification ✓/X
						GV 03/20/2024	

Form Complete Signature/Date: GV 03/25/2024