



Belair Systems Inc.
 Service, Installation, and Repairs Division
 28364 S. Western Ave. #302
 Rancho Palos Verdes, Ca. 90275

Office: (310) 528-4293
 Fax: (310) 831-1684

Date: 10/11/2024

Service Report

Address & Description: 500 JEFFERSON STREET 0 Houston, TX 77002	Company: HOUSTON FORENSIC SCIENCE Contact: Dayong Lee Telephone #: dlee@hfsctx.gov Fax/Email: #REF!						
	<table border="1"> <tr> <th>Cuts. PO#</th> <th>Machine Type</th> <th>Serial No.</th> </tr> <tr> <td></td> <td>generator: : H2PEMPD-510-10</td> <td>PLD5146, 19PLD5147, 19PLD5150</td> </tr> </table>	Cuts. PO#	Machine Type	Serial No.		generator: : H2PEMPD-510-10	PLD5146, 19PLD5147, 19PLD5150
	Cuts. PO#	Machine Type	Serial No.				
		generator: : H2PEMPD-510-10	PLD5146, 19PLD5147, 19PLD5150				
	<table border="1"> <tr> <th>Job No.</th> <th>Job Completion Date</th> <th>Location</th> </tr> <tr> <td>8237</td> <td>10/7/2024</td> <td></td> </tr> </table>	Job No.	Job Completion Date	Location	8237	10/7/2024	
Job No.	Job Completion Date	Location					
8237	10/7/2024						
<table border="1"> <tr> <th>Warranty</th> <th>Event</th> <th>Technician</th> </tr> <tr> <td></td> <td>121381</td> <td>O. Delgado</td> </tr> </table>	Warranty	Event	Technician		121381	O. Delgado	
Warranty	Event	Technician					
	121381	O. Delgado					
Reason For Visit: Generator: : H2PEMPD-510-100 S/N 19PLD5146, 19PLD5147, 19PLD5150 Service	Recommendation to Customer: Call Belair Systems Inc. for all of your preventative maintenance, installation, and certification needs!						

Model #:	Generator: : H2PEMPD-510-100
Serial #'s:	S/N 19PLD5146, 19PLD5147, 19PLD5150

Description of Service/Installation:

The technician found two generators running with 90 PSI inlet pressure and one generator about to start, the technician isolated the generators to depressurize them, once the generators were depressurized the technician replaced the deionizer cartridge, 100 micron water filter, and environmental filter, once the PM service was completed the technician left the generators running properly leaving the same pressures. next service will be 04/2025.

Parts	Quantity	Description
Parts Used: 3 MKH2PEMPD-6M 30071883		

Time Table							Hourly Total	Notes:
Mon	Tues	Wed	Thur	Fri	Sat	Sun		

Date:	10/7/2024	Customer Signature:	
-------	-----------	---------------------	--