Belair Systems ***Josh Pray*** ***603-264-2954***	**Date**	11/14/2022
	Company	HOUSTON FORENSIC SCIENCE
	Contact	Dayong Lee
	Fax/Email	dlee@hfsctx.gov
	Phone #	(832) 993-0175
	Event #	118048
	Machine Type	H2PEMPD-510-100
	Serial Number	19PLD5146, 19PLD5147, 19PLD5150
	Location	Houston, TX 77002
	Technician	Josh Pray

Reason For Visit	**Reccomendations**
6 Month PM	

Description of service/ Installation

All three units on and running well at 90psi SET/ACT, Hours on all three units between 24000 and 25000, tech isolated each unit, replaced environmental filters, resin beds and the water filter, each unit running to spec, placed all unused parts from 24M PM kits next to the generators

Part Used	Quantity	Description
MKH2PEMPD-6M	3	6 MONTH SERVICE KIT

Date	**Hours**	
11/14/2022	0.75	

Signature

x