

Service Report



Service Job No : 58337
Serial No : 771053385
Item No : 10-6025
Description : Genius NM32LA-A 230V
Visit Date : 03/09/2023
Service Call Cat : SC - Planned Maintenance
Contract Type : Complete Plan 48/72
Engineer : Robert Hebert (TX)

Customer and Generator Location Details

Customer Billing Address
Houston Forensic Science Center

500 Jefferson St.
FI 13
Houston
USA
77002

Generator Location Details

CassandraDuvall-1
832-975-1058

Measurements

Compressor Run Hours:4228
Compressor Serial Number: 31900062905 Thomas 3rd refit
Output Flow Setting: Meets Specification
Output Pressure: Meets Specification
Software Version: 4.10
First Time Fix?: Yes
Compressor: Refitted
Compressor2:Refitted
Compressor3:N/A
Compressor4:N/A
Compressor Inlet Filter Element:Replaced
Drain Valves:Replaced
Leak Check: Pass
Off-Loader Valves:Replaced
Operational Testing:Pass
Output Settings:Pass
RAC Filter Element:Replaced
Supply Voltage : Meets Specification
System Run Hours : 28895
Ambient Humidity : Meets Specification
Ambient temperature :Meets Specification
Application : Unknown
System Run Hour : 28895
Coalescing Filter Elements:Replaced
Compressor 2 Serial Number:31900062921 Thomas 3rd refit
Compressor 2 Run Hours:4139
Compressor 3 Serial Number:N/a
Compressor 3 Run Hours:N/a
Compressor 4 Serial Number:N/a

Visit Notes

Completed annual pm.
Performed 3rd refit on compressors 1@2.
Replaced all solenoid valves, relays, and filters.
Tested generator per qd09 specs.
No major issues to report.
Generator is functioning as it should.

Labour

Labour - Hrs and Mins	Labour Hrs Qty	Labour Rate PH	Labour Total
0000:03:45:00	3.75	0.00	0.00
Labour Total			0.00

Sales Order Items

Related Order	Item no.	Item name	Qty.
20016442	3300180	PS-Genius NM32LA-A Complete	1.00

Customer Signature

No rep

Cassandra
Duvall

Digitally signed by Cassandra Duvall
DN: cn=Cassandra Duvall, o=Houston
Forensic Science Center, ou=Toxicology,
email=cduvall@houstonforensicscience.
org, c=US
Date: 2023.03.13 12:12:43 -0500