

Service Report



Service Job No.: 90656
Contract ID: US-32382
Serial No.: 771053385
Item No.: 10-6025
Description: Genius NM32LA-A 230V

Visit Date: 07/22/2024
Service Call Cat: Complete Contract Breakdown
Contract Type: Complete Plan 48/72
Engineer: Robert Hebert (TX)
Contract PO: 24-11-006

Customer and Generator Location Details

Customer Billing Address

Houston Forensic Science Center Inc

Fl 13
500 Jefferson St.
Houston
Texas
USA
77002--7359

Generator Location Details

Houston Forensic Science Center

Floor 13
500 Jefferson St.
Houston
Texas
USA
77002
USA

Contact Person: Cassandra Duvall
Contact No.: 832-975-1058

Visit Notes

Customer states the generator is leaking. Potentially coming from one of the compressors.

Upon inspection found a 6mm adapter push fitting for a Nrv that was leaking .

Replaced damaged adapter.

Tested generator per qd09 specs with no noticeable issues found.

Returned generator back to customers control.

Measurements and Actions

System Run Hour: 40818

Compressor Serial Number: 22302024923

Compressor 2 Serial Number: 2230202933

Compressor 3 Serial Number: N/a

Compressor 4 Serial Number: N/a

Ambient Temperature: Meets Specification

Ambient Humidity: Meets Specification

Output Pressure: Meets Specification

Output Flow Setting: Meets Specification

Software Version: 4.10

Supply Voltage: Meets Specification

Application: Unknown

Compressor: Pass

Compressor 2: Pass

Compressor 3: N/A

Compressor 4: N/A

Filters: Pass

Off-Loader Valves: Pass

Drain Valves: Pass

Nitrogen Membrane: Pass

PLC: Pass

Leak Check: Pass

Operational Testing: Pass

First Time Fix?: Yes

Parts Used

Item no.	Item name	Qty.	List Price	Line Item Total
3300715	Reducer 10mm to 6mm push fit	1.00	0.00	0.00
Parts Used Total				0.00

Service Report

Service Job No.: 90656



Customer Signature **Cassandra Duvall** Digitally signed by
Cassandra Duvall
Date: 2024.07.23 11:13:15
-05'00'

A handwritten signature in black ink, appearing to be "RH" with a horizontal line underneath.