

Service Report



Service Job No : 69002
Serial No : 771053079
Item No : 10-6025
Description : Genius NM32LA-A 230V
Visit Date : 11/04/2022
Service Call Cat : SC - CSP Breakdown
Contract Type : Complete Plan 48/72
Engineer : Robert Hebert (TX)

Customer and Generator Location Details

Customer Billing Address

Houston Forensic Science Center

500 Jefferson St.
Fl 13
Houston
USA
77002

Generator Location Details

500 Jefferson St.
Floor 13
Houston
TX
USA
77002

CassandraDuvall-1
832-975-1058

Measurements

Compressor Run Hours:N/a
Compressor Serial Number: 11900061804
Output Flow Setting: Meets Specification
Output Pressure: Meets Specification
Software Version: 4.10
First Time Fix?: Yes
Compressor: Pass
Compressor2:Pass
Compressor3:N/A
Compressor4:N/A
Drain Valves:Pass
Additional Parts Used:Yes
Filters:Pass
Leak Check: Pass
Nitrogen Membrane:Replaced
Off-Loader Valves:Pass
Operational Testing:Pass
PLC:Pass
Compressor 4 Run Hours:N/a
Supply Voltage : Meets Specification
System Run Hours : N/a
Ambient Humidity : Meets Specification
Ambient temperature :Meets Specification
Application : Unknown
System Run Hour : N/a
Compressor 2 Serial Number:11900061804
Compressor 2 Run Hours:N/a
Compressor 3 Serial Number:N/a
Compressor 3 Run Hours:N/a
Compressor 4 Serial Number:N/a

Visit Notes

M2 membrane failed:
Replaced failed n2 membrane.
Tested generator per qd09 specs.
Generator is functioning as it should.

Labour

Labour - Hrs and Mins	Labour Hrs Qty	Labour Rate PH	Labour Total
0000:11:15:00	11.25	0.00	0.00
Labour Total			0.00

Parts Used

Item no.	Item name	Qty.	List Price	Line Item Total
06-4418	Membrane N2 Dia54mmODx610mm	1.00	0.00	0.00

Parts Used Total

0.00

Customer Signature

No rep