



Houston Forensic Science Center

Comparative and Analytical Division - Toxicology

Temperature Log

Device: Freezer -3

Manufacturer: VWR

Model Number: MFV-17

Serial Number: VWR-WB74750285-1802

Date	Temperature (°C)	Acceptable	Initials	Comments
4/26/2022	-25.6	yes	BAM	DicksonOne device No. 16356678. DicksonOne reading at 9:06AM. Everything stored in Freezer-3 was moved to Freezer-4. The DicksonOne's warm alert alarm was paused on Freezer-3. Freezer-3 was unplugged at 10:54AM for defrosting.
4/27/2022	BAM 4/27/2022			Freezer-3 was defrosted. Replacing the door gasket. Freezer-3 was plugged back in at 11:50AM after defrost + door gasket being replaced.
4/28/2022	-24.1	yes	BAM	Freezer-3 was left overnight to get to the appropriate temperature. The DicksonOne's warm alert alarm for Freezer-3 was unpaused. DicksonOne reading at 8:43AM.
BAM 4/26/2022				Freezer-3 is back in-service.
BAM 4/26/2022				

Signature: Bruce Mendel

Date Completed: 4/28/2022