



Validation of Quantitative Methods

Analyte	Amphetamines (amphetamine, methamphetamine, 3,4-methylenedioxyamphetamine (MDA), 3,4-methylenedioxymethamphetamine (MDMA), and 3,4-methylenedioxy-N-ethylamphetamine (MDEA))
Unit of Measure	ng/mL
Analyst Performing Validation Studies	Sara Dempsey, Pucheng Ke, and Corissa L. Rodgers
Responsible Supervisor	Dayong Lee
Start Date	August 14, 2020
Completion Date	September 24, 2020
Primary Matrix	Blood
Secondary Matrix	N/A
Low Calibrator Concentration	10 ng/mL
Highest Calibrator Concentration	500 ng/mL for amphetamine, MDA, and MDEA; 1000 ng/mL for methamphetamine and MDMA
Equipment/Instrument	LCMS-1
Instrument Serial Number	SG1939G104
Method	AMP.M

Validation Approval

Analyst: Sara Dempsey Digitally signed by Sara Dempsey
Date: 2020.11.20 11:58:45 -06'00' 11/20/2020 Date

Analyst: Pucheng Ke Digitally signed by Pucheng Ke
Date: 2020.11.20 16:51:15 -06'00' 11/20/2020 Date

Analyst: Corissa L. Rodgers, M.S. Digitally signed by Corissa L. Rodgers, M.S.
Date: 2020.11.20 13:03:19 -06'00' 11/20/2020 Date

Responsible Supervisor: _____ Date

Quality Director: _____ Date

This method validation study, even though it may depict an approval date by both the Toxicology section and the Quality Division, will not be used in casework until ANAB grants an expansion in our scope of accreditation to include this test method and instrument technology.

METHOD VALIDATION PROTOCOL AND RESULTS

Analyte: Amphetamine
 Units: ng/mL
 Method: AMP.M
 Instrument: LCMS-1
 SOP Reference: Toxicology Analytical Manual v3.5

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

VALIDATION EXPERIMENT		SOP CRITERIA	RESULTS	COMMENTS
1	Weight Verification	The least complex weighting scheme that minimizes $\sum %RE $	%RE = 91.51 ($1/x^2$)	1/x ² weighting will be used. %RE for unweighted and 1/x weighting are 209.09 and 112.01, respectively. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.
1	Validation Calibration	%RE Calibrators $\pm 20\%$ of target	Max %RE = 8.91	N/A
1	Case Work Calibration	%RE Calibrators $\pm 20\%$ of target	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
1	Comparison of Validation Calibration to Casework Calibration	95% CI of slope includes 1 95% CI of intercept includes 0	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
2	Limit of Detection (LOD)	Signal to Noise ≥ 3.3 Acceptable detection and identification criteria	LOD = 10 ng/mL S:N= 1403	The lowest non-zero calibrator was defined as the LOD.
3	Limit of Quantitation (LOQ)	Bias: $\leq 20\%$ Within-Run Precision: CV $\leq 20\%$ Between-Run Precision: CV $\leq 20\%$	Bias = 2.76% Within-Run Precision = 10.56% Between-Run Precision = 5.97%	Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Number in red was $> 20\%$ target value. Numbers with * were $> 20\%$ target value due to insufficient IS.
4	Bias & Precision	%Bias $\leq 20\%$ Within-Run %CV $\leq 20\%$ Between-Run %CV $\leq 20\%$	Max Bias = -10.15% Max Within-Run Precision = 18.13% Max Between-Run Precision = 12.48%	Number in red was not within 20% of the target value.
5	Carryover	No analyte carryover is observed in the matrix blank samples; response in blank samples is $< LOQ$ of the method.	No significant carryover observed following samples containing analyte at up to 2000 ng/mL	N/A
6	Matrix Interference	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020
6	Interference from stable isotope internal standard	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	
6	Exogenous Substances Interferences	Concentrations of analytes of interest within $\pm 20\%$ of the average concentration obtained in the Bias and Precision studies	No significant interference observed.	N/A
7	Dilution Integrity	Average %Bias must be less than 20%	It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.	N/A
8	Processed Sample Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	N/A	Study not performed. Sample preparation will be completed once started without prolonged interruptions.
8	Autosampler Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	Punctured and unpunctured samples were shown to be stable for 48 hours.	AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not stable; therefore the 72 hr data for amphetamine were not included. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.
9	Matrix Effects	Average suppression/enhancement $\leq 25\%$ or the \leq CV of the suppression/enhancement 20%	Average percent suppression values for LQC and HQC were -6% and -5%, respectively.	N/A

* Alternate criteria is any deviation from that described in the SOP as determined by the Supervisor

ADDITIONAL STUDIES (Describe any additional studies required for this validation.)

Validation Study 1

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1



LINEARITY
 Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Study Date	Target (x)	Target (x) ²	Calculated (y)	Predicted	Residual	%RE
AMP_20200814B_SD	10	100	10.00	10.40	-0.41	0.02
	20	400	19.86	20.18	-0.33	0.71
	50	2500	50.74	49.60	1.13	1.47
	100	10000	102.93	98.90	4.03	2.93
	250	62500	231.20	248.90	-17.60	7.52
	500	250000	528.94	505.28	23.65	5.79
AMP_20200814B_PK	10	100	9.83	10.40	-0.57	1.70
	20	400	21.31	20.18	1.12	6.53
	50	2500	45.54	49.60	-4.06	8.91
	100	10000	99.96	98.90	1.06	0.04
	250	62500	267.96	248.80	19.17	7.19
	500	250000	484.74	505.28	-20.54	3.05
AMP_20200828B_PK	10	100	9.83	10.40	-0.58	1.72
	20	400	20.90	20.18	0.71	4.49
	50	2500	49.18	49.60	-0.42	1.64
	100	10000	97.24	98.90	-1.66	2.76
	250	62500	255.38	248.80	6.58	2.15
	500	250000	497.35	505.28	-7.94	0.53
AMP_20200828B_SD	10	100	9.92	10.40	-0.48	0.77
	20	400	20.10	20.18	-0.09	0.48
	50	2500	52.35	49.60	2.75	4.70
	100	10000	96.83	98.90	-2.07	3.17
	250	62500	241.70	248.80	-7.10	3.32
	500	250000	511.06	505.28	5.78	2.21
AMP_20200901B_CLR	10	100	9.81	10.40	-0.59	1.89
	20	400	21.08	20.18	0.90	5.42
	50	2500	47.46	49.60	-2.14	5.07
	100	10000	102.53	98.90	3.62	2.53
	250	62500	245.40	248.80	-3.40	1.84
	500	250000	504.74	505.28	-0.54	0.95

Regression Statistics	
Multiple R	0.998956279
R Square	0.997913647
Adjusted R Square	0.997759102
Standard Error	8.459442209
Observations	30

ANOVA					
	df	SS	MS	F	Significance F
Regression	2	924170.9047	462085.4523	6457.119744	6.48241E-37
Residual	27	1932.178387	71.56216249		
Total	29	926103.0831			

	Coefficients	Standard Error	t Stat	P-value	Lower 95%	Upper 95%	Lower 95.0%	Upper 95.0%
Intercept	0.636927298	2.662898457	0.239185725	0.81276791	-4.826889018	6.100743613	-4.826889018	6.100743613
target x	0.97599925	0.036569563	26.68884094	6.08314E-21	0.900964705	1.051033796	0.900964705	1.051033796
x^2	6.6589E-05	7.04575E-05	0.94509442	0.352992162	-7.79778E-05	0.000211156	-7.79778E-05	0.000211156



Results: The reported %RE is using 1/x² weighting. Σ|%RE| 91.51
 Max %RE 8.91

Comments: 1/x² weighting will be used. %RE for unweighted and 1/x weighting are 209.09 and 112.01, respectively. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.

Acceptance Criteria:

The least complex weighting scheme that minimizes Σ|%RE|
 %RE Calibrators ±20% of target

Validation Study 2

Sensitivity (LOD)

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

		Signal to Noise Ratio		
Concentration (ng/mL)	10			
AMP_20200814B_SD	1273			
	732			
	667			
	2315			
	1199			
	2541			
	2679			
	2447			
AMP_20200828B_PK	1504			
	338			
	1197			
	505			
	953			
	820			
	1742			
	1275			
AMP_20200901B_PK	2773			
	704			
	2098			
	709			
	238			
	1944			
	1338			
	1933			
AMP_20200901B_CLR	1187			
	688			
	2072			
	1935			
	630			
	1360			
	1617			
	2223			
Average Signal to Noise:	1403			

Established LOD: 10 ng/mL
 Signal to Noise: 1403

Comments: The lowest non-zero calibrator was defined as the LOD.

Acceptance Criteria:

Signal to Noise ≥ 3.3
 Acceptable detection and identification criteria

Validation Study 3

SENSITIVITY (LOQ)

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LOQ
<i>Target Concentration (ng/mL):</i>		10
AMP_20200814B_SD	1-1	10.32
	1-2	10.54
	1-3	9.76
	1-4	10.05
	1-5	10.03
	1-6	11.02
	1-7	10.44
	1-8	10.17
	1-9	9.95
	<i>Within Run</i>	Mean
	SD	0.38
	%CV	3.69%
	% Bias	2.53%
AMP_20200901B_PK	2-1	10.52
	2-2	10.41
	2-3	10.16
	2-4	10.25
	2-5	10.65
	2-6	11.09
	2-7	10.72
	2-8	10.79
	2-9	10.71
	<i>Within Run</i>	Mean
	SD	0.29
	%CV	2.70%
	% Bias	5.87%
AMP_20200901B_CLR	3-1	57.75
	3-2	10.04
	3-3	9.77
	3-4	10.00
	3-5	9.29
	3-6	10.51
	3-7	12.90
	3-8	10.15
	3-9	10.05
	<i>Within Run</i>	Mean
	SD	1.09
	%CV	10.56%
	% Bias	3.38%
AMP_20200828B_PK	4-1	10.28
	4-2	9.81
	4-3	10.18
	4-4	9.54
	4-5	9.76
	4-6	3.93*
	4-7	9.92
	4-8	10.10
	4-9	9.82
	<i>Within Run</i>	Mean
	SD	0.24
	%CV	2.45%
	% Bias	-0.73%

Mean		10.28
SD		0.61
Precision (%CV)	<i>Max Within-Run</i>	10.56%
	<i>Between-Run</i>	5.97%
% Bias		2.76%

Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Number in red was > 20% target value. Numbers with * were > 20% target value due to insufficient IS.

Acceptance Criteria: Bias: ≤20%
 Within-Run Precision: CV ≤20%
 Between-Run Precision: CV ≤20%

Validation Study 4

BIAS AND PRECISION

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LQC	MQC	UTAK	HQC	
<i>Target Concentration (ng/mL):</i>		25	100	73.97	400	
AMP_20200814B_SD	1-1	25.18	92.86	63.75	356.77	
	1-2	23.61	88.67	68.46	347.61	
	1-3	24.71	91.58	65.43	347.62	
	1-4			69.28		
	<i>Within Run</i>	Mean	24.50	91.04	66.73	350.67
		SD	0.80	2.15	2.59	5.28
		%CV	3.28%	2.36%	3.87%	1.51%
	% Bias	-1.99%	-8.96%	-9.79%	-12.33%	
AMP_20200814B_PK	2-1	25.15	81.17	79.66	392.65	
	2-2	28.47	83.01	80.62	403.58	
	2-3	27.54	89.47	92.96	390.02	
	2-4			95.46		
	<i>Within Run</i>	Mean	27.05	84.55	87.18	395.42
		SD	1.71	4.36	8.20	7.19
		%CV	6.34%	5.16%	9.40%	1.82%
	% Bias	8.21%	-15.45%	17.85%	-1.14%	
AMP_20200828B_PK	3-1	25.51	90.29	67.88	365.48	
	3-2	27.62	97.78	70.78	363.46	
	3-3	24.80	86.74	71.94	370.54	
	3-4			66.47		
	<i>Within Run</i>	Mean	25.98	91.61	69.27	366.49
		SD	1.47	5.64	2.53	3.65
		%CV	5.66%	6.15%	3.65%	0.99%
	% Bias	3.76%	-9.16%	-6.79%	-9.14%	
AMP_20200828B_SD	4-1	23.00	98.12	73.63	360.25	
	4-2	24.10	93.39	77.05	355.84	
	4-3	24.22	95.83	69.30	354.35	
	4-4			70.85		
	<i>Within Run</i>	Mean	23.77	95.78	72.71	356.81
		SD	0.67	2.37	3.40	3.07
		%CV	2.83%	2.47%	4.68%	0.86%
	% Bias	-4.91%	-4.22%	-1.71%	-10.80%	
AMP_20200901B_CLR	5-1	27.25	109.59		356.50	
	5-2	29.13	76.21		367.53	
	5-3	25.16	90.95		362.03	
	<i>Within Run</i>	Mean	27.18	92.25		362.02
		SD	1.98	16.73		5.52
		%CV	7.30%	18.13%		1.52%
		% Bias	8.73%	-7.75%		-9.49%
AMP_20200901B_PK	6-1	24.67	80.30		370.86	
	6-2	24.64	87.92		361.30	
	6-3	24.32	83.34		360.92	
	<i>Within Run</i>	Mean	24.54	83.85		364.36
		SD	0.19	3.83		5.63
		%CV	0.79%	4.57%		1.55%
		% Bias	-1.82%	-16.15%		-8.91%
Mean		25.51	89.85	73.97	365.96	
SD		1.74	7.82	9.23	15.22	
Precision (%CV)	Max Within-Run	7.30%	18.13%	9.40%	1.82%	
	Between-Run	6.82%	8.70%	12.48%	4.16%	
	% Bias	2.02%	-10.15%	0.00%	-8.51%	

Comments: Number in red was not within 20% of the target value.

Acceptance Criteria: %Bias ≤20%
 Within-Run %CV ≤20%
 Between-Run %CV ≤20%

Validation Study 5**CARRYOVER**

Analyte: Amphetamine
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Average LOQ Response*: _____ 136552 _____

Study Date:	Response		
	AMP_20200814B_SD	AMP_20200828B_PK	AMP_20200901B_PK
Concentrated Sample (2000 ng/mL)	11401848	14582622	10479640
Blank	3916	2620	7013
%LOD Response	2.87%	1.92%	5.14%

Maximum Response in Blank: **5.1%**

Results: No significant carryover observed following samples containing analyte at up to 2000 ng/mL.
Comments: N/A

Acceptance Criteria: No analyte carryover is observed in the matrix blank samples; response in blank samples is <LOQ of the method.

Validation Study 6

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

MATRIX & IS INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

	LOQ Response	
	Analyte	IS
Run 1	169346	627849
Run 2	92912	413897
Run 3	182678	700000
Run 4	106895	466605
Run 5	130930	502265
Average	136552	542123

Matrix Interference Study Date: AMP_20200814B_PK

Matrix Source	Peak at Analyte RT	
	Peak Response	Percent of LOQ Response
4	920	0.67%
5	1067	0.78%
7	1404	1.03%
8	1171	0.86%
9	1737	1.27%
10	1776	1.30%
11	1587	1.16%
12	1041	0.76%
13	1307	0.96%
14	1657	1.21%

Interference from Stable-Isotope Internal Standards Study Date: AMP_20200828B_SD

Experiment	Peak at Analyte RT		Peak at IS RT	
	Peak Response	Percent of LOQ Response	Peak Response	Percent of LOQ Response
Matrix with IS but no D0 (IS = 50 ng/mL)	2822.47	2.07%	N/A	N/A
Matrix with D0 but no IS (D0 = 2000 ng/mL)	N/A	N/A	213.67	0.04%

Matrix Interference: No significant interference observed.

IS Interference: No significant interference observed.

Comments: LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020

Acceptance Criteria:

Response of blank matrix is less than 20% the average response of LOQ

Validation Study 6

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

EXOGENOUS SUBSTANCE INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target LQC Concentration (µg/mL): 25.51
 Control Acceptance: 20%
 Run Date: AMP 20200828B SD

Group	Compound	Compound Concentration (µg/mL)	Calculated LQC Concentration (ng/mL)	% Difference from Target	Comment	
Benzodiazepines	Alprazolam	1	24.39	-4%	No significant interference	
	α-Hydroxyalprazolam					
	Clonazepam					
	7-Aminoclonazepam					
	Diazepam					
	Nordiazepam					
	Temazepam					
	Lorazepam					
	Oxazepam					
Zolpidem						
Phencyclidine	Phencyclidine	1	26.05	-2%	No significant interference	
	Morphine	2				
Opioids	Hydrocodone	1	26.33	3%	No significant interference	
	Buprenorphine	0.5				
	Norbuprenorphine					
	Codeine					
	Fentanyl					
	Norfentanyl oxalate					
	Hydromorphone					
	Methadone					
	EDDP					
	Oxycodone					
	Oxymorphone					
	Tramadol					
	o-Desmethyltramadol					
	Cocaine and Metabolites					Benzoylcegonine
Cocaine			0.5			
Cocacethylene						
Carisoprodol/Meprobamate	Carisoprodol	1	25.07	-2%	No significant interference	
	Meprobamate					
Basic and Neutral Mix	Amtriptyline	5	24.96	-2%	No significant interference	
	Benzpiperazine					
	Chlorpheniramine					
	Cyclobenzaprine					
	Dextromethorphan					
	Diphenhydramine					
	Doxylamine					
	Fluoxetine					
	Imipramine					
	Ketamine					
	Norketamine					
	Meperidine					
	Nortriptyline					
	Propoxyphene					
	Sertraline					
	Trazodone					
	Venlafaxine					
	Zopiclone					
Over-the-Counter Drugs	Acetaminophen	10	23.62	-7%	No significant interference	
	Caffeine					
	Ibuprofen					
	Naproxen					
	Pseudoephedrine					
	Theobromine					1
Cannabinoids	Δ9-THC	0.5	25.23	-1%	No significant interference	
	11-Hydroxy-Δ9-THC					
	Δ9-THC-COOH					
	Δ8-THC					
	Cannabinol					
Cannabinolic acid						

Conclusions: No significant interference observed.

Comments: N/A

Acceptance Criteria: Concentrations of analytes of interest within ±20% of the average concentration obtained in the Bias and Precision studies

Validation Study 7

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

DILUTION INTEGRITY

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target Concentration (ng/mL): 347.84

Run Date	Dilution Factor	Diluted Result (ng/mL)	Calculated Result (µg/mL)	Mean	SD	%Bias	Absolute % Bias	Within-Run Precision (per Dilution)
AMP_20200814B_SD	2	174.76	349.52	334.78	19.21	0.48	0.48	5.74
	2	170.88	341.77			-1.75	1.75	
	2	156.53	313.05			-10.00	10.00	
	5	71.21	356.04	341.60	16.67	2.36	2.36	4.88
	5	69.08	345.41			-0.70	0.70	
	5	64.67	323.35			-7.04	7.04	
	10	37.78	377.81	350.07	33.83	8.62	8.62	9.66
	10	36.00	360.02			3.50	3.50	
	10	31.24	312.38			-10.19	10.19	
	AMP_20200828B_PK	2	185.55	371.09	397.18	24.63	6.69	6.69
2		200.20	400.40	15.11			15.11	
2		210.02	420.04	20.76			20.76	
5		75.29	376.43	403.67	24.71	8.22	8.22	6.12
5		84.93	424.66			22.09	22.09	
5		81.98	409.92			17.85	17.85	
10		36.02	360.17	397.91	32.97	3.55	3.55	8.29
10		42.12	421.15			21.08	21.08	
10		41.24	412.40			18.56	18.56	
AMP_20200901B_PK		2	191.73	383.47	396.07	11.64	0.74	0.74
	2	203.21	406.42	6.77			6.77	
	2	199.16	398.32	4.64			4.64	
	5	72.88	364.40	389.94	22.53	-4.27	4.27	5.78
	5	79.69	398.44			4.67	4.67	
	5	81.40	407.00			6.92	6.92	
	10	39.89	398.92	385.87	22.15	4.80	4.80	5.74
	10	36.03	360.30			-5.35	5.35	
	10	39.84	398.38			4.65	4.65	
	AMP_20200828B_SD	2	182.29	364.58	361.56	9.25	4.81	4.81
2		184.47	368.93	6.06			6.06	
2		175.59	351.18	0.96			0.96	
5		74.11	370.55	364.92	12.53	6.53	6.53	3.43
5		74.73	373.66			7.42	7.42	
5		70.11	350.56			0.78	0.78	
10		32.51	325.06	342.41	19.55	-6.55	6.55	5.71
10		36.36	363.59			4.53	4.53	
10		33.86	338.57			-2.66	2.66	
AMP_20200901B_CLR		2	151.91	303.81	343.13	37.67	-12.66	12.66
	2	189.45	378.91	8.93			8.93	
	2	173.33	346.67	-0.34			0.34	
	5	73.11	365.53	379.82	12.44	5.09	5.09	3.27
	5	77.14	385.72			10.89	10.89	
	5	77.64	388.21			11.61	11.61	
	10	37.12	371.17	361.88	9.15	6.71	6.71	2.53
	10	35.29	352.87			1.45	1.45	
	10	36.16	361.60			3.96	3.96	

Dilution Factor	2
Avg %Bias	6.71
Max Within-Run Precision (%CV)	10.98

Dilution Factor	5
Avg %Bias	7.76
Max Within-Run Precision (%CV)	6.12

Dilution Factor	10
Avg %Bias	7.08
Max Within-Run Precision (%CV)	9.66

Results: It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.
 Comments: N/A

Acceptance Criteria:

Avg %Bias must be less than 20%

Validation Study 8Analyte: Amphetamine
Units: ng/mL
Instrument: LCMS-1**PROCESSED SAMPLE STABILITY**Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Run Date	Sample	Time Zero Response	24H Response	48H Response	72H Response	Acceptable Range		Pass/Unstable
						Low	High	
	LQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			
	HQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			

Results: N/A**Comments:** Study not performed. Sample preparation will be completed once started without prolonged interruptions.**Acceptance Criteria:**

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 8

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

AUTOSAMPLER STABILITY

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours			% Difference from Time Zero (Response Ratio)
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	
AMP_20200923B_SD & AMP_20200924B_SD	Calibrator 1	10.07	0.29	10.03	9.97	0.26	-8%
	Calibrator 2	19.41	0.53	19.42	18.87	0.48	-9%
	Calibrator 3	52.51	1.36	52.21	50.77	1.24	-9%
	Calibrator 4	98.57	2.47	99.06	96.56	2.28	-8%
	Calibrator 5	242.13	5.54	243.33	232.58	5.04	-9%
	Calibrator 6	511.64	9.68	517.12	495.73	8.97	-7%

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours			% Difference from Time Zero	48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio		Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	Calibrator 1	9.99	0.27	N/A	N/A	N/A	N/A	9.89	10.15	0.27	-1%
	Calibrator 2	20.26	0.50	N/A	N/A	N/A	N/A	20.08	19.05	0.50	-1%
	Calibrator 3	47.76	1.10	N/A	N/A	N/A	N/A	47.58	43.29	1.10	0%
	Calibrator 4	103.18	2.28	N/A	N/A	N/A	N/A	104.10	94.05	2.30	1%
	Calibrator 5	252.38	5.22	N/A	N/A	N/A	N/A	257.84	241.25	5.32	2%
	Calibrator 6	495.41	9.26	N/A	N/A	N/A	N/A	502.15	545.95	9.36	1%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours			% Difference from Time Zero (Response Ratio)
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	
AMP_20200923B_SD & AMP_20200924B_SD	LOC	24.01	0.65	23.94	23.43	0.59	-9%
		24.04	0.65	23.97	23.52	0.59	-8%
		24.09	0.65	24.06	23.50	0.59	-9%
	HQC	355.37	7.53	358.66	343.70	6.93	-8%
		353.53	7.50	362.93	343.36	6.92	-8%
		354.29	7.52	358.23	343.16	6.92	-8%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours			% Difference from Time Zero	48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio		Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	LOC	24.59	0.60	N/A	N/A	N/A	N/A	24.56	22.98	0.60	0%
		24.58	0.60	N/A	N/A	N/A	N/A	24.83	23.22	0.60	1%
		24.86	0.60	N/A	N/A	N/A	N/A	24.87	23.25	0.60	0%
	HQC	350.17	6.95	N/A	N/A	N/A	N/A	358.81	349.91	7.10	2%
		350.32	6.96	N/A	N/A	N/A	N/A	355.78	346.44	7.05	1%
		352.84	7.00	N/A	N/A	N/A	N/A	355.84	346.50	7.05	1%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours			% Difference from Time Zero (Response Ratio)
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	
AMP_20200923B_SD & AMP_20200924B_SD	LOC	24.01	0.65	23.89	23.37	0.59	-9%
		24.04	0.65	23.92	23.62	0.59	-8%
		24.09	0.65	23.83	23.38	0.59	-9%
	HQC	355.37	7.53	362.06	346.22	6.97	-8%
		353.53	7.50	359.17	339.87	6.87	-8%
		354.29	7.52	359.49	342.61	6.91	-8%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours			% Difference from Time Zero	48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio		Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	LOC	24.59	0.60	N/A	N/A	N/A	N/A	24.60	23.01	0.60	0%
		24.58	0.60	N/A	N/A	N/A	N/A	24.71	23.11	0.60	0%
		24.86	0.60	N/A	N/A	N/A	N/A	24.57	22.99	0.60	-1%
	HQC	350.17	6.95	N/A	N/A	N/A	N/A	355.33	345.92	7.04	1%
		350.32	6.96	N/A	N/A	N/A	N/A	358.10	349.10	7.09	2%
		352.84	7.00	N/A	N/A	N/A	N/A	357.76	348.71	7.08	1%

Results: Punctured and unpunctured samples were shown to be stable for 48 hours.
 AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not
Comments: stable; therefore the 72 hr data for amphetamine were not included. These data can be found in a supplemental
 Excel located in the Toxicology section shared electronic storage.

Acceptance Criteria: Average signal (peak area, or ratio of peak area analyte(s) compared to time 0 is within 20%.

Validation Study 9

Analyte: Amphetamine
 Units: ng/mL
 Instrument: LCMS-1

MATRIX EFFECTS

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Neat Response at Analyte RT		
	LQC	HQC
1	263201	2841207
2	243472	2860394
3	254761	2885998
4	258382	2862797
5	243900	2743578
6	245218	2879888
Mean	251489	2845644
SD	8445	52441
%CV	3%	2%

Neat Response at IS RT		
	LQC	HQC
	408972	433611
	378283	433593
	435809	441189
	422956	421527
	388212	448823
	399309	412403
Mean	405590	431858
SD	21510	13162
%CV	5%	3%

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at Analyte RT			% suppression/enhancement
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	241477	-4%	2557616	-10%	
5	231617	-8%	2823217	-1%	
7	229002	-9%	2796095	-2%	
8	242140	-4%	2676641	-6%	
9	206062	-18%	2639858	-7%	
10	236577	-6%	2524344	-11%	
11	245566	-2%	2679573	-6%	
12	248017	-1%	2754737	-3%	
13	234754	-7%	2679574	-6%	
14	244456	-3%	2891628	2%	
Mean	235967		2702328		
SD	12209		115569		
%CV	5%		4%		
Average % suppression/enhancement	-6%		-5%		

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at IS RT			% suppression/enhancement
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	426963	5%	397656	-8%	
5	407552	0%	446633	3%	
7	413441	2%	442097	2%	
8	436401	8%	422638	-2%	
9	365055	-10%	416398	-4%	
10	416819	3%	395819	-8%	
11	425983	5%	420782	-3%	
12	434760	7%	434337	1%	
13	414119	2%	414211	-4%	
14	433702	7%	455849	6%	
Mean	417480		424642		
SD	20968		20030		
%CV	5%		5%		
Average % suppression/enhancement	3%		-2%		

Results: Average percent suppression values for LQC and HQC were -6% and -5%, respectively.

Comments: N/A

Acceptance Criteria:

Average suppression/enhancement \leq 25% or the \leq %CV of the suppression/enhancement 20%

SUMMARY OF VALIDATION PERFORMANCE

Analyte: Amphetamine
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

The intent of this summary is to capture and document important information about the performance of this method outside the required measurements for validation.

Failed Runs (include dates/reasons):

Date	Reason
N/A	N/A

Deviations from SOP: N/A

Other Observations:

Working Standards Verified in Validation:
Calibrators: 200805C-C-10, 200805C-C-2.5, 200805C-C-0.2, 200923C-C-10, 200923C-C-2.5, 200923C-C-0.2
Controls: 200807L-Q-Mix, 200813L-Q-0.5, 200813L-St-5
Internal Standard: 200805C-IS-1, 200903C-IS-1, 200915C-IS-1

Sample Preparation Steps:

Refer to Toxicology Analytical Manual v3.5, "Amphetamines Confirmation by Liquid Chromatography-Tandem Mass Spectrometry" section titled "Extraction Procedure".

Location of Raw Data:

Toxicology section shared electronic storage.

Recommended Maximum Run Length (Unknown Samples):

30

Conclusion:

This method is fit for use on casework for amphetamine confirmation analysis in blood.

METHOD VALIDATION PROTOCOL AND RESULTS

Analyte: Methamphetamine
 Units: ng/mL
 Method: AMP.M
 Instrument: LCMS-1
 SOP Reference: Toxicology Analytical Manual v3.5

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

VALIDATION EXPERIMENT	SOP CRITERIA	RESULTS	COMMENTS	
1	Weight Verification	The least complex weighting scheme that minimizes $\sum %RE $	%RE = 130.99 (1/x)	1/x weighting will be used. %RE for unweighted and 1/x2 weighting are 235.48 and 131.99, respectively. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.
1	Validation Calibration	%RE Calibrators $\pm 20\%$ of target	Max %RE = 11.66	N/A
1	Case Work Calibration	%RE Calibrators $\pm 20\%$ of target	N/A	Not performed. All 7 calibrators used in validation will be used for casework.
1	Comparison of Validation Calibration to Casework Calibration	95% CI of slope includes 1 95% CI of intercept includes 0	N/A	Not performed. All 7 calibrators used in validation will be used for casework.
2	Limit of Detection (LOD)	Signal to Noise ≥ 3.3 Acceptable detection and identification criteria	LOD = 10 ng/mL S:N= 1373	The lowest non-zero calibrator was defined as the LOD.
3	Limit of Quantitation (LOQ)	Bias: $\leq 20\%$ Within-Run Precision: CV $\leq 20\%$ Between-Run Precision: CV $\leq 20\%$	Bias = 5.43% Within-Run Precision = 21.57% Between-Run Precision = 17.25%	Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Numbers in red were $> 20\%$ target value. Numbers with * were $> 20\%$ target value due to insufficient IS. Max within-run precision is greater than 20% due to the values of 15.15 ng/mL and 18.38 ng/mL being greater than 20% of the target value in AMP_20200901B_CLR and AMP_202008014B_SD respectively. LOQ data are deemed acceptable because a minimum of three values passing acceptance criteria were obtained for each source used over the four runs as defined in the SOP.
4	Bias & Precision	%Bias $\leq 20\%$ Within-Run %CV $\leq 20\%$ Between-Run %CV $\leq 20\%$	Max Bias = -6.64% Max Within-Run Precision = 18.28% Max Between-Run Precision = 17.09%	N/A
5	Carryover	No analyte carryover is observed in the matrix blank samples; response in blank samples is $< LOQ$ of the method.	No significant carryover observed following samples containing analyte at up to 2000 ng/mL.	N/A
6	Matrix Interference	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.
6	Interference from stable isotope internal standard	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	
6	Exogenous Substances Interferences	Concentrations of analytes of interest within $\pm 20\%$ of the average concentration obtained in the Bias and Precision studies	No significant interference observed.	
7	Dilution Integrity	Average %Bias must be less than 20%	It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.	N/A
8	Processed Sample Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	N/A	Study not performed. Sample preparation will be completed once started without prolonged interruptions.
8	Autosampler Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	Punctured and unpunctured samples were shown to be stable for 48 hours.	Cal 7 relative response was beyond the range of the calibration curve and therefore was not quantified. The samples were deemed stable at 48 hrs since response ratio of analyte to IS was within 20% compared to time 0. AMP_20200904B_SD showed the punctured and unpunctured samples at 72 hr were not stable. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.
9	Matrix Effects	Average suppression/enhancement $\leq 25\%$ or the $\leq \%CV$ of the suppression/enhancement 20%	Average percent suppression values for LQC and HQC were -11% and -8%, respectively.	N/A

Validation Study 1

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1



LINEARITY
 Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Study Date	Target (x)	Target (x) ²	Calculated (y)	Predicted	Residual	%RE
AMP_20200814B_PK	10	100	10.53	12.52	-1.99	5.31
	20	400	20.76	22.14	-1.38	3.80
	50	2500	44.41	51.04	-6.63	11.17
	100	10000	97.78	99.38	-1.61	2.22
	250	62500	266.82	245.73	21.09	6.73
	500	250000	486.15	494.01	-7.86	2.77
	1000	1000000	1004.44	1007.00	-2.56	0.44
AMP_20200828B_PK	10	100	9.56	12.52	-2.96	4.37
	20	400	20.71	22.14	-1.43	3.55
	50	2500	50.05	51.04	-0.99	0.10
	100	10000	99.08	99.38	-0.30	0.92
	250	62500	260.98	245.73	15.25	4.39
	500	250000	482.13	494.01	-11.88	3.57
	1000	1000000	1009.04	1007.00	2.04	0.90
AMP_20200828B_SD	10	100	9.24	12.52	-3.28	7.58
	20	400	20.45	22.14	-1.69	2.23
	50	2500	53.83	51.04	2.80	7.67
	100	10000	99.19	99.38	-0.20	0.81
	250	62500	245.62	245.73	-0.11	1.75
	500	250000	500.71	494.01	6.69	0.14
	1000	1000000	1001.58	1007.00	-5.41	0.16
AMP_20200901B_PK	10	100	9.89	12.52	-2.63	1.13
	20	400	20.48	22.14	-1.65	2.41
	50	2500	46.62	51.04	-4.41	6.75
	100	10000	104.84	99.38	5.45	4.84
	250.0	62500	259.08	245.73	13.35	3.63
	500	250000	480.41	494.01	-13.61	3.92
	1000	1000000	1010.75	1007.00	3.75	1.07
AMP_20200901B_CLR	10	100	8.83	12.52	-3.69	11.66
	20	400	21.48	22.14	-0.66	1.41
	50	2500	49.06	51.04	-1.98	1.87
	100	10000	108.14	99.38	8.76	8.14
	250	62500	253.17	245.73	7.44	1.27
	500	250000	476.79	494.01	-17.23	4.64
	1000	1000000	1016.49	1007.00	9.50	1.65

Regression Statistics	
Multiple R	0.999746586
R Square	0.999493236
Adjusted R Square	0.999461563
Standard Error	7.953397165
Observations	35

ANOVA					
	df	SS	MS	F	Significance F
Regression	2	3992357.738	1996178.869	31556.88402	1.89187E-53
Residual	32	2024.208847	63.25652646		
Total	34	3994381.947			

	Coefficients	Standard Error	t Stat	P-value	Lower 95%	Upper 95%	Lower 95.0%	Upper 95.0%
Intercept	2.913355172	2.117251758	1.376007912	0.17837367	-1.399345531	7.226055875	-1.399345531	7.226055875
target x	0.9603227	0.015249158	62.9754582	3.98739E-35	0.929261182	0.991384218	0.929261182	0.991384218
x^2	4.37607E-05	1.50407E-05	2.909494521	0.006535296	1.31239E-05	7.43975E-05	1.31239E-05	7.43975E-05



Σ|%RE| 130.99
 Max %RE 11.66

Results: The reported %RE is for 1/x weighting.

Comments: 1/x weighting will be used. %RE for unweighted and 1/x² weighting are 235.48 and 131.99, respectively. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.

Acceptance Criteria:

The least complex weighting scheme that minimizes Σ|%RE|
 %RE Calibrators ≤20% of target

Validation Study 2

Sensitivity (LOD)

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Concentration (ng/mL)	Signal to Noise Ratio			
	10			
AMP_20200814B_SD	1812			
	1673			
	1245			
	2995			
	3785			
	2549			
	3191			
	704			
AMP_20200828_PK	2057			
	1276			
	765			
	1453			
	381			
	341			
	981			
	684			
AMP_20200901_PK	1223			
	1392			
	803			
	1326			
	206			
	1546			
	843			
	1640			
AMP_20200901_CLR	1767			
	858			
	1745			
	897			
	1369			
	2178			
	991			
	789			
1618				
1700				
328				
305				
Average Signal to Noise:	1373			

Established LOD: 10 ng/mL
Signal to Noise: 1373

Comments: The lowest non-zero calibrator was defined as the LOD.

Acceptance Criteria: Signal to Noise ≥ 3.3
 Acceptable detection and identification criteria

Validation Study 3 **SENSITIVITY (LOQ)**

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LOQ	
<i>Target Concentration (ng/mL):</i>		10	
AMP_202008014B_SD	1-1	10.56	
	1-2	10.82	
	1-3	12.49	
	1-4	11.08	
	1-5	12.27	
	1-6	11.33	
	1-7	11.17	
	1-8	10.56	
	1-9	18.38	
	Within Run	Mean	12.07
		SD	2.46
		%CV	20.39%
		% Bias	20.74%
AMP_20200828_PK	2-1	9.83	
	2-2	10.22	
	2-3	9.59	
	2-4	9.88	
	2-5	3.86*	
	2-6	9.85	
	2-7	10.30	
	2-8	9.81	
	2-9	10.19	
	Within Run	Mean	9.96
	SD	0.25	
	%CV	2.50%	
	% Bias	-0.41%	
AMP_20200901B_PK	3-1	9.83	
	3-2	9.78	
	3-3	9.96	
	3-4	10.22	
	3-5	10.29	
	3-6	10.34	
	3-7	10.37	
	3-8	10.18	
	3-9	9.90	
	Within Run	Mean	10.10
	SD	0.23	
	%CV	2.27%	
	% Bias	0.96%	
AMP_20200901B_CLR	4-1	58.58*	
	4-2	8.79	
	4-3	9.51	
	4-4	9.47	
	4-5	8.05	
	4-6	10.16	
	4-7	15.15	
	4-8	9.95	
	4-9	9.26	
	Within Run	Mean	10.04
	SD	2.17	
	%CV	21.57%	
	% Bias	0.41%	

Mean		10.57
SD		1.82
Precision (%CV)	Max Within-Run	21.57%
	Between-Run	17.25%
% Bias		5.43%

Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Numbers in red were > 20% target value. Numbers with * were > 20% target value due to insufficient IS. Max within-run precision is greater than 20% due to the values of 15.15 ng/mL and 18.38 ng/mL being greater than 20% of the target value in AMP_20200901B_CLR and AMP_202008014B_SD respectively. LOQ data are deemed acceptable because a minimum of three values passing acceptance criteria were obtained for each source used over the four runs as defined in the SOP.

Acceptance Criteria: Bias: ≤20%
Within-Run Precision: CV ≤20%
Between-Run Precision: CV ≤20%

Validation Study 4

BIAS AND PRECISION

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LQC	MQC	UTAK	HQC	
<i>Target Concentration (ng/mL):</i>		25	100	84.03	800	
AMP_202008014B_SD	1-1	25.20	94.11	66.64	NA	
	1-2	23.42	89.34	72.17	NA	
	1-3	24.54	90.71	67.83	NA	
	1-4			78.13		
	Within Run	Mean	24.38	91.39	71.19	
		SD	0.90	2.46	5.20	
		%CV	3.69%	2.69%	7.30%	
	% Bias	-2.47%	-8.61%	-15.28%		
AMP_202008014B_PK	2-1	25.22	82.88	104.41	889.93	
	2-2	29.26	84.44	102.64	842.23	
	2-3	28.41	91.83	108.70	816.73	
	2-4			110.71		
	Within Run	Mean	27.63	86.39	106.62	849.63
		SD	2.13	4.78	3.73	37.15
		%CV	7.72%	5.54%	3.50%	4.37%
	% Bias	10.52%	-13.61%	26.88%	6.20%	
AMP_20200828B_PK	3-1	26.86	NA	82.66	757.03	
	3-2	29.00	NA	78.04	768.73	
	3-3	25.95	NA	76.71	771.34	
	3-4			77.58		
	Within Run	Mean	27.27		78.75	765.70
		SD	1.57		2.67	7.62
		%CV	5.75%		3.39%	1.00%
	% Bias	9.08%		-6.29%	-4.29%	
AMP_20200828B_SD	4-1	23.30	103.46	81.53	806.91	
	4-2	26.00	99.80	84.69	775.76	
	4-3	25.04	101.17	74.99	815.42	
	4-4			77.06		
	Within Run	Mean	24.78	101.48	79.57	799.36
		SD	1.37	1.85	4.37	20.88
		%CV	5.52%	1.82%	5.49%	2.61%
	% Bias	-0.88%	1.48%	-5.31%	-0.08%	
AMP_20200901B_PK	5-1	24.95	83.71		779.76	
	5-2	25.24	91.55		766.31	
	5-3	25.42	87.62		801.08	
	Within Run	Mean	25.20	87.63		782.38
		SD	0.24	3.92		17.53
		%CV	0.94%	4.47%		2.24%
		% Bias	0.81%	-12.37%		-2.20%
AMP_20200901B_CLR	6-1	27.84	118.49		758.56	
	6-2	29.55	81.98		794.22	
	6-3	26.01	99.24		778.26	
	Within Run	Mean	27.80	99.90		777.01
		SD	1.77	18.26		17.86
		%CV	6.38%	18.28%		2.30%
		% Bias	11.20%	-0.10%		-2.87%
Mean		26.18	93.36	84.03	794.82	
SD		1.91	9.80	14.36	35.89	
Precision (%CV)	Max Within-Run	7.72%	18.28%	7.30%	4.37%	
	Between-Run	7.29%	10.50%	17.09%	4.52%	
% Bias		4.71%	-6.64%	0.00%	-0.65%	

Comments: N/A

Acceptance Criteria: %Bias ≤20%
 Within-Run %CV ≤20%
 Between-Run %CV ≤20%

Validation Study 5**CARRYOVER**

Analyte: Methamphetamine
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Average LOQ Response*: 77171

Study Date:	Response		
	AMP_202008014B_SD	AMP_202008014B_PK	AMP_20200828B_PK
Concentrated Sample (2000 ng/mL)	6244867	6192835	8224432
Blank	4855	4455	2071
%LOD Response	6.29%	5.77%	2.68%

Maximum Response in Blank: **6.3%**

Results: No significant carryover observed following samples containing analyte at up to 2000 ng/mL.

Comments: N/A

Acceptance Criteria: No analyte carryover is observed in the matrix blank samples; response in blank samples is <LOQ of the method.

Validation Study 6

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

MATRIX & IS INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

	LOQ Response	
	Analyte	IS
Run 1	65958	270023
Run 2	88777	300000
Run 3	86390	372619
Run 4	55281	242518
Run 5	89447	327379
Average	77171	302508

Matrix Interference

Study Date: AMP_20200814B_PK

Matrix Source	Peak at Analyte RT	
	Peak Response	Percent of LOQ Response
4	2497	3.24%
5	3193	4.14%
7	4038	5.23%
8	2723	3.53%
9	3404	4.41%
10	2909	3.77%
11	3246	4.21%
12	2972	3.85%
13	3924	5.08%
14	4387	5.68%

Interference from Stable-Isotope Internal Standards

Study Date: AMP_20200828B_SD

Experiment	Peak at Analyte RT		Peak at IS RT	
	Peak Response	Percent of LOQ Response	Peak Response	Percent of LOQ Response
Matrix with IS but no D0 (IS = 50 ng/mL)	2095	2.71%	N/A	N/A
Matrix with D0 but no IS (D0 = 2000 ng/mL)	N/A	N/A	139	0.05%

Matrix Interference: No significant interference observed.

IS Interference: No significant interference observed.

Comments: LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.

Acceptance Criteria:

Response of blank matrix is less than 20% the average response of LOQ

Validation Study 6

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

EXOGENOUS SUBSTANCE INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target LQC Concentration (µg/mL): 26.18
 Control Acceptance: 20%
 Run Date: AMP 20200828B SD

Group	Compound	Compound Concentration (µg/mL)	Calculated LQC Concentration (ng/mL)	% Difference from Target	Comment
Benzodiazepines	Alprazolam	1	25.01	-4%	No significant interference
	α-Hydroxyalprazolam				
	Clonazepam				
	7-Aminoclonazepam				
	Diazepam				
	Nordiazepam				
	Temazepam				
	Lorazepam				
	Oxazepam				
Zolpidem					
Phencyclidine	Phencyclidine	1	27.06	3%	No significant interference
Opioids	Morphine	2	25.61	-2%	No significant interference
	Hydrocodone	1			
	Buprenorphine	0.5			
	Norbuprenorphine				
	Codeine				
	Fentanyl				
	Norfentanyl oxalate				
	Hydromorphone				
	Metadone				
	EDDP				
	Oxycodone				
	Oxymorphone				
	Tramadol				
	o-Desmethyltramadol				
Cocaine and Metabolites	Benzoylcegonine		2	26.10	0%
	Cocaine	0.5			
	Cocacethylene				
Carisoprodol/Meprobamate	Carisoprodol	1	25.94	-1%	No significant interference
	Meprobamate				
Basic and Neutral Mix	Amitriptyline	5	25.77	-2%	No significant interference
	Benzylpiperazine				
	Chlorpheniramine				
	Cyclobenzaprine				
	Dextromethorphan				
	Diphenhydramine				
	Doxylamine				
	Fluoxetine				
	Imipramine				
	Ketamine				
	Norketamine				
	Meperidine				
	Nortriptyline				
	Propoxyphene				
	Sertraline				
	Trazodone				
	Venlafaxine				
	Zopiclone				
Over-the-Counter Drugs	Acetaminophen	10	24.00	-8%	No significant interference
	Caffeine				
	Ibuprofen				
	Naproxen				
	Pseudoephedrine	1			
Theobromine	1				
Cannabinoids	Δ9-THC	0.5	25.93	-1%	No significant interference
	11-Hydroxy-Δ9-THC				
	Δ9-THC-COOH				
	Δ8-THC				
	Cannabinol				
Cannabinolic acid					

Conclusions: No significant interference observed.

Comments: N/A

Acceptance Criteria: Concentrations of analytes of interest within ±20% of the average concentration obtained in the Bias and Precision studies

Validation Study 7

DILUTION INTEGRITY

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target Concentration (ng/mL): 359.36

Run Date	Dilution Factor	Diluted Result (ng/mL)	Calculated Result (µg/mL)	Mean	SD	%Bias	Absolute % Bias	Within-Run Precision (per Dilution)
AMP_202008014B_SD	2	173.96	347.92	334.63	20.02	-3.18	3.18	5.98
	2	172.19	344.39			-4.17	4.17	
	2	155.80	311.60			-13.29	13.29	
	5	68.41	342.05	326.11	19.01	-4.82	4.82	5.83
	5	66.24	331.21			-7.83	7.83	
	5	61.01	305.07			-15.11	15.11	
	10	35.64	356.45	331.55	30.01	-0.81	0.81	9.05
	10	34.00	339.96			-5.40	5.40	
	10	29.82	298.23			-17.01	17.01	
	AMP_20200828B_PK	2	188.76	377.52	402.21	22.52	5.05	5.05
2		203.75	407.50	13.40			13.40	
2		210.81	421.62	17.33			17.33	
5		75.64	378.21	407.81	26.47	5.24	5.24	6.49
5		85.83	429.17			19.43	19.43	
5		83.21	416.06			15.78	15.78	
10		36.26	362.60	402.48	34.54	0.90	0.90	8.58
10		42.28	422.83			17.66	17.66	
10		42.20	422.02			17.44	17.44	
AMP_20200828B_SD		2	184.47	368.93	367.95	8.90	2.66	2.66
	2	188.16	376.32	4.72			4.72	
	2	179.30	358.60	-0.21			0.21	
	5	75.14	375.68	374.16	14.25	4.54	4.54	3.81
	5	77.52	387.60			7.86	7.86	
	5	71.84	359.21			-0.04	0.04	
	10	32.23	322.25	352.71	26.49	-10.33	10.33	7.51
	10	36.55	365.47			1.70	1.70	
	10	37.04	370.40			3.07	3.07	
	AMP_20200901B_PK	2	197.77	395.53	409.19	12.16	10.07	10.07
2		209.41	418.83	16.55			16.55	
2		206.61	413.22	14.99			14.99	
5		74.36	371.82	402.81	27.22	3.47	3.47	6.76
5		82.75	413.77			15.14	15.14	
5		84.57	422.83			17.66	17.66	
10		40.81	408.11	394.19	25.89	13.57	13.57	6.57
10		36.43	364.32			1.38	1.38	
10		41.02	410.15			14.13	14.13	
AMP_20200901B_CLR		2	155.15	310.30	354.05	40.75	-13.65	13.65
	2	195.47	390.93	8.79			8.79	
	2	180.46	360.91	0.43			0.43	
	5	76.84	384.22	395.01	11.25	6.92	6.92	2.85
	5	78.83	394.15			9.68	9.68	
	5	81.33	406.67			13.17	13.17	
	10	37.90	379.04	367.68	9.91	5.48	5.48	2.70
	10	36.08	360.79			0.40	0.40	
	10	36.32	363.22			1.07	1.07	

Dilution Factor	2
Avg %Bias	8.57
Max Within-Run Precision (%CV)	11.51

Dilution Factor	5
Avg %Bias	9.78
Max Within-Run Precision (%CV)	6.76

Dilution Factor	10
Avg %Bias	7.36
Max Within-Run Precision (%CV)	9.05

Results: It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.
 Comments: N/A

Acceptance Criteria:

Avg %Bias must be less than 20%

Validation Study 8Analyte: Methamphetamine
Units: ng/mL
Instrument: LCMS-1**PROCESSED SAMPLE STABILITY**Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Run Date	Sample	Time Zero Response	24H Response	48H Response	72H Response	Acceptable Range		Pass/Unstable
						Low	High	
	LQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			
	HQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			

Results:**Comments:** Study not performed. Sample preparation will be completed once started without prolonged interruptions.**Acceptance Criteria:**

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 8

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

AUTOSAMPLER STABILITY

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	Calibrator 1	9.14	0.28	8.97	8.61	0.28	-1%
	Calibrator 2	19.15	0.51	19.28	18.93	0.51	1%
	Calibrator 3	55.63	1.33	55.12	54.28	1.32	-1%
	Calibrator 4	104.71	2.39	105.17	103.60	2.40	0%
	Calibrator 5	248.82	5.25	247.96	242.65	5.24	0%
	Calibrator 6	477.88	8.99	474.66	456.28	8.94	-1%
	Calibrator 7	1037.02	13.93	1084.70	893.14	14.08	1%

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours			48 Hours				
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20201019B_SD & AMP_20201021B_SD	Calibrator 1	9.86	0.30	N/A	N/A	N/A	N/A	9.79	8.80	0.30	-1%
	Calibrator 2	19.62	0.52	N/A	N/A	N/A	N/A	19.69	18.44	0.52	0%
	Calibrator 3	49.51	1.18	N/A	N/A	N/A	N/A	48.85	46.86	1.17	-1%
	Calibrator 4	105.60	2.38	N/A	N/A	N/A	N/A	106.44	102.97	2.40	1%
	Calibrator 5	252.11	5.24	N/A	N/A	N/A	N/A	257.46	250.01	5.34	2%
	Calibrator 6	483.64	8.96	N/A	N/A	N/A	N/A	499.49	485.29	9.17	2%
	Calibrator 7	1022.04	13.77	N/A	N/A	N/A	N/A	N/A*	N/A*	14.55	6%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	LQC	25.51	0.66	25.69	25.17	0.66	1%
		25.57	0.66	25.18	24.67	0.65	-1%
		25.48	0.65	25.74	25.23	0.66	1%
	HQC	771.64	12.32	781.48	717.47	12.41	1%
		771.52	12.32	791.01	724.69	12.49	1%
		784.15	12.43	789.84	723.81	12.48	0%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20201019B_SD & AMP_20201021B_SD	LQC	25.53	0.65	N/A	N/A	N/A	N/A	25.43	24.04	0.65	0%
		25.38	0.65	N/A	N/A	N/A	N/A	25.55	24.16	0.65	1%
		25.48	0.65	N/A	N/A	N/A	N/A	25.59	24.20	0.65	0%
	HQC	771.25	12.19	N/A	N/A	N/A	N/A	818.65	793.63	12.58	3%
		773.44	12.21	N/A	N/A	N/A	N/A	815.76	790.86	12.56	3%
		760.24	12.10	N/A	N/A	N/A	N/A	817.24	792.28	12.57	4%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	LQC	25.51	0.66	25.71	25.20	0.66	1%
		25.57	0.66	25.47	24.95	0.65	0%
		25.48	0.65	25.59	25.07	0.66	0%
	HQC	771.64	12.32	794.41	727.24	12.51	2%
		771.52	12.32	772.44	710.96	12.33	0%
		784.15	12.43	772.80	710.84	12.33	-1%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20201019B_SD & AMP_20201021B_SD	LQC	25.53	0.65	N/A	N/A	N/A	N/A	25.42	24.03	0.65	0%
		25.38	0.65	N/A	N/A	N/A	N/A	25.53	24.14	0.65	1%
		25.48	0.65	N/A	N/A	N/A	N/A	25.56	24.17	0.65	0%
	HQC	771.25	12.19	N/A	N/A	N/A	N/A	825.83	800.51	12.63	4%
		773.44	12.21	N/A	N/A	N/A	N/A	834.36	808.68	12.70	4%
		760.24	12.10	N/A	N/A	N/A	N/A	819.64	794.58	12.59	4%

Results: Punctured and unpunctured samples were shown to be stable for 48 hours.

Comments: Cal 7 relative response was beyond the range of the calibration curve and therefore was not quantified. The samples were deemed stable at 48 hrs since response ratio of analyte to IS was within 20% compared to time 0. AMP_20200904B_SD showed the punctured and unpunctured samples at 12 hr were not stable. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.

Acceptance Criteria:

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 9

Analyte: Methamphetamine
 Units: ng/mL
 Instrument: LCMS-1

MATRIX EFFECTS

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Neat Response at Analyte RT		
	LQC	HQC
1	186790	3120037
2	175901	3140257
3	183088	3149629
4	186966	3139183
5	173831	3023923
6	174046	3166044
Mean	180104	3123179
SD	6236	50885
%CV	3%	2%

Neat Response at IS RT		
	LQC	HQC
	280576	300386
	257202	301385
	303981	307923
	293909	297130
	265101	311246
	272994	280962
Mean	278961	299839
SD	17625	10600
%CV	6%	4%

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at Analyte RT			% suppression/enhancement
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	168063	-7%	2829211	-9%	
5	147048	-18%	3000414	-4%	
7	146034	-19%	2944539	-6%	
8	163598	-9%	2804180	-10%	
9	144713	-20%	2851479	-9%	
10	159546	-11%	2600611	-17%	
11	167479	-7%	2891832	-7%	
12	183269	2%	2897534	-7%	
13	151469	-16%	2836859	-9%	
14	167610	-7%	3093322	-1%	
Mean	159883		2874998		
SD	12484		130567		
%CV	8%		5%		
Average % suppression/enhancement	-11%		-8%		

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at IS RT			% suppression/enhancement
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	284611	2%	274368	-8%	
5	256822	-8%	302497	1%	
7	256976	-8%	298344	0%	
8	279249	0%	276282	-8%	
9	248414	-11%	285105	-5%	
10	273596	-2%	251192	-16%	
11	280138	0%	288037	-4%	
12	296923	6%	297712	-1%	
13	256768	-8%	278564	-7%	
14	286837	3%	313916	5%	
Mean	272033		286602		
SD	16214		17761		
%CV	6%		6%		
Average % suppression/enhancement	-2%		-4%		

Results: Average percent suppression values for LQC and HQC were -11% and -8%, respectively.
Comments: N/A

Acceptance Criteria: Average suppression/enhancement ≤25% or the ≤%CV of the suppression/enhancement 20%

SUMMARY OF VALIDATION PERFORMANCE

Analyte: Methamphetamine
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

The intent of this summary is to capture and document important information about the performance of this method outside the required measurements for validation.

Failed Runs (include dates/reasons):

Date	Reason
N/A	N/A

Deviations from SOP: N/A

Other Observations: Working Standards Verified in Validation:
Calibrators: 200805C-C-10, 200805C-C-2.5, 200805C-C-0.2, 200923C-C-10, 200923C-C-2.5, 200923C-C-0.2
Controls: 200807L-Q-Mix, 200813L-Q-0.5, 200813L-St-5
Internal Standard: 200805C-IS-1, 200903C-IS-1, 200915C-IS-1

Sample Preparation Steps: Refer to Toxicology Analytical Manual v3.5, "Amphetamines Confirmation by Liquid Chromatography-Tandem Mass Spectrometry" section titled "Extraction Procedure".

Location of Raw Data: Toxicology section shared electronic storage.

Recommended Maximum Run Length (Unknown Samples): 30

Conclusion: This method is fit for use on casework for methamphetamine confirmation analysis in blood.

METHOD VALIDATION PROTOCOL AND RESULTS

Analyte: MDA
 Units: ng/mL
 Method: AMP.M
 Instrument: LCMS-1
 SOP Reference: Toxicology Analytical Manual v3.5

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

VALIDATION EXPERIMENT		SOP CRITERIA	RESULTS	COMMENTS
1	Weight Verification	The least complex weighting scheme that minimizes $\sum %RE $	Unweighted: = 560.76 1/x Weighting: = 332.63 1/x2 Weighting: = 329.77	Data were processed using 1/x2 weighting.
1	Linearity	95% CI of slope includes 1 95% CI of intercept includes 0	95% CI of slope = 1.0477 - 1.0937 95% CI of Intercept = -11.0805 - -0.3719	The linearity of the method is acceptable because the individual calibration curves on all five days met the acceptance criteria. Also, other analytical data including the ion ratios and quantification values of the controls and calibrators met acceptance criteria. □
1	Validation Calibration	%RE Calibrators $\pm 20\%$ of target	Max %RE = 15.38	N/A
1	Case Work Calibration	%RE Calibrators $\pm 20\%$ of target	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
1	Comparison of Validation Calibration to Casework Calibration	95% CI of slope includes 1 95% CI of intercept includes 0	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
2	Limit of Detection (LOD)	Signal to Noise ≥ 3.3 Acceptable detection and identification criteria	LOD = 10 ng/mL S:N= 2622	Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Blood source 5 had an interference with the 77.1 m/z qualifier ion. This ion was removed from the acquisition method on 9/10/2020.
3	Limit of Quantitation (LOQ)	Bias: $\leq 20\%$ Within-Run Precision: CV $\leq 20\%$ Between-Run Precision: CV $\leq 20\%$	Bias = 4.62% Within-Run Precision = 10.05% Between-Run Precision = 5.64%	Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Blood source 5 had an interference with the 77.1 m/z qualifier ion. This ion was removed from the acquisition method on 9/10/2020. Number with * was $> 20\%$ target value due to insufficient IS. Number in red was $> 20\%$ target value.
4	Bias & Precision	%Bias $\leq 20\%$ Within-Run %CV $\leq 20\%$ Between-Run %CV $\leq 20\%$	Max Bias = -10.09% Max Within-Run Precision = 18.75% Max Between-Run Precision = 16.27%	Numbers in red were not within 20% of the target value.
5	Carryover	No analyte carryover is observed in the matrix blank samples; response in blank samples is $< LOQ$ of the method.	No significant carryover observed following samples containing analyte at up to 2000 ng/mL.	N/A
6	Matrix Interference	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.
6	Interference from stable isotope internal standard	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	
6	Exogenous Substances Interferences	Concentrations of analytes of interest within $\pm 20\%$ of the average concentration obtained in the Bias and Precision studies	No significant interference observed.	N/A
7	Dilution Integrity	Average %Bias must be less than 20%	It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.	N/A
8	Processed Sample Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	N/A	Study not performed. Sample preparation will be completed once started without prolonged interruptions.
8	Autosampler Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	Punctured and unpunctured samples were shown to be stable for 48 hours.	AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not stable; therefore the 72 hr data for MDA were not included. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.
9	Matrix Effects	Average suppression/enhancement $\leq 25\%$ or the $\leq CV$ of the suppression/enhancement 20%	Average percent suppression values for LQC and HQC were -8% and -4%, respectively.	N/A
10	Additional Experiment	%RE Calibrators $\pm 20\%$ of target Bias: $\leq 20\%$ Within-Run Precision: CV $\leq 20\%$	All acceptance criteria were met.	The purpose of this experiment was to verify the change in qualifier ions from 105.1 m/z and 77.1 m/z to 133.1 m/z and 79.1 m/z. Blank blood sources 4 (lot 310355), 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study.

Validation Study 1

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

STANDARD CURVE WEIGHT VERIFICATION

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Unweighted: 560.76
 1/x Weighting: 332.63
 1/x² Weighting: 329.77

	C _{nom}	y	w	wxy	wx	wy	wx ²	wy ²	C _{found}	%RE	%RE
AMP_20200814B_SD	10	0.24036	1	2.40355	10	0.240355	100	0.057771	14.89155	48.9155	48.9155
	20	0.47392	1	9.4784	20	0.47392	400	0.2246	24.5532	22.76602	22.76602
	50	1.22437	1	61.21835	50	1.224367	2500	1.499075	55.59621	11.19242	11.19242
	100	2.58049	1	258.0485	100	2.580485	10000	6.658903	111.6934	11.69341	11.69341
	250	6.06747	1	1516.868	250	6.067473	62500	36.81423	255.9362	2.374489	2.374489
	500	13.39617	1	6698.085	500	13.39617	250000	179.4574	559.0952	11.81905	11.81905
AMP_20200814B_PK	10	0.21536	1	2.153614	10	0.21536141	100	0.046381	13.85767	38.57665	38.57665
	20	0.39552	1	7.910335	20	0.39551676	400	0.156434	21.30997	6.549872	6.549872
	50	0.80632	1	40.31583	50	0.80631669	2500	0.650147	38.30313	-23.3937	23.39374
	100	1.72844	1	172.8443	100	1.72844332	10000	2.987516	76.44784	-23.5522	23.55216
	250	5.00909	1	1252.272	250	5.00908786	62500	25.09096	212.155	-15.138	15.13798
	500	10.34035	1	5170.173	500	10.3403466	250000	106.9228	432.688	-13.4624	13.4624
AMP_20200828B_SD	10	0.25424	1	2.54236	10	0.254236	100	0.064636	15.46575	54.65752	54.65752
	20	0.48059	1	9.6118	20	0.48059	400	0.230967	24.82912	24.14558	24.14558
	50	1.22095	1	61.04725	50	1.220945	2500	1.490707	55.45465	-10.90931	10.90931
	100	2.27145	1	227.1448	100	2.271448	10000	5.159476	98.90978	-1.09022	1.090219
	250	5.79295	1	1448.238	250	5.79295	62500	33.55827	244.5803	-2.16788	2.167881
	500	13.02508	1	6512.54	500	13.02508	250000	169.6527	543.7447	8.748944	8.748944
AMP_20200828B_PK	10	0.21531	1	2.153097	10	0.21530971	100	0.046358	13.85553	38.55526	38.55526
	20	0.41191	1	8.238193	20	0.41190967	400	0.16967	21.98808	9.940421	9.940421
	50	0.92830	1	46.41484	50	0.92829689	2500	0.861735	43.34897	-13.3021	13.30207
	100	1.82790	1	182.7904	100	1.82790406	10000	3.341233	80.56213	-19.4379	19.43787
	250	5.16230	1	1290.575	250	5.16229927	62500	26.64933	218.4928	-12.6029	12.60288
	500	10.53041	1	5265.204	500	10.5304083	250000	110.8895	440.5501	-11.89	11.88988
AMP_20200901B_CLR	10	0.24454	1	2.445443	10	0.2445443	100	0.059802	15.06485	50.64845	50.64845
	20	0.51138	1	10.22766	20	0.51136323	400	0.261513	26.10291	30.51455	30.51455
	50	1.19108	1	59.55381	50	1.19107613	2500	1.418662	54.2191	8.438197	8.438197
	100	2.70390	1	270.3896	100	2.70389566	10000	7.311052	116.7984	16.79842	16.79842
	250	6.35689	1	1589.222	250	6.35688939	62500	40.41004	267.9082	7.16329	7.16329
	500	13.21491	1	6607.454	500	13.2149071	250000	174.6338	551.5971	10.31942	10.31942
Sum	4650.0	108.8219	5	38787.56	4650	108.821915	1627500	936.7756			560.764
Slope											
Intercept											
R ²											

Comments: Data were processed using 1/x² weighting.

Acceptance Criteria: The least complex weighting scheme that minimizes Σ|%RE|

Validation Study 1

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

VALIDATION CURVE CALIBRATION


Analyte: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Batch Name	Target	Calculated Result	% RE	y (Response Ratio)
AMP_20200814B_SD	10	10.19	1.85	0.240
	20	19.37	3.14	0.474
	50	48.89	2.22	1.224
	100	102.23	2.23	2.580
	250	239.38	4.25	6.067
	500	527.64	5.53	13.396
AMP_20200814B_PK	10	10.31	3.13	0.215
	20	20.07	0.34	0.396
	50	42.31	15.38	0.806
	100	92.24	7.76	1.728
	250	269.88	7.95	5.009
	500	558.56	11.71	10.340
AMP_20200828B_PK	10	10.14	1.38	0.215
	20	20.17	0.85	0.412
	50	46.52	6.96	0.928
	100	92.42	7.58	1.828
	250	262.56	5.02	5.162
	500	536.46	7.29	10.530
AMP_20200828B_SD	10	10.13	1.35	0.254
	20	19.59	2.06	0.481
	50	50.51	1.02	1.221
	100	94.39	5.61	2.271
	250	241.47	3.41	5.793
	500	543.54	8.71	13.025
AMP_20200901B_CLR	10	10.02	0.16	0.2445
	20	20.31	1.54	0.5114
	50	46.53	6.95	1.1911
	100	104.88	4.88	2.7039
	250	245.78	1.69	6.3569
	500	510.31	2.06	13.2149


Max %RE = 15.38




Slope 0.02663
 Intercept -0.13003
 R² 0.99893




Slope 0.02084
 Intercept -0.14735
 R² 0.99938



Slope 0.02117
 Intercept -0.10148
 R² 0.99964



Slope 0.02586
 Intercept -0.16691
 R² 0.99836



Slope 0.02638
 Intercept -0.05243
 R² 0.99975

Comments: N/A

Acceptance Criteria:

%RE Calibrators ±20% of target

Validation Study 1

LINEARITY

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1



Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Study Date	Target (x)	Calculated (y)	Predicted	Residual
AMP_20200814B_SD	10	10.19	4.98	5.20
	20	19.37	15.69	3.68
	50	48.89	47.81	1.08
	100	102.23	101.34	0.89
	250	239.38	261.95	-22.57
	500	527.64	529.62	-1.98
AMP_20200814B_PK	10	10.31	4.98	5.33
	20	20.07	15.69	4.38
	50	42.31	47.81	-5.50
	100	92.24	101.34	-9.10
	250	269.88	261.95	7.93
	500	558.56	529.62	28.94
AMP_20200828B_PK	10	10.14	4.98	5.16
	20	20.17	15.69	4.48
	50	46.52	47.81	-1.29
	100	92.42	101.34	-8.92
	250	262.56	261.95	0.61
	500	536.46	529.62	6.84
AMP_20200828B_SD	10	10.13	4.98	5.15
	20	19.59	15.69	3.90
	50	50.51	47.81	2.70
	100	94.39	101.34	-6.96
	250	241.47	261.95	-20.48
	500	543.54	529.62	13.92
AMP_20200901B_CLR	10	10.02	4.98	5.03
	20	20.31	15.69	4.62
	50	46.53	47.81	-1.28
	100	104.88	101.34	3.53
	250	245.78	261.95	-16.17
	500	510.31	529.62	-19.32

Slope	1.0707
Std err in slope, S_b	0.0112
Degrees freedom	28
Confidence level	95%
Student t	2.0484
Confidence interval	0.023
Slope	1.0707 ± 0.023
Range	1.0477 - 1.0937

Intercept	-5.7262
Std err in Intercept	2.6139
Degrees freedom	28
Confidence Level	95%
Student t	2.0484
Confidence interval	5.354
Intercept	-5.7262 ± 5.3543
Lower	-11.0805 - -0.3719

NO NO



Comments: The linearity of the method is acceptable because the individual calibration curves on all five days met the acceptance criteria. Also, other analytical data including the ion ratios and quantification values of the controls and calibrators met acceptance criteria. □

Acceptance Criteria:

95% CI of slope includes 1
 95% CI of intercept includes 0

Validation Study 2

Sensitivity (LOD)

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Concentration (ng/mL)	Signal to Noise Ratio			
	10			
AMP_20200814B_SD	NA			
	1619			
	2831			
	1029			
	5118			
	2699			
	4250			
	6302			
	894			
AMP_20200828B_PK	2547			
	NA			
	1988			
	3799			
	1398			
	1719			
	4532			
	5331			
	2724			
AMP_20200901B_CLR	NA			
	1103			
	3718			
	1141			
	2061			
	1917			
	2087			
	3133			
	3765			
AMP_20200901B_PK	797			
	NA			
	1541			
	1455			
	2836			
	3285			
	3781			
	838			
1682				
Average Signal to Noise:	2622			

Established LOD: 10 ng/mL
Signal to Noise: 2622

Comments: Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Blood source 5 had an interference with the 77.1 m/z qualifier ion. This ion was removed from the acquisition method on 9/10/2020.

Acceptance Criteria:

Signal to Noise ≥ 3.3
Acceptable detection and identification criteria

Validation Study 3

SENSITIVITY (LOQ)

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LOQ
<i>Target Concentration (ng/mL):</i>		10
AMP_20200814B_SD	1-1	NA
	1-2	10.85
	1-3	10.22
	1-4	10.55
	1-5	10.86
	1-6	11.44
	1-7	10.53
	1-8	10.64
	1-9	10.47
	<i>Within Run</i>	Mean
SD		0.36
%CV		3.41%
% Bias		6.94%
AMP_20200828B_PK	2-1	NA
	2-2	10.56
	2-3	10.16
	2-4	10.35
	2-5	5.00*
	2-6	10.29
	2-7	10.63
	2-8	10.39
	2-9	10.52
	<i>Within Run</i>	Mean
SD		0.17
%CV		1.60%
% Bias		4.13%
AMP_20200901B_CLR	3-1	NA
	3-2	10.42
	3-3	9.82
	3-4	10.11
	3-5	9.36
	3-6	10.53
	3-7	12.89
	3-8	10.35
	3-9	10.18
	<i>Within Run</i>	Mean
SD		1.05
%CV		10.05%
% Bias		4.57%
AMP_20200901B_PK	4-1	NA
	4-2	9.88
	4-3	9.91
	4-4	10.34
	4-5	11.05
	4-6	10.34
	4-7	10.38
	4-8	10.25
	4-9	10.12
	<i>Within Run</i>	Mean
SD		0.36
%CV		3.54%
% Bias		2.85%

Mean		10.46
SD		0.59
Precision (%CV)	Max Within-Run	10.05%
	Between-Run	5.64%
% Bias		4.62%

Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Blood source 5 had an interference with the 77.1 m/z qualifier ion. This ion was removed from the acquisition method on 9/10/2020. Number with * was > 20% target value due to insufficient IS. Number in red was > 20% target value.

Acceptance Criteria: Bias: ≤20%
 Within-Run Precision: CV ≤20%
 Between-Run Precision: CV ≤20%

Validation Study 4

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

BIAS AND PRECISION

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LQC	MQC	UTAK	HQC	
Target Concentration (ng/mL):		25	100	78.71	400	
AMP_20200814B_SD	1-1	25.02	94.30	66.97	402.83	
	1-2	23.79	89.77	71.87	404.40	
	1-3	24.61	93.02	68.31	394.66	
	1-4			72.49		
	Within Run	Mean	24.47	92.36	69.91	400.63
	SD	0.63	2.33	2.69	5.23	
	%CV	2.56%	2.53%	3.85%	1.31%	
% Bias	-2.10%	-7.64%	-11.18%	0.16%		
AMP_20200814B_PK	2-1	22.89	76.35	94.69	457.18	
	2-2	27.26	79.08	94.61	504.72	
	2-3	26.06	85.16	101.84	470.01	
	2-4			105.71		
	Within Run	Mean	25.40	80.19	99.21	477.30
	SD	2.26	4.51	5.50	24.60	
	%CV	8.90%	5.62%	5.54%	5.15%	
% Bias	1.62%	-19.81%	26.05%	19.33%		
AMP_20200828B_PK	3-1	25.39	89.73	73.81	420.92	
	3-2	27.46	97.57	72.36	400.50	
	3-3	24.56	86.10	71.12	398.53	
	3-4			70.42		
	Within Run	Mean	25.80	91.13	71.93	406.65
	SD	1.49	5.86	1.49	12.39	
	%CV	5.79%	6.43%	2.07%	3.05%	
% Bias	3.21%	-8.87%	-8.61%	1.66%		
AMP_20200828B_SD	4-1	23.08	98.05	75.43	390.83	
	4-2	24.14	92.95	79.57	386.11	
	4-3	24.20	95.35	68.89	374.02	
	4-4			71.21		
	Within Run	Mean	23.81	95.45	73.77	383.65
	SD	0.63	2.55	4.72	8.67	
	%CV	2.65%	2.67%	6.40%	2.26%	
% Bias	-4.77%	-4.55%	-6.27%	-4.09%		
AMP_20200901B_CLR	5-1	28.14	114.14		371.26	
	5-2	29.60	78.77		394.29	
	5-3	25.39	92.47		390.31	
	Within Run	Mean	27.71	95.13		385.29
	SD	2.14	17.84		12.31	
	%CV	7.71%	18.75%		3.19%	
	% Bias	10.84%	-4.87%		-3.68%	
AMP_20200901B_PK	6-1	24.17	80.84		404.33	
	6-2	24.74	89.34		388.53	
	6-3	23.89	85.38		380.12	
	Within Run	Mean	24.27	85.19		390.99
	SD	0.43	4.26		12.29	
	%CV	1.79%	5.00%		3.14%	
	% Bias	-2.92%	-14.81%		-2.25%	
Mean		25.24	89.91	78.71	407.42	
SD		1.81	8.92	12.80	35.22	
Precision (%CV)	Max Within-Run	8.90%	18.75%	6.40%	5.15%	
	Between-Run	7.18%	9.92%	16.27%	8.65%	
% Bias		0.98%	-10.09%	0.00%	1.85%	

Comments: Numbers in red were not within 20% of the target value.

Acceptance Criteria: %Bias ≤20%
 Within-Run %CV ≤20%
 Between-Run %CV ≤20%

Validation Study 5**CARRYOVER**

Analyte: MDA
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Average LOQ Response*: 202648

Study Date:	Response		
	AMP_20200814B_SD	AMP_20200814B_PK	AMP_20200828B_PK
Concentrated Sample (2000 ng/mL)	14714024	13840992	16918292
Blank	3987	3572	4366
%LOD Response	1.97%	1.76%	2.15%

Maximum Response in Blank: **2.2%**

Results: No significant carryover observed following samples containing analyte at up to 2000 ng/mL.

Comments: N/A

Acceptance Criteria: No analyte carryover is observed in the matrix blank samples; response in blank samples is <LOQ of the method.

Validation Study 6

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

MATRIX & IS INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

	LOQ Response	
	Analyte	IS
Run 1	222811	927008
Run 2	146270	679186
Run 3	246190	1000000
Run 4	211291	981337
Run 5	186675	763357
Average	202648	870178

Matrix Interference

Study Date: AMP_20200814B_PK

Matrix Source	Peak at Analyte RT	
	Peak Response	Percent of LOQ Response
4	551	0.27%
5	774	0.38%
7	848	0.42%
8	1167	0.58%
9	1189	0.59%
10	981	0.48%
11	1029	0.51%
12	500	0.25%
13	651	0.32%
14	1334	0.66%

Interference from Stable-Isotope Internal Standards

Study Date: AMP_20200828B_SD

Experiment	Peak at Analyte RT		Peak at IS RT	
	Peak Response	Percent of LOQ Response	Peak Response	Percent of LOQ Response
Matrix with IS but no D0 (IS = 50 ng/mL)	5476	2.70%	N/A	N/A
Matrix with D0 but no IS (D0 = 2000 ng/mL)	N/A	N/A	352	0.04%

Matrix Interference: No significant interference observed.

IS Interference: No significant interference observed.

Comments: LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.

Acceptance Criteria:

Response of blank matrix is less than 20% the average response of LOQ

Validation Study 6

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

EXOGENOUS SUBSTANCE INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target LQC Concentration (µg/mL): 25.24
 Control Acceptance: 20%
 Run Date: AMP 20200828B SD

Group	Compound	Compound Concentration (µg/mL)	Calculated LQC Concentration (ng/mL)	% Difference from Target	Comment
Benzodiazepines	Alprazolam	1	24.66	-2%	No significant interference
	α-Hydroxyalprazolam				
	Clonazepam				
	7-Aminoclonazepam				
	Diazepam				
	Nordiazepam				
	Temazepam				
	Lorazepam				
	Oxazepam				
Zolpidem					
Phencyclidine	Phencyclidine	1	25.75	-2%	No significant interference
Opioids	Morphine	2	25.63	2%	No significant interference
	Hydrocodone	1			
	Buprenorphine	0.5			
	Norbuprenorphine				
	Codeine				
	Fentanyl				
	Norfentanyl oxalate				
	Hydromorphone				
	Methadone				
	EDDP				
	Oxycodone				
	Oxymorphone				
	Tramadol				
o-Desmethyltramadol					
Cocaine and Metabolites	Benzoylcegonine	2	24.61	-3%	No significant interference
	Cocaine	0.5			
	Cocaethylene				
Carisoprodol/Meprobamate	Carisoprodol	1	24.63	-2%	No significant interference
	Meprobamate				
Basic and Neutral Mix	Amiripityline	5	24.43	-3%	No significant interference
	Benzpiperazine				
	Chlorpheniramine				
	Cyclobenzaprine				
	Dextromethorphan				
	Diphenhydramine				
	Doxylamine				
	Fluoxetine				
	Imipramine				
	Ketamine				
	Norketamine				
	Meperidine				
	Nortriptyline				
	Propoxyphene				
	Sertraline				
	Trazodone				
	Venlafaxine				
Zopiclone					
Over-the-Counter Drugs	Acetaminophen	10	23.24	-8%	No significant interference
	Caffeine				
	Ibuprofen				
	Naproxen				
	Pseudoephedrine				
Theobromine	1				
Cannabinoids	Δ9-THC	0.5	25.25	0%	No significant interference
	11-Hydroxy-Δ9-THC				
	Δ9-THC-COOH				
	Δ8-THC				
	Cannabinol				
Cannabinolic acid					

Conclusions: No significant interference observed.

Comments: N/A

Acceptance Criteria: Concentrations of analytes of interest within ±20% of the average concentration obtained in the Bias and Precision studies

Validation Study 7

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

DILUTION INTEGRITY

Analyt: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target Concentration (ng/mL): 366.55

Run Date	Dilution Factor	Diluted Result (ng/mL)	Calculated Result (µg/mL)	Mean	SD	%Bias	Absolute % Bias	Within-Run Precision (per Dilution)
AMP_20200814B_SD	2	177.66	355.33	339.85	20.53	-3.06	3.06	6.04
	2	173.83	347.65			-5.16	5.16	
	2	158.28	316.56			-13.64	13.64	
	5	70.66	353.32	333.79	20.50	-3.61	3.61	6.14
	5	67.12	335.60			-8.44	8.44	
	5	62.49	312.44			-14.76	14.76	
	10	36.29	362.86	338.24	33.04	-1.01	1.01	9.77
	10	35.12	351.17			-4.20	4.20	
	10	30.07	300.69			-17.97	17.97	
	AMP_20200828B_PK	2	184.93	369.86	394.31	22.39	0.90	0.90
2		199.62	399.24	8.92			8.92	
2		206.91	413.82	12.90			12.90	
5		70.90	354.50	381.13	23.14	-3.29	3.29	6.07
5		79.26	396.30			8.12	8.12	
5		78.52	392.60			7.11	7.11	
10		34.16	341.60	378.03	31.61	-6.81	6.81	8.36
10		39.81	398.10			8.61	8.61	
10		39.44	394.40			7.60	7.60	
AMP_20200828B_SD		2	180.84	361.68	359.26	6.74	-1.33	1.33
	2	182.23	364.45	-0.57			0.57	
	2	175.82	351.65	-4.06			4.06	
	5	72.00	360.01	353.79	14.05	-1.78	1.78	3.97
	5	72.73	363.66			-0.79	0.79	
	5	67.54	337.69			-7.87	7.87	
	10	31.56	315.63	329.74	15.95	-13.89	13.89	4.84
	10	34.70	347.04			-5.32	5.32	
	10	32.65	326.54			-10.91	10.91	
	AMP_20200901B_PK	2	198.36	396.71	407.00	9.03	8.23	8.23
2		206.81	413.63	12.84			12.84	
2		205.33	410.65	12.03			12.03	
5		72.83	364.14	390.95	23.43	-0.66	0.66	5.99
5		80.25	401.23			9.46	9.46	
5		81.50	407.49			11.17	11.17	
10		39.46	394.56	379.86	22.23	7.64	7.64	5.85
10		35.43	354.29			-3.34	3.34	
10		39.07	390.74			6.60	6.60	
AMP_20200901B_CLR		2	153.53	307.06	344.17	38.72	-16.23	16.23
	2	192.16	384.31	4.85			4.85	
	2	170.58	341.15	-6.93			6.93	
	5	71.43	357.17	379.95	19.85	-2.56	2.56	5.22
	5	77.83	389.15			6.16	6.16	
	5	78.71	393.53			7.36	7.36	
	10	36.12	361.23	355.84	4.99	-1.45	1.45	1.40
	10	35.14	351.39			-4.14	4.14	
	10	35.49	354.89			-3.18	3.18	

Dilution Factor	2
Avg %Bias	7.44
Max Within-Run Precision (%CV)	11.25

Dilution Factor	5
Avg %Bias	6.21
Max Within-Run Precision (%CV)	6.14

Dilution Factor	10
Avg %Bias	6.84
Max Within-Run Precision (%CV)	9.77

Results: It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.
Comments: N/A

Acceptance Criteria:

Avg %Bias must be less than 20%

Validation Study 8Analyte: MDA
Units: ng/mL
Instrument: LCMS-1**PROCESSED SAMPLE STABILITY**Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Run Date	Sample	Time Zero Response	24H Response	48H Response	72H Response	Acceptable Range		Pass/Unstable
						Low	High	
	LQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			
	HQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			

Results:**Comments:** Study not performed. Sample preparation will be completed once started without prolonged interruptions.

Acceptance Criteria:

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 8

Analyte: MDA
Units: ng/mL
Instrument: LCMS-1

AUTOSAMPLER STABILITY

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	Calibrator 1	10.17	0.26	10.25	10.18	0.26	1%
	Calibrator 2	19.12	0.48	19.12	18.75	0.48	0%
	Calibrator 3	51.40	1.28	51.95	50.49	1.30	1%
	Calibrator 4	98.78	2.45	100.57	97.48	2.50	2%
	Calibrator 5	249.03	6.18	248.65	240.60	6.17	0%
	Calibrator 6	507.28	12.57	510.65	493.84	12.65	1%

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20201019B_SD & AMP_20201021B_SD	Calibrator 1	10.29	0.28	N/A	N/A	N/A	N/A	10.26	10.28	0.28	0%
	Calibrator 2	19.48	0.51	N/A	N/A	N/A	N/A	19.49	19.09	0.51	0%
	Calibrator 3	45.50	1.16	N/A	N/A	N/A	N/A	45.49	43.93	1.16	0%
	Calibrator 4	99.53	2.52	N/A	N/A	N/A	N/A	99.49	95.50	2.51	0%
	Calibrator 5	258.94	6.51	N/A	N/A	N/A	N/A	259.15	248.00	6.52	0%
	Calibrator 6	528.08	13.27	N/A	N/A	N/A	N/A	523.10	500.11	13.14	-1%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	LQC	24.38	0.61	24.53	23.99	0.62	1%
		24.29	0.61	24.64	24.09	0.62	1%
		24.27	0.61	24.44	23.89	0.61	1%
	HQC	393.02	9.74	396.93	383.92	9.84	1%
		394.78	9.79	396.35	383.36	9.82	0%
		395.82	9.81	395.06	382.11	9.79	0%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20201019B_SD & AMP_20201021B_SD	LQC	24.67	0.64	N/A	N/A	N/A	N/A	24.58	23.95	0.64	0%
		24.69	0.64	N/A	N/A	N/A	N/A	24.75	24.11	0.64	0%
		24.66	0.64	N/A	N/A	N/A	N/A	24.77	24.13	0.64	0%
	HQC	401.55	10.09	N/A	N/A	N/A	N/A	394.43	377.20	9.91	-2%
		400.16	10.06	N/A	N/A	N/A	N/A	395.42	378.16	9.94	-1%
		398.60	10.02	N/A	N/A	N/A	N/A	397.47	380.11	9.99	0%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	LQC	24.38	0.61	24.38	23.84	0.61	0%
		24.29	0.61	24.56	24.01	0.62	1%
		24.27	0.61	24.62	24.07	0.62	1%
	HQC	393.02	9.74	395.69	382.73	9.81	1%
		394.78	9.79	397.53	384.50	9.85	1%
		395.82	9.81	398.57	385.51	9.88	1%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20201019B_SD & AMP_20201021B_SD	LQC	24.67	0.64	N/A	N/A	N/A	N/A	24.63	24.00	0.64	0%
		24.69	0.64	N/A	N/A	N/A	N/A	24.67	24.04	0.64	0%
		24.66	0.64	N/A	N/A	N/A	N/A	24.76	24.13	0.64	0%
	HQC	401.55	10.09	N/A	N/A	N/A	N/A	396.30	378.99	9.96	-1%
		400.16	10.06	N/A	N/A	N/A	N/A	397.63	380.27	9.99	-1%
		398.60	10.02	N/A	N/A	N/A	N/A	394.17	376.96	9.91	-1%

Results: Punctured and unpunctured samples were shown to be stable for 48 hours.

Comments: AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not stable; therefore the 72 hr data for MDA were not included. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.

Acceptance Criteria:

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 9

Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

MATRIX EFFECTS

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Neat Response at Analyte RT

	LQC	HQC
1	362346	3562600
2	341261	3601120
3	354361	3637204
4	358215	3591569
5	338333	3441000
6	337385	3643637
Mean	348650	3579522
SD	10951	74195
%CV	3%	2%

Neat Response at IS RT

	LQC	HQC
	568223	370576
	531651	368744
	609232	372199
	592467	360812
	543609	390241
	555689	346856
Mean	566812	368238
SD	29523	14268
%CV	5%	4%

Matrix Effect Study Date: AMP_20200904B_SD

Matrix Source	Peak at Analyte RT			
	Peak Response at LQC	% suppression/enhancement	Peak Response at HQC	% suppression/enhancement
4	325007	-7%	3285537	-8%
5	310820	-11%	3553173	-1%
7	313815	-10%	3559901	-1%
8	336070	-4%	3446874	-4%
9	281484	-19%	3341211	-7%
10	317463	-9%	3254067	-9%
11	328510	-6%	3391030	-5%
12	337111	-3%	3486792	-3%
13	312584	-10%	3499839	-2%
14	332928	-5%	3609470	1%
Mean	319579		3442790	
SD	16604		121212	
%CV	5%		4%	
Average % suppression/enhancement	-8%		-4%	

Matrix Effect Study Date: AMP_20200904B_SD

Matrix Source	Peak at IS RT			
	Peak Response at LQC	% suppression/enhancement	Peak Response at HQC	% suppression/enhancement
4	591599	4%	355658	-3%
5	577217	2%	382886	4%
7	583710	3%	387330	5%
8	612124	8%	374005	2%
9	515791	-9%	361226	-2%
10	577038	2%	358241	-3%
11	594049	5%	368026	0%
12	616578	9%	378096	3%
13	564120	0%	377426	2%
14	614439	8%	384105	4%
Mean	584666		372700	
SD	29954		11324	
%CV	5%		3%	
Average % suppression/enhancement	3%		1%	

Results: Average percent suppression values for LQC and HQC were -8% and -4%, respectively.
Comments: N/A

Acceptance Criteria: Average suppression/enhancement ≤25% or the ≤%CV of the suppression/enhancement 20%

Validation Study 10


Analyte: MDA
 Units: ng/mL
 Instrument: LCMS-1

Additional Experiment

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

LINEARITY

Batch Name	Target	Calculated Result	% RE	y (Response Ratio)
AMP_202000910B_SD	10	9.98	-0.23	0.242
	20	20.26	1.29	0.493
	50	50.20	0.41	1.224
	100	94.74	-5.26	2.311
	250	244.44	-2.22	5.964
	500	530.11	6.02	12.936



Slope 0.02576
 Intercept -0.13176
 R² 0.99916

LOQ

Run Date	Run Order	LOQ
<i>Target Concentration (ng/mL):</i>		10
AMP_202000910B_SD	1-1	10.15
	1-2	9.79
	1-3	10.22
	1-4	10.24
	1-5	10.34
	1-6	10.70
	1-7	10.44
	1-8	10.18
	1-9	10.62
	1-10	10.48
<i>Within Run</i>	Mean	10.32
	SD	0.26
	%CV	2.55%
	% Bias	3.15%

BIAS & PRECISION

Run Date	Run Order	LQC	MQC	HQC
<i>Target Concentration (ng/mL):</i>		25	100	400
AMP_202000910B_SD	1-1	24.79	107.81	404.17
	1-2	24.82	104.38	396.80
	1-3	24.98	107.37	416.18
	Mean	24.86	106.52	405.72
<i>Within Run</i>	SD	0.10	1.87	9.78
	%CV	0.40%	1.76%	2.41%
	% Bias	-0.55%	6.52%	1.43%

The purpose of this experiment was to verify the change in qualifier ions from 105.1 m/z and 77.1 m/z to 133.1 m/z and 79.1 m/z. Blank blood sources 4 (lot 310355), 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study.

Acceptance Criteria: %RE Calibrators $\pm 20\%$ of target
 Bias: $\leq 20\%$
 Within-Run Precision: CV $\leq 20\%$

SUMMARY OF VALIDATION PERFORMANCE

Analyte: MDA
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

The intent of this summary is to capture and document important information about the performance of this method outside the required measurements for validation.

Failed Runs (include dates/reasons):

Date	Reason
N/A	N/A

Deviations from SOP:

On 9/10/2020 the qualifier ions for MDA were changed from 105.1 m/z and 77.1 m/z to the new ions of 133.1 m/z and 79.1 m/z due to interferences with the original ions. An additional experiment was performed on 9/10/2020 that included linearity, bias & precision, and LOQ. All parameters met acceptance criteria; therefore, the new ions were deemed acceptable for use.

Other Observations:

Working Standards Verified in Validation:
Calibrators: 200805C-C-10, 200805C-C-2.5, 200805C-C-0.2, 200923C-C-10, 200923C-C-2.5, 200923C-C-0.2
Controls: 200807L-Q-Mix, 200813L-Q-0.5, 200813L-St-5
Internal Standard: 200805C-IS-1, 200903C-IS-1, 200915C-IS-1

Sample Preparation Steps:

Refer to Toxicology Analytical Manual v3.5, "Amphetamines Confirmation by Liquid Chromatography-Tandem Mass Spectrometry" section titled "Extraction Procedure".

Location of Raw Data:

Toxicology section shared electronic storage.

Recommended Maximum Run Length (Unknown Samples):

30

Conclusion:

This method is fit for use on casework for MDA confirmation analysis in blood.

METHOD VALIDATION PROTOCOL AND RESULTS

Analyte: MDMA
 Units: ng/mL
 Method: AMP.M
 Instrument: LCMS-1
 SOP Reference: Toxicology Analytical Manual v3.5

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

VALIDATION EXPERIMENT		SOP CRITERIA	RESULTS	COMMENTS
1	Weight Verification	The least complex weighting scheme that minimizes $\sum %RE $	Unweighted: = 629.22 1/x Weighting: = 385.11 1/x2 Weighting: = 377.46	Data were processed using 1/x2 weighting.
1	Linearity	95% CI of slope includes 1 95% CI of intercept includes 0	95% CI of slope = 0.9526 - 0.9932 95% CI of Intercept = -4.4564 - 13.2224	The linearity of the method is acceptable because the individual calibration curves on all five days met the acceptance criteria. Also, other analytical data including the ion ratios and quantification values of the controls and calibrators met acceptance criteria.
1	Validation Calibration	%RE Calibrators $\pm 20\%$ of target	Max %RE = 11.95	N/A
1	Case Work Calibration	%RE Calibrators $\pm 20\%$ of target	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
1	Comparison of Validation Calibration to Casework Calibration	95% CI of slope includes 1 95% CI of intercept includes 0	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
2	Limit of Detection (LOD)	Signal to Noise ≥ 3.3 Acceptable detection and identification criteria	LOD = 10 ng/mL S:N= 3270	The lowest non-zero calibrator was defined as the LOD.
3	Limit of Quantitation (LOQ)	Bias: $\leq 20\%$ Within-Run Precision: CV $\leq 20\%$ Between-Run Precision: CV $\leq 20\%$	Bias = 2.11% Within-Run Precision= 10.43% Between-Run Precision = 5.74%	Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Number in red was $> 20\%$ target value. Numbers with * were $> 20\%$ target value due to insufficient IS.
4	Bias & Precision	%Bias $\leq 20\%$ Within-Run %CV $\leq 20\%$ Between-Run %CV $\leq 20\%$	Max Bias = -9.02% Max Within-Run Precision = 18.79% Max Between-Run Precision = 14.78%	Numbers in red were not within 20% of the target value.
5	Carryover	No analyte carryover is observed in the matrix blank samples; response in blank samples is $< LOQ$ of the method.	No significant carryover observed following samples containing analyte at up to 2000 ng/mL.	N/A
6	Matrix Interference	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.
6	Interference from stable isotope internal standard	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	
6	Exogenous Substances Interferences	Concentrations of analytes of interest within $\pm 20\%$ of the average concentration obtained in the Bias and Precision studies	No significant interference observed.	N/A
7	Dilution Integrity	Average %Bias must be less than 20%	It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.	N/A
8	Processed Sample Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	N/A	Study not performed. Sample preparation will be completed once started without interruptions.
8	Autosampler Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	Punctured and unpunctured samples were shown to be stable for 48 hours.	AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not stable; therefore the 72 hr data for MDMA were not included. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.
9	Matrix Effects	Average suppression/enhancement $\leq 25\%$ or the \leq CV of the suppression/enhancement 20%	Average percent suppression values for LQC and HQC were -16% and -8%, respectively.	N/A

Validation Study 1

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

STANDARD CURVE WEIGHT VERIFICATION

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Unweighted: 629.22
 1/x Weighting: 385.11
 1/x² Weighting: 377.46

	C _{nom}	y	w	wxy	wx	wy	wx ²	wy ²	C _{found}	%RE	%RE
AMP_20200814B_SD	10	0.25225	1	2.522546	10	0.25225461	100	0.063632	5.7247	-42.753	42.753
	20	0.48722	1	9.744497	20	0.48722485	400	0.237388	16.78629	-16.0686	16.06857
	50	1.24941	1	62.47074	50	1.2494147	2500	1.561037	52.66755	5.3351	5.3351
	100	2.56309	1	256.3095	100	2.56309467	10000	6.569454	114.5111	14.51106	14.51106
	250	5.55288	1	1388.219	250	5.55287569	62500	30.83443	255.2596	2.103846	2.103846
	500	11.11483	1	5557.417	500	11.1148334	250000	123.5395	517.0974	3.419471	3.419471
	1000	20.70035	1	20700.35	1000	20.7003519	1000000	428.5046	968.3504	-3.16496	3.164957
AMP_20200814B_PK	10	0.22491	1	2.24908	10	0.224908	100	0.050584	4.437316	-55.6268	55.62684
	20	0.41306	1	8.26112	20	0.413056	400	0.170615	13.29467	-33.5266	33.52663
	50	0.83125	1	41.56255	50	0.831251	2500	0.690978	32.98185	-34.0363	34.0363
	100	1.73628	1	173.6282	100	1.736282	10000	3.014675	75.58758	-24.4124	24.41242
	250	4.88321	1	1220.803	250	4.883212	62500	23.84576	223.7342	-10.5063	10.50633
	500	9.36278	1	4681.388	500	9.362776	250000	87.66157	434.6166	-13.0767	13.07669
	1000	18.02597	1	18025.97	1000	18.02597	1000000	324.9356	842.4498	-15.755	15.75502
AMP_20200828B_SD	10	0.26652	1	2.665204	10	0.26652035	100	0.071033	6.396282	-36.0372	36.03718
	20	0.49992	1	9.998468	20	0.49992341	400	0.249923	17.38409	-13.0795	13.07954
	50	1.27470	1	63.73509	50	1.27470179	2500	1.624865	53.85798	7.715957	7.715957
	100	2.34730	1	234.7298	100	2.34729814	10000	5.509809	104.3521	4.352107	4.352107
	250	5.83258	1	1458.146	250	5.83258294	62500	34.01902	268.4273	7.370906	7.370906
	500	12.29250	1	6146.248	500	12.292497	250000	151.1055	572.5377	14.50754	14.50754
	1000	22.07351	1	22073.51	1000	22.0735145	1000000	487.24	1032.994	3.299418	3.299418
AMP_20200828B_PK	10	0.22748	1	2.274758	10	0.2274758	100	0.051745	4.558199	-54.418	54.41801
	20	0.43194	1	8.638729	20	0.43193646	400	0.186569	14.1835	-29.0825	29.0825
	50	0.97897	1	48.9485	50	0.97896995	2500	0.956382	39.93595	-20.1281	20.12811
	100	1.91075	1	191.0751	100	1.91075082	10000	3.650969	83.80097	-16.199	16.19903
	250	5.11929	1	1279.822	250	5.11928852	62500	26.20711	234.8478	-6.06087	6.060868
	500	9.91325	1	4956.625	500	9.91325098	250000	98.27255	460.531	-7.8938	7.893798
	1000	20.18783	1	20187.83	1000	20.1878343	1000000	407.5487	944.2229	-5.57771	5.577713
AMP_20200901B_CLR	10	0.25469	1	2.546862	10	0.2546862	100	0.064865	5.839171	-41.6083	41.60829
	20	0.52399	1	10.47983	20	0.52399136	400	0.274567	18.51713	-7.41436	7.414365
	50	1.23633	1	61.81673	50	1.23633452	2500	1.528523	52.05178	4.10356	4.10356
	100	2.74417	1	274.417	100	2.74417048	10000	7.530472	123.0355	23.03548	23.03548
	250	6.37900	1	1594.75	250	6.37900008	62500	40.69164	294.1507	17.66028	17.66028
	500	12.72574	1	6362.872	500	12.7257434	250000	161.9445	592.9334	18.58669	18.58669
	1000	24.94011	1	24940.11	1000	24.940108	1000000	622.009	1167.943	16.79435	16.79435
Sum	9650.0	209.5581	5	142042.1	9650	209.558084	6627500	3082.42			629.2219
Slope											0.02124200
Intercept											0.13065052
R ²											0.97932871

Comments: Data were processed using 1/x² weighting.

Acceptance Criteria: The least complex weighting scheme that minimizes Σ|%RE|

Validation Study 1

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

VALIDATION CURVE CALIBRATION


Analyte: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Batch Name	Target	Calculated Result	% RE	y (Response Ratio)
AMP_20200814B_SD	10	9.77	2.31	0.252
	20	20.05	0.27	0.487
	50	53.42	6.84	1.249
	100	110.93	10.93	2.563
	250	241.81	3.28	5.553
	500	485.29	2.94	11.115
AMP_20200814B_PK	10	10.09	0.94	0.225
	20	20.62	3.11	0.413
	50	44.03	11.95	0.831
	100	94.67	5.33	1.736
	250	270.77	8.31	4.883
	500	521.45	4.29	9.363
AMP_20200828B_SD	10	9.93	0.66	0.267
	20	19.87	0.64	0.500
	50	52.86	5.73	1.275
	100	98.54	1.46	2.347
	250	246.94	1.22	5.833
	500	522.02	4.40	12.292
AMP_20200828B_PK	10	10.00	0.00	0.227
	20	20.39	1.95	0.432
	50	48.17	3.66	0.979
	100	95.50	4.50	1.911
	250	258.47	3.39	5.119
	500	501.97	0.39	9.913
AMP_20200901B_CLR	10	9.91	0.88	0.255
	20	20.40	2.01	0.524
	50	48.15	3.70	1.236
	100	106.89	6.89	2.744
	250	248.48	0.61	6.379
	500	495.71	0.86	12.726
	1000	971.51	2.85	24.940


Max %RE = 11.95

Comments: N/A

Acceptance Criteria: %RE Calibrators ±20% of target



Slope 0.02070
 Intercept 0.28258
 R² 0.99931




Slope 0.01887
 Intercept -0.01698
 R² 0.99947



Slope 0.01778
 Intercept 0.47014
 R² 0.99751



Slope 0.02253
 Intercept 0.01179
 R² 0.99826



Slope 0.01999
 Intercept 0.10778
 R² 0.99989

Validation Study 1 **LINEARITY**

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1



Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Study Date	Target (x)	Calculated (y)	Predicted	Residual
AMP_20200814B_SD	10	9.77	14.11	-4.34
	20	20.05	23.84	-3.79
	50	53.42	53.03	0.39
	100	110.93	101.67	9.25
	250	241.81	247.61	-5.80
	500	485.29	490.83	-5.55
AMP_20200814B_PK	10	10.09	14.11	-4.02
	20	20.62	23.84	-3.22
	50	44.03	53.03	-9.00
	100	94.67	101.67	-7.00
	250	270.77	247.61	23.17
	500	521.45	490.83	30.62
AMP_20200828B_PK	10	10.00	14.11	-4.11
	20	20.39	23.84	-3.45
	50	48.17	53.03	-4.86
	100	95.50	101.67	-6.17
	250	258.47	247.61	10.86
	500	501.97	490.83	11.14
AMP_20200828B_SD	10	9.93	14.11	-4.18
	20	19.87	23.84	-3.97
	50	52.86	53.03	-0.16
	100	98.54	101.67	-3.14
	250	246.94	247.61	-0.66
	500	522.02	490.83	31.18
AMP_20200901B_CLR	10	9.91	14.11	-4.20
	20	20.40	23.84	-3.44
	50	48.15	53.03	-4.88
	100	106.89	101.67	5.21
	250	248.48	247.61	0.87
	500	495.71	490.83	4.88
	1000	971.51	977.28	-5.77

Slope	0.9729
Std err in slope, S_b	0.0100
Degrees freedom	33
Confidence level	95%
Student t	2.0345
Confidence interval	0.020
Slope	0.9729 ± 0.0203
Range	0.9526 - 0.9932

Intercept	4.3830
Std err in Intercept	4.3447
Degrees freedom	33
Confidence Level	95%
Student t	2.0345
Confidence interval	8.839
Intercept	4.383 ± 8.8394
Lower	-4.4564 - 13.2224

NO YES



Comments: The linearity of the method is acceptable because the individual calibration curves on all five days met the acceptance criteria. Also, other analytical data including the ion ratios and quantification values of the controls and calibrators met acceptance criteria.

Acceptance Criteria:

95% CI of slope includes 1
 95% CI of intercept includes 0

Validation Study 2

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

Sensitivity (LOD)

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Concentration (ng/mL)	Signal to Noise Ratio			
	10			
AMP_20200814B_SD	2929			
	4659			
	4762			
	4789			
	1388			
	7028			
	7771			
	6318			
	3705			
AMP_20200828B_PK	2512			
	1968			
	2860			
	3959			
	1702			
	3215			
	1398			
	2219			
401				
AMP_20200901B_CLR	3511			
	544			
	2252			
	5585			
	1373			
	1935			
	3270			
	2079			
4154				
AMP_20200901B_PK	1282			
	484			
	1761			
	827			
	2775			
	3512			
	523			
	2043			
2258				
Average Signal to Noise:	3270			

Established LOD: 10 ng/mL
Signal to Noise: 3270

Comments: The lowest non-zero calibrator was defined as the LOD.

Acceptance Criteria:

Signal to Noise ≥3.3
Acceptable detection and identification criteria

Validation Study 3

SENSITIVITY (LOQ)

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LOQ
<i>Target Concentration (ng/mL):</i>		
		10
AMP_20200814B_SD	1-1	10.16
	1-2	10.33
	1-3	9.71
	1-4	9.89
	1-5	10.05
	1-6	10.94
	1-7	10.17
	1-8	9.95
	1-9	9.78
	Within Run	Mean
SD		0.37
%CV		3.64%
% Bias		1.10%
AMP_20200828B_PK	2-1	10.28
	2-2	10.41
	2-3	9.98
	2-4	10.20
	2-5	4.54*
	2-6	10.25
	2-7	10.69
	2-8	10.07
	2-9	10.57
	Within Run	Mean
SD		0.24
%CV		2.34%
% Bias		3.07%
AMP_20200901B_CLR	3-1	52.40*
	3-2	10.35
	3-3	9.67
	3-4	9.97
	3-5	9.21
	3-6	10.65
	3-7	12.84
	3-8	10.12
	3-9	10.28
	Within Run	Mean
SD		1.08
%CV		10.43%
% Bias		3.87%
AMP_20200901B_PK	4-1	10.16
	4-2	9.80
	4-3	9.73
	4-4	9.48
	4-5	9.96
	4-6	10.61
	4-7	10.29
	4-8	10.24
	4-9	10.07
	Within Run	Mean
SD		0.34
%CV		3.39%
% Bias		0.38%

Mean		10.20
SD		0.59
Precision (%CV)	Max Within-Run	10.43%
	Between-Run	5.74%
% Bias		2.11%

Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Number in red was > 20% target value. Numbers with * were > 20% target value due to insufficient IS.

Acceptance Criteria: Bias: ≤20%
 Within-Run Precision: CV ≤20%
 Between-Run Precision: CV ≤20%

Validation Study 4

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

BIAS AND PRECISION

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LQC	MQC	UTAK	HQC
Target Concentration (ng/mL):		25	100	79.64	800
AMP_202008014B_SD	1-1	26.31	98.72	69.97	730.22
	1-2	24.26	95.22	74.11	722.29
	1-3	25.79	97.51	71.59	720.42
	1-4			75.15	
	Within Run Mean	25.46	97.15	72.71	724.31
	SD	1.06	1.78	2.36	5.20
	%CV	4.18%	1.83%	3.24%	0.72%
% Bias	1.82%	-2.85%	-8.71%	-9.46%	
AMP_202008014B_PK	2-1	23.22	77.69	94.45	854.80
	2-2	27.42	79.95	93.43	890.21
	2-3	26.24	87.58	101.32	841.35
	2-4			104.21	
	Within Run Mean	25.63	81.74	98.35	862.12
	SD	2.16	5.18	5.25	25.24
	%CV	8.44%	6.34%	5.34%	2.93%
% Bias	2.51%	-18.26%	23.50%	7.77%	
AMP_20200828B_PK	3-1	25.51	NA	73.91	770.63
	3-2	27.65	NA	71.58	758.67
	3-3	24.66	NA	71.17	768.48
	3-4			69.80	
	Within Run Mean	25.94		71.61	765.93
	SD	1.54		1.71	6.37
	%CV	5.94%		2.39%	0.83%
% Bias	3.76%		-10.08%	-4.26%	
AMP_20200828B_SD	4-1	23.38	99.15	77.71	762.49
	4-2	23.90	94.28	81.35	767.76
	4-3	24.29	96.93	71.01	746.61
	4-4			73.47	
	Within Run Mean	23.85	96.79	75.88	758.95
	SD	0.46	2.44	4.58	11.01
	%CV	1.91%	2.52%	6.03%	1.45%
% Bias	-4.58%	-3.21%	-4.71%	-5.13%	
AMP_20200901B_PK	5-1	24.07	80.07		770.54
	5-2	24.72	88.74		763.70
	5-3	24.37	84.70		744.42
	Within Run Mean	24.39	84.50		759.55
	SD	0.32	4.34		13.54
	%CV	1.32%	5.13%		1.78%
	% Bias	-2.45%	-15.50%		-5.06%
AMP_20200901B_CLR	6-1	28.12	112.89		729.58
	6-2	29.17	77.32		781.06
	6-3	25.48	93.93		772.69
	Within Run Mean	27.59	94.71		761.11
	SD	1.90	17.80		27.62
	%CV	6.89%	18.79%		3.63%
	% Bias	10.37%	-5.29%		-4.86%

Mean		25.48	90.98	79.64	772.00
SD		1.71	9.94	11.77	46.13
Precision (%CV)	Max Within-Run	8.44%	18.79%	6.03%	3.63%
	Between-Run	6.72%	10.92%	14.78%	5.98%
% Bias		1.91%	-9.02%	0.00%	-3.50%

Comments: Numbers in red were not within 20% of the target value.

Acceptance Criteria:

%Bias ≤20%
 Within-Run %CV ≤20%
 Between-Run %CV ≤20%

Validation Study 5**CARRYOVER**

Analyte: MDMA
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Average LOQ Response*: 206683

Study Date:	Response		
	AMP_202008014B_SD	AMP_202008014B_PK	AMP_202008028B_PK
Concentrated Sample (2000 ng/mL)	15224166	14850080	20045769
Blank	4726	4480	5462
%LOD Response	2.29%	2.17%	2.64%

Maximum Response in Blank: **2.6%**

Results: No significant carryover observed following samples containing analyte at up to 2000 ng/mL.

Comments: N/A

Acceptance Criteria: No analyte carryover is observed in the matrix blank samples; response in blank samples is <LOQ of the method.

Validation Study 6

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

MATRIX & IS INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

	LOQ Response	
	Analyte	IS
Run 1	249596	989459
Run 2	151193	672245
Run 3	202304	759057
Run 4	216388	951259
Run 5	213932	839983
Average	206683	842401

Matrix Interference

Study Date: AMP_20200814B_PK

Matrix Source	Peak at Analyte RT	
	Peak Response	Percent of LOQ Response
4	287	0.14%
5	394	0.19%
7	341	0.16%
8	1728	0.84%
9	1145	0.55%
10	861	0.42%
11	696	0.34%
12	474	0.23%
13	565	0.27%
14	1315	0.64%

Interference from Stable-Isotope Internal Standards

Study Date: AMP_20200828B_SD

Experiment	Peak at Analyte RT		Peak at IS RT	
	Peak Response	Percent of LOQ Response	Peak Response	Percent of LOQ Response
Matrix with IS but no D0 (IS = 50 ng/mL)	5966	2.89%	N/A	N/A
Matrix with D0 but no IS (D0 = 2000 ng/mL)	N/A	N/A	312	0.04%

Matrix Interference: No significant interference observed.

IS Interference: No significant interference observed.

Comments: LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.

Acceptance Criteria:

Response of blank matrix is less than 20% the average response of LOQ

Validation Study 6

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

EXOGENOUS SUBSTANCE INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target LQC Concentration (µg/mL): 25.48
 Control Acceptance: 20%
 Run Date: AMP 20200828B SD

Group	Compound	Compound Concentration (µg/mL)	Calculated LQC Concentration (ng/mL)	% Difference from Target	Comment
Benzodiazepines	Alprazolam	1	24.52	-4%	No significant interference
	α-Hydroxyalprazolam				
	Clonazepam				
	7-Aminoclonazepam				
	Diazepam				
	Nordiazepam				
	Temazepam				
	Lorazepam				
	Oxazepam				
Zolpidem					
Phencyclidine	Phencyclidine	1	25.88	2%	No significant interference
Opioids	Morphine	2	24.34	-4%	No significant interference
	Hydrocodone	1			
	Buprenorphine	0.5			
	Norbuprenorphine				
	Codeine				
	Fentanyl				
	Norfentanyl oxalate				
	Hydromorphone				
	Methadone				
	EDDP				
	Oxycodone				
	Oxymorphone				
	Tramadol				
	o-Desmethytramadol				
Cocaine and Metabolites	Benzoylcegonine		2	24.46	-4%
	Cocaine	0.5			
	Cocaehtylene				
Carisoprodol/Meprobamate	Carisoprodol	1	24.48	-4%	No significant interference
	Meprobamate				
Basic and Neutral Mix	Amitriptyline	5	24.48	-4%	No significant interference
	Benzylpiperazine				
	Chlorpheniramine				
	Cyclobenzaprine				
	Dextromethorphan				
	Diphenhydramine				
	Doxylamine				
	Fluoxetine				
	Imipramine				
	Ketamine				
	Norketamine				
	Meperidine				
	Nortriptyline				
	Propoxyphene				
	Sertraline				
	Trazodone				
	Venlafaxine				
Zopiclone					
Over-the-Counter Drugs	Acetaminophen	10	23.11	-9%	No significant interference
	Caffeine				
	Ibuprofen				
	Naproxen				
	Pseudoephedrine	1			
Theobromine					
Cannabinoids	Δ9-THC	0.5	24.91	-2%	No significant interference
	11-Hydroxy-Δ9-THC				
	Δ9-THC-COOH				
	Δ8-THC				
	Cannabinol				
Cannabinolic acid					

Conclusions: No significant interference observed.

Comments: N/A

Acceptance Criteria: Concentrations of analytes of interest within ±20% of the average concentration obtained in the Bias and Precision studies

Validation Study 7

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

DILUTION INTEGRITY

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target Concentration (ng/mL): 359.98

Run Date	Dilution Factor	Diluted Result (ng/mL)	Calculated Result (µg/mL)	Mean	SD	%Bias	Absolute % Bias	Within-Run Precision (per Dilution)
AMP_202008014B_SD	2	183.03	366.05	353.65	18.38	1.69	1.69	5.20
	2	181.18	362.36			0.66	0.66	
	2	166.26	332.53			-7.63	7.63	
	5	75.39	376.93	359.71	20.91	4.71	4.71	5.81
	5	73.15	365.75			1.60	1.60	
	5	67.29	336.44			-6.54	6.54	
	10	38.88	388.76	362.71	34.32	8.00	8.00	9.46
	10	37.55	375.54			4.32	4.32	
	10	32.38	323.83			-10.04	10.04	
	AMP_20200828B_PK	2	185.35	370.70	398.09	26.40	2.98	2.98
2		200.09	400.18	11.17			11.17	
2		211.69	423.38	17.61			17.61	
5		73.05	365.25	391.78	22.98	1.46	1.46	5.87
5		81.07	405.35			12.60	12.60	
5		80.95	404.75			12.44	12.44	
10		34.87	348.70	388.30	34.48	-3.13	3.13	8.88
10		41.17	411.70			14.37	14.37	
10		40.45	404.50			12.37	12.37	
AMP_20200828B_SD		2	185.82	371.63	367.55	8.37	3.24	3.24
	2	186.55	373.10	3.64			3.64	
	2	178.97	357.93	-0.57			0.57	
	5	73.97	369.86	364.23	14.98	2.74	2.74	4.11
	5	75.12	375.58			4.33	4.33	
	5	69.45	347.26			-3.53	3.53	
	10	32.53	325.28	340.71	17.68	-9.64	9.64	5.19
	10	36.00	360.01			0.01	0.01	
	10	33.68	336.84			-6.43	6.43	
	AMP_20200901B_PK	2	200.46	400.92	407.67	6.41	11.37	11.37
2		206.84	413.68	14.92			14.92	
2		204.21	408.42	13.46			13.46	
5		74.10	370.50	397.30	23.48	2.92	2.92	5.91
5		81.42	407.12			13.10	13.10	
5		82.85	414.27			15.08	15.08	
10		40.10	401.02	389.11	23.90	11.40	11.40	6.14
10		36.16	361.59			0.45	0.45	
10		40.47	404.72			12.43	12.43	
AMP_20200901B_CLR		2	154.68	309.37	348.02	38.49	-14.06	14.06
	2	193.17	386.34	7.32			7.32	
	2	174.18	348.36	-3.23			3.23	
	5	73.72	368.59	388.54	18.30	2.39	2.39	4.71
	5	78.49	392.47			9.03	9.03	
	5	80.91	404.55			12.38	12.38	
	10	36.87	368.75	361.68	6.97	2.44	2.44	1.93
	10	35.48	354.81			-1.44	1.44	
	10	36.15	361.49			0.42	0.42	

Dilution Factor	2
Avg %Bias	7.57
Max Within-Run Precision (%CV)	11.06

Dilution Factor	5
Avg %Bias	6.99
Max Within-Run Precision (%CV)	5.91

Dilution Factor	10
Avg %Bias	6.46
Max Within-Run Precision (%CV)	9.46

Results: It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.
 Comments: N/A

Acceptance Criteria:

Avg %Bias must be less than 20%

Validation Study 8Analyte: MDMA
Units: ng/mL
Instrument: LCMS-1**PROCESSED SAMPLE STABILITY**Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Run Date	Sample	Time Zero Response	24H Response	48H Response	72H Response	Acceptable Range		Pass/Unstable
						Low	High	
	LQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			
	HQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			

Results:**Comments:** Study not performed. Sample preparation will be completed once started without interruptions.

Acceptance Criteria:

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 8

Analyte: MDMA
Units: ng/mL
Instrument: LCMS-1

AUTOSAMPLER STABILITY

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	Calibrator 1	9.92	0.26	10.04	9.86	0.26	1%
	Calibrator 2	19.66	0.48	19.83	19.36	0.48	1%
	Calibrator 3	53.76	1.26	53.68	52.20	1.26	0%
	Calibrator 4	102.48	2.37	102.10	99.17	2.37	0%
	Calibrator 5	245.29	5.64	248.62	241.33	5.72	1%
	Calibrator 6	502.16	11.52	501.53	486.70	11.50	0%
	Calibrator 7	939.29	21.51	938.11	910.27	21.49	0%

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	Calibrator 1	10.06	0.27	N/A	N/A	N/A	N/A	9.94	9.92	0.27	-1%
	Calibrator 2	20.07	0.50	N/A	N/A	N/A	N/A	20.17	19.72	0.50	0%
	Calibrator 3	47.16	1.12	N/A	N/A	N/A	N/A	47.13	45.55	1.12	0%
	Calibrator 4	102.15	2.38	N/A	N/A	N/A	N/A	103.84	99.89	2.42	2%
	Calibrator 5	254.66	5.88	N/A	N/A	N/A	N/A	254.36	244.11	5.87	0%
	Calibrator 6	519.93	11.97	N/A	N/A	N/A	N/A	519.39	498.04	11.95	0%
	Calibrator 7	966.98	22.22	N/A	N/A	N/A	N/A	964.94	924.94	22.17	0%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	Response Ratio
AMP_20200923B_SD & AMP_20200924B_SD	LQC	25.19	0.61	25.33	24.70	0.61	1%
		25.31	0.61	25.26	24.63	0.61	0%
		24.95	0.60	25.16	24.53	0.61	1%
	HQC	734.98	16.84	744.83	722.75	17.07	1%
		737.45	16.90	739.70	717.76	16.95	0%
		742.44	17.01	745.10	723.01	17.07	0%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	LQC	25.09	0.61	N/A	N/A	N/A	N/A	25.38	24.71	0.62	1%
		24.96	0.61	N/A	N/A	N/A	N/A	25.31	24.64	0.62	1%
		25.03	0.61	N/A	N/A	N/A	N/A	25.22	24.56	0.62	1%
	HQC	749.51	17.23	N/A	N/A	N/A	N/A	762.67	731.14	17.53	2%
		756.89	17.40	N/A	N/A	N/A	N/A	758.84	727.48	17.45	0%
		751.99	17.29	N/A	N/A	N/A	N/A	761.19	729.73	17.50	1%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	Response Ratio
AMP_20200923B_SD & AMP_20200924B_SD	LQC	25.19	0.61	25.20	24.57	0.61	0%
		25.31	0.61	25.37	24.73	0.61	0%
		24.95	0.60	25.32	24.69	0.61	1%
	HQC	734.98	16.84	743.87	721.82	17.04	1%
		737.45	16.90	745.37	723.27	17.08	1%
		742.44	17.01	742.68	720.66	17.02	0%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero	Concentration (Time Zero Curve)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	LQC	25.09	0.61	N/A	N/A	N/A	N/A	25.14	24.48	0.62	0%
		24.96	0.61	N/A	N/A	N/A	N/A	25.15	24.49	0.62	1%
		25.03	0.61	N/A	N/A	N/A	N/A	25.36	24.69	0.62	1%
	HQC	749.51	17.23	N/A	N/A	N/A	N/A	760.46	729.03	17.48	1%
		756.89	17.40	N/A	N/A	N/A	N/A	756.52	725.25	17.39	0%
		751.99	17.29	N/A	N/A	N/A	N/A	756.14	724.89	17.38	1%

Results: Punctured and unpunctured samples were shown to be stable for 48 hours.

Comments: AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not stable, therefore the 72 hr data for MDMA were not included. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.

Acceptance Criteria:

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 9

Analyte: MDMA
 Units: ng/mL
 Instrument: LCMS-1

MATRIX EFFECTS

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Neat Response at Analyte RT		
	LQC	HQC
1	393349	6488376
2	371460	6444722
3	382566	6491170
4	386798	6509970
5	369330	6214191
6	368034	6503590
Mean	378589	6442003
SD	10479	113918
%CV	3%	2%

Neat Response at IS RT		
	LQC	HQC
	631729	413530
	595093	411600
	689097	420179
	660047	405451
	605456	442964
	621645	383501
Mean	633845	412871
SD	35214	19384
%CV	6%	5%

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at Analyte RT			% suppression/enhancement
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	339410	-10%	5850994	-9%	
5	291989	-23%	6097799	-5%	
7	297308	-21%	6018770	-7%	
8	331180	-13%	5716567	-11%	
9	293660	-22%	5807287	-10%	
10	327260	-14%	5284226	-18%	
11	332241	-12%	5955490	-8%	
12	353836	-7%	6082211	-6%	
13	291461	-23%	6057075	-6%	
14	333293	-12%	6227281	-3%	
Mean	319164		5909770		
SD	23158		267533		
%CV	7%		5%		
Average % suppression/enhancement	-16%		-8%		

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at IS RT			% suppression/enhancement
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	623545	-2%	389332	-6%	
5	546517	-14%	405616	-2%	
7	561610	-11%	403889	-2%	
8	613070	-3%	386885	-6%	
9	548885	-13%	390660	-5%	
10	613306	-3%	351112	-15%	
11	604344	-5%	392959	-5%	
12	657766	4%	403492	-2%	
13	539597	-15%	396946	-4%	
14	624249	-2%	414299	0%	
Mean	593289		393519		
SD	40813		17209		
%CV	7%		4%		
Average % suppression/enhancement	-6%		-5%		

Results: Average percent suppression values for LQC and HQC were -16% and -8%, respectively.

Comments: N/A

Acceptance Criteria: Average suppression/enhancement ≤25% or the ≤%CV of the suppression/enhancement 20%

SUMMARY OF VALIDATION PERFORMANCE

Analyte: MDMA
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

The intent of this summary is to capture and document important information about the performance of this method outside the required measurements for validation.

Failed Runs (include dates/reasons):

Date	Reason
N/A	N/A

Deviations from SOP: N/A

Other Observations:

Working Standards Verified in Validation:
Calibrators: 200805C-C-10, 200805C-C-2.5, 200805C-C-0.2, 200923C-C-10, 200923C-C-2.5, 200923C-C-0.2
Controls: 200807L-Q-Mix, 200813L-Q-0.5, 200813L-St-5
Internal Standard: 200805C-IS-1, 200903C-IS-1, 200915C-IS-1

Sample Preparation Steps:

Refer to Toxicology Analytical Manual v3.5, "Amphetamines Confirmation by Liquid Chromatography-Tandem Mass Spectrometry" section titled "Extraction Procedure".

Location of Raw Data:

Toxicology section shared electronic storage.

Recommended Maximum Run Length (Unknown Samples):

30

Conclusion:

This method is fit for use on casework for MDMA confirmation analysis in blood.

METHOD VALIDATION PROTOCOL AND RESULTS

Analyte: MDEA
 Units: ng/mL
 Method: AMP.M
 Instrument: LCMS-1
 SOP Reference: Toxicology Analytical Manual v3.5

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

VALIDATION EXPERIMENT		SOP CRITERIA	RESULTS	COMMENTS
1	Weight Verification	The least complex weighting scheme that minimizes $\sum %RE $	Unweighted: = 535.96 1/x Weighting: = 363.82 1/x2 Weighting: = 349.31	Data were processed using 1/x2 weighting.
1	Linearity	95% CI of slope includes 1 95% CI of intercept includes 0	95% CI of slope = 0.9349 - 0.9812 95% CI of Intercept = -1.6720 - 9.1182	The linearity of the method is acceptable because the individual calibration curves on all five days met the acceptance criteria. Also, other analytical data including the ion ratios and quantification values of the controls and calibrators met acceptance criteria.
1	Validation Calibration	%RE Calibrators $\pm 20\%$ of target	Max %RE = 9.87	N/A
1	Case Work Calibration	%RE Calibrators $\pm 20\%$ of target	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
1	Comparison of Validation Calibration to Casework Calibration	95% CI of slope includes 1 95% CI of intercept includes 0	N/A	Not performed. All 6 calibrators used in validation will be used for casework.
2	Limit of Detection (LOD)	Signal to Noise ≥ 3.3 Acceptable detection and identification criteria	LOD = 10 ng/mL S:N = 4157	The lowest non-zero calibrator was defined as the LOD.
3	Limit of Quantitation (LOQ)	Bias: $\leq 20\%$ Within-Run Precision: CV $\leq 20\%$ Between-Run Precision: CV $\leq 20\%$	Bias = 1.75% Within-Run Precision = 10.57% Between-Run Precision = 5.64%	Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Number in red was $> 20\%$ target value. Numbers with * were $> 20\%$ target value due to insufficient IS.
4	Bias & Precision	%Bias $\leq 20\%$ Within-Run %CV $\leq 20\%$ Between-Run %CV $\leq 20\%$	Max Bias = -7.93% Max Within-Run Precision = 19.33% Max Between-Run Precision = 17.19%	Number in red was not within 20% of the target value.
5	Carryover	No analyte carryover is observed in the matrix blank samples; response in blank samples is $< LOQ$ of the method.	No significant carryover observed following samples containing analyte at up to 2000 ng/mL.	N/A
6	Matrix Interference	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.
6	Interference from stable isotope internal standard	Response of blank matrix is less than 20% the average response of LOQ	No significant interference observed.	
6	Exogenous Substances Interferences	Concentrations of analytes of interest within $\pm 20\%$ of the average concentration obtained in the Bias and Precision studies	No significant interference observed.	N/A
7	Dilution Integrity	Average %Bias must be less than 20%	It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.	N/A
8	Processed Sample Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	N/A	Study not performed. Sample preparation will be completed once started without prolonged interruptions.
8	Autosampler Stability	Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%	Punctured and unpunctured samples were shown to be stable for 48 hours.	AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not stable; therefore the 72 hr data for MDEA were not included. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.
9	Matrix Effects	Average suppression/enhancement $\leq 25\%$ or the \leq CV of the suppression/enhancement 20%	Average percent suppression values for LQC and HQC were -20% and -10%, respectively.	N/A

Validation Study 1

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

STANDARD CURVE WEIGHT VERIFICATION

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Unweighted: 535.96
 1/x Weighting: 363.82
 1/x² Weighting: 349.31

	C _{nom}	y	w	wxy	wx	wy	wx ²	wy ²	C _{found}	%RE	%RE
AMP_20200814B_SD	10	0.28553	1	2.855284	10	0.28552844	100	0.081526	6.506682	-34.9332	34.93318
	20	0.55939	1	11.18784	20	0.55939209	400	0.31292	18.35313	-8.23434	8.234341
	50	1.42739	1	71.36973	50	1.42739461	2500	2.037455	55.90011	11.80019	11.80019
	100	2.86432	1	286.4322	100	2.86432161	10000	8.204338	118.0569	18.05688	18.05688
	250	6.28665	1	1571.663	250	6.2866534	62500	39.52201	266.0958	6.438326	6.438326
	500	11.80453	1	5902.263	500	11.8045262	250000	139.3468	504.781	0.956207	0.956207
AMP_20200814B_PK	10	0.25475	1	2.54753	10	0.254753	100	0.064899	5.175437	-48.2456	48.24563
	20	0.46347	1	9.26934	20	0.463467	400	0.214802	14.20372	-28.9814	28.98138
	50	0.94908	1	47.45375	50	0.949075	2500	0.900743	35.20955	-29.5809	29.58091
	100	1.95758	1	195.7582	100	1.957582	10000	3.832127	78.83428	-21.1657	21.16572
	250	5.44247	1	1360.617	250	5.442468	62500	29.62046	229.5791	-8.16936	8.169359
	500	9.78285	1	4891.423	500	9.782845	250000	95.70406	417.3297	-16.5341	16.53406
AMP_20200828B_SD	10	0.29981	1	2.998103	10	0.29981026	100	0.089886	7.124467	-28.7553	28.75533
	20	0.59322	1	11.18642	20	0.5932075	400	0.31284	18.35005	-8.24977	8.249769
	50	1.40843	1	70.42129	50	1.40842579	2500	1.983663	55.07957	10.15913	10.15913
	100	2.8069	1	258.0687	100	2.806867	10000	6.659944	105.7878	5.787758	5.787758
	250	6.36041	1	1590.102	250	6.36040886	62500	40.4548	269.2862	7.714495	7.714495
	500	12.69487	1	6347.435	500	12.6948691	250000	161.1597	543.2944	8.658874	8.658874
AMP_20200828B_PK	10	0.25360	1	2.535996	10	0.25359964	100	0.064313	5.125546	-48.7445	48.74454
	20	0.48865	1	9.77291	20	0.48864551	400	0.238774	15.29287	-23.5357	23.53567
	50	1.08332	1	54.16606	50	1.08332118	2500	1.173585	41.0166	-17.9668	17.9668
	100	2.09596	1	209.5959	100	2.09595938	10000	4.393046	84.82003	-15.18	15.17997
	250	5.53419	1	1383.548	250	5.5341931	62500	30.62729	233.5468	-6.58127	6.581267
	500	10.23135	1	5115.677	500	10.2313536	250000	104.6806	436.7307	-12.6539	12.65386
AMP_20200901B_CLR	10	0.29213	1	2.921285	10	0.29212855	100	0.085339	6.792181	-32.0782	32.07819
	20	0.60799	1	12.15979	20	0.60798958	400	0.369651	20.4553	2.276507	2.276507
	50	1.41181	1	70.59072	50	1.41181448	2500	1.99322	55.22615	10.4523	10.4523
	100	3.11917	1	311.9174	100	3.11917403	10000	9.729247	129.081	29.08097	29.08097
	250	7.11643	1	1779.107	250	7.1164296	62500	50.64357	301.9892	20.79569	20.79569
	500	13.33480	1	6667.4	500	13.3348005	250000	177.8169	570.9757	14.19514	14.19514
Sum	4650.0	111.5509	5	38252.44	4650	111.550937	1627500	912.3186			535.9615
Slope						0.02311778					
Intercept						0.13510839					
R ²						0.97400078					

Comments: Data were processed using 1/x² weighting.

Acceptance Criteria: The least complex weighting scheme that minimizes Σ|%RE|

Validation Study 1

LINEARITY

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1



Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Study Date	Target (x)	Calculated (y)	Predicted	Residual
AMP_20200814B_SD	10	9.75	13.30	-3.55
	20	20.24	22.88	-2.65
	50	53.46	51.62	1.83
	100	108.45	99.52	8.93
	250	239.44	243.22	-3.78
AMP_20200814B_PK	10	10.01	13.30	-3.29
	20	20.65	22.88	-2.23
	50	45.41	51.62	-6.21
	100	96.82	99.52	-2.70
	250	274.48	243.22	31.26
AMP_20200828B_PK	10	9.80	13.30	-3.50
	20	20.98	22.88	-1.91
	50	49.25	51.62	-2.37
	100	97.40	99.52	-2.12
	250	260.89	243.22	17.66
AMP_20200828B_SD	10	9.91	13.30	-3.39
	20	19.99	22.88	-2.89
	50	52.98	51.62	1.36
	100	98.53	99.52	-0.99
	250	245.39	243.22	2.17
AMP_20200901B_CLR	10	9.78	13.30	-3.52
	20	20.80	22.88	-2.08
	50	48.84	51.62	-2.78
	100	108.41	99.52	8.89
	250	247.87	243.22	4.64
	500	464.81	482.72	-17.91

Slope	0.9580
Std err in slope, S_b	0.0113
Degrees freedom	28
Confidence level	95%
Student t	2.0484
Confidence interval	0.023
Slope	0.958 ± 0.0232
Range	0.9349 - 0.9812

Intercept	3.7231
Std err in Intercept	2.6338
Degrees freedom	28
Confidence Level	95%
Student t	2.0484
Confidence interval	5.395
Intercept	3.7231 ± 5.3951
Lower	-1.6720 - 9.1182

NO YES



Comments: The linearity of the method is acceptable because the individual calibration curves on all five days met the acceptance criteria. Also, other analytical data including the ion ratios and quantification values of the controls and calibrators met acceptance criteria.

Acceptance Criteria:

95% CI of slope includes 1
 95% CI of intercept includes 0

Validation Study 1

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1


VALIDATION CURVE CALIBRATION

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood


Batch Name	Target	Calculated Result	% RE	y (Response Ratio)
AMP_20200814B_SD	10	9.75	2.45	0.286
	20	20.24	1.18	0.559
	50	53.46	6.92	1.427
	100	108.45	8.45	2.864
	250	239.44	4.22	6.287
	500	450.63	9.87	11.805
AMP_20200814B_PK	10	10.01	0.14	0.255
	20	20.65	3.27	0.463
	50	45.41	9.18	0.949
	100	96.82	3.18	1.958
	250	274.48	9.79	5.442
	500	495.75	0.85	9.783
AMP_20200828B_PK	10	9.80	1.99	0.254
	20	20.98	4.88	0.489
	50	49.25	1.49	1.083
	100	97.40	2.60	2.096
	250	260.89	4.36	5.534
	500	484.23	3.15	10.231
AMP_20200828B_SD	10	9.91	0.92	0.300
	20	19.99	0.04	0.559
	50	52.98	5.97	1.408
	100	98.53	1.47	2.581
	250	245.39	1.84	6.360
	500	491.52	1.70	12.695
AMP_20200901B_CLR	10	9.78	2.20	0.2921
	20	20.80	4.00	0.6080
	50	48.84	2.31	1.4118
	100	108.41	8.41	3.1192
	250	247.87	0.85	7.1164
	500	464.81	7.04	13.3348

Max %RE = 9.87


Comments: N/A




Slope 0.02341
 Intercept 0.24270
 R² 0.99902




Slope 0.01977
 Intercept 0.07735
 R² 0.99835



Slope 0.02053
 Intercept 0.09844
 R² 0.99923



Slope 0.02523
 Intercept 0.07352
 R² 0.99997



Slope 0.02665
 Intercept 0.18352
 R² 0.99913

Acceptance Criteria: %RE Calibrators ±20% of target

Validation Study 2

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

Sensitivity (LOD)

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Concentration (ng/mL)	Signal to Noise Ratio			
	10			
AMP_20200814B_SD	8441			
	7976			
	2980			
	3684			
	10782			
	9267			
	9171			
	6879			
AMP_20200828B_PK	2863			
	2017			
	1609			
	2447			
	2554			
	4242			
	6556			
	6512			
AMP_20200901B_PK	1328			
	2342			
	2215			
	3339			
	2370			
	7272			
	4066			
	2492			
AMP_20200901B_CLR	4689			
	3738			
	4437			
	1264			
	4223			
	1790			
	4398			
	981			
845				
4450				
2314				
3130				
Average Signal to Noise:	4157			

Established LOD: 10 ng/mL

Signal to Noise: 4157

Comments: The lowest non-zero calibrator was defined as the LOD.

Acceptance Criteria:

Signal to Noise ≥ 3.3
Acceptable detection and identification criteria

Validation Study 3

SENSITIVITY (LOQ)

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LOQ
<i>Target Concentration (ng/mL):</i>		10
AMP_20200814B_SD	1-1	10.30
	1-2	10.37
	1-3	9.76
	1-4	9.94
	1-5	10.14
	1-6	11.14
	1-7	10.27
	1-8	10.09
	1-9	9.86
	<i>Within Run</i>	Mean
SD		0.41
%CV		3.98%
% Bias		2.08%
AMP_20200828B_PK	2-1	10.32
	2-2	9.91
	2-3	10.15
	2-4	9.87
	2-5	9.89
	2-6	3.95*
	2-7	10.10
	2-8	10.25
	2-9	9.85
	<i>Within Run</i>	Mean
SD		0.19
%CV		1.86%
% Bias		0.42%
AMP_20200901B_CLR	3-1	53.71*
	3-2	10.17
	3-3	9.78
	3-4	9.98
	3-5	9.00
	3-6	10.44
	3-7	12.77
	3-8	10.08
	3-9	10.15
	<i>Within Run</i>	Mean
SD		1.09
%CV		10.57%
% Bias		2.97%
AMP_20200901B_PK	4-1	9.97
	4-2	9.78
	4-3	9.70
	4-4	10.19
	4-5	10.64
	4-6	10.37
	4-7	10.44
	4-8	10.09
	4-9	10.20
	<i>Within Run</i>	Mean
SD		0.31
%CV		3.01%
% Bias		1.53%

Mean		10.18
SD		0.57
Precision (%CV)	Max Within-Run	10.57%
	Between-Run	5.64%
% Bias		1.75%

Blank blood sources 5 (lot 264184), 7 (lot 228467), 8 (lot 367648), 9 (lot 331307), 10 (lot 335183), 11 (lot 233762), 12 (lot 235659), 13 (lot 223450), and 14 (lot 279720) were used for this study. Number in red was > 20% target value. Numbers with * were > 20% target value due to insufficient IS.

Acceptance Criteria: Bias: ≤20%
 Within-Run Precision: CV ≤20%
 Between-Run Precision: CV ≤20%

Validation Study 4

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

BIAS AND PRECISION

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Run Date	Run Order	LQC	MQC	UTAK	HQC
<i>Target Concentration (ng/mL):</i>		25	100	82.34	400
AMP_20200814B_SD	1-1	25.66	96.60	69.69	358.47
	1-2	24.43	93.24	74.01	350.50
	1-3	25.30	95.17	70.39	344.17
	1-4			74.35	
	Mean	25.13	95.00	72.11	351.04
	SD	0.63	1.68	2.41	7.17
	%CV	2.52%	1.77%	3.34%	2.04%
AMP_20200814B_PK	2-1	24.26	80.62	102.78	426.73
	2-2	28.26	82.98	101.94	439.33
	2-3	27.20	90.29	107.27	415.71
	2-4			109.20	
	Mean	26.58	84.63	105.30	427.26
	SD	2.07	5.04	3.50	11.82
	%CV	7.79%	5.96%	3.32%	2.77%
AMP_20200828B_PK	3-1	25.63	91.83	77.54	381.13
	3-2	28.33	100.48	75.89	368.66
	3-3	24.85	88.94	73.10	375.05
	3-4			75.10	
	Mean	26.27	93.75	75.41	374.95
	SD	1.82	6.00	1.84	6.24
	%CV	6.95%	6.41%	2.45%	1.66%
AMP_20200828B_SD	4-1	23.38	100.15	78.93	365.75
	4-2	24.06	96.29	82.66	362.48
	4-3	24.40	96.33	70.66	360.60
	4-4			73.85	
	Mean	23.95	97.59	76.53	362.95
	SD	0.52	2.21	5.32	2.61
	%CV	2.18%	2.27%	6.96%	0.72%
AMP_20200901B_CLR	5-1	27.90	115.45		353.38
	5-2	29.76	78.96		370.27
	5-3	25.96	92.23		374.32
	Mean	27.87	95.54		365.99
	SD	1.90	18.47		11.10
	%CV	6.82%	19.33%		3.03%
	% Bias	11.49%	-4.46%		-8.50%
AMP_20200901B_PK	6-1	24.54	81.84		374.15
	6-2	24.87	91.08		368.35
	6-3	24.82	84.84		363.33
	Mean	24.74	85.92		368.61
	SD	0.17	4.71		5.41
	%CV	0.71%	5.48%		1.47%
	% Bias	-1.03%	-14.08%		-7.85%

Mean		25.76	92.07	82.34	375.13
SD		1.79	8.77	14.15	26.00
Precision (%CV)	Max Within-Run	7.79%	19.33%	6.96%	3.03%
	Between-Run	6.93%	9.52%	17.19%	6.93%
% Bias		3.02%	-7.93%	0.00%	-6.22%

Comments: Number in red was not within 20% of the target value.

Acceptance Criteria:

%Bias ≤20%
 Within-Run %CV ≤20%
 Between-Run %CV ≤20%

Validation Study 5**CARRYOVER**

Analyte: MDEA
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Average LOQ Response*: 203508

Study Date:	Response		
	AMP_20200814B_SD	AMP_20200814B_PK	AMP_20200828B_PK
Concentrated Sample (2000 ng/mL)	13331962	12785352	17480700
Blank	4214	5571	7809
%LOD Response	2.07%	2.74%	3.84%

Maximum Response in Blank: **3.8%**

Results: No significant carryover observed following samples containing analyte at up to 2000 ng/mL.

Comments: N/A

Acceptance Criteria: No analyte carryover is observed in the matrix blank samples; response in blank samples is <LOQ of the method.

Validation Study 6

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

MATRIX & IS INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

	LOQ Response	
	Analyte	IS
Run 1	255002	893087
Run 2	145439	570902
Run 3	182272	600000
Run 4	214297	845019
Run 5	220528	754901
Average	203508	732782

Matrix Interference		AMP_20200814B_PK	
Matrix Source	Peak at Analyte RT		
	Peak Response	Percent of LOQ Response	
4	214	0.11%	
5	266	0.13%	
7	229	0.11%	
8	821	0.40%	
9	970	0.48%	
10	722	0.35%	
11	608	0.30%	
12	363	0.18%	
13	558	0.27%	
14	1222	0.60%	

Interference from Stable-Isotope Internal Standards			Study Date: AMP_20200828B_SD	
Experiment	Peak at Analyte RT		Peak at IS RT	
	Peak Response	Percent of LOQ Response	Peak Response	Percent of LOQ Response
Matrix with IS but no D0 (IS = 50 ng/mL)	8177	4.02%	N/A	N/A
Matrix with D0 but no IS (D0 = 400 ng/mL)	N/A	N/A	246	0.03%

Matrix Interference: No significant interference observed.

IS Interference: No significant interference observed.

Comments: LOQ reponse per run was taken from the lowest calibrator in analytical runs between 8/14/2020 and 9/1/2020.

Acceptance Criteria:

Response of blank matrix is less than 20% the average response of LOQ

Validation Study 6

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

EXOGENOUS SUBSTANCE INTERFERENCE

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target LQC Concentration (µg/mL): 25.76
 Control Acceptance: 20%
 Run Date: AMP 20200828B SD

Group	Compound	Compound Concentration (µg/mL)	Calculated LQC Concentration (ng/mL)	% Difference from Target	Comment	
Benzodiazepines	Alprazolam	1	24.66	-4%	No significant interference	
	α-Hydroxyalprazolam					
	Clonazepam					
	7-Aminoclonazepam					
	Diazepam					
	Nordiazepam					
	Temazepam					
	Lorazepam					
	Oxazepam					
Zolpidem						
Phencyclidine	Phencyclidine	1	26.06	1%	No significant interference	
Opioids	Morphine	2	24.38	-5%	No significant interference	
	Hydrocodone	1				
	Buprenorphine	0.5				
	Norbuprenorphine					
	Codeine					
	Fentanyl					
	Norfentanyl oxalate					
	Hydromorphone					
	Methadone					
	EDDP					
	Oxycodone					
	Oxymorphone					
	Tramadol					
	o-Desmethyltramadol					
Cocaine and Metabolites	Benzoylcegonine		2	24.67	-4%	No significant interference
	Cocaine		0.5			
	Cocaineethylene					
Carisoprodol/Meprobamate	Carisoprodol	1	24.60	-4%	No significant interference	
	Meprobamate					
Basic and Neutral Mix	Amitriptyline	5	24.77	-4%	No significant interference	
	Benzpiperazine					
	Chlorpheniramine					
	Cyclobenzaprine					
	Dextromethorphan					
	Diphenhydramine					
	Doxylamine					
	Fluoxetine					
	Imipramine					
	Ketamine					
	Norketamine					
	Mepredine					
	Nortriptyline					
	Propoxyphene					
	Sertraline					
	Trazodone					
	Venlafaxine					
Zopiclone						
Over-the-Counter Drugs	Acetaminophen	10	23.25	-10%	No significant interference	
	Caffeine					
	Ibuprofen					
	Naproxen					
	Pseudoephedrine					
Theobromine	1					
Cannabinoids	Δ9-THC	0.5	25.14	-2%	No significant interference	
	11-Hydroxy-Δ9-THC					
	Δ9-THC-COOH					
	Δ8-THC					
	Cannabinol					
Cannabinolic acid						

Conclusions: No significant interference observed.

Comments: N/A

Acceptance Criteria: Concentrations of analytes of interest within ±20% of the average concentration obtained in the Bias and Precision studies

Validation Study 7

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

DILUTION INTEGRITY

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Target Concentration (ng/mL): 342.17

Run Date	Dilution Factor	Diluted Result (ng/mL)	Calculated Result (µg/mL)	Mean	SD	%Bias	Absolute % Bias	Within-Run Precision (per Dilution)
AMP_20200814B_SD	2	173.02	346.04	333.83	18.06	1.13	1.13	5.41
	2	171.18	342.37			0.06	0.06	
	2	156.55	313.09			-8.50	8.50	
	5	70.76	353.79	339.35	18.31	3.40	3.40	5.40
	5	69.10	345.51			0.98	0.98	
	5	63.75	318.76			-6.84	6.84	
	10	37.03	370.33	345.31	33.58	8.23	8.23	9.72
	10	35.84	358.44			4.75	4.75	
	10	30.71	307.15			-10.24	10.24	
	AMP_20200828B_PK	2	181.92	363.84	385.45	21.91	6.33	6.33
2		192.43	384.86	12.48			12.48	
2		203.82	407.64	19.13			19.13	
5		71.61	358.04	382.43	21.13	4.64	4.64	5.52
5		79.00	395.01			15.44	15.44	
5		78.85	394.23			15.22	15.22	
10		34.14	341.35	375.27	29.47	-0.24	0.24	7.85
10		39.47	394.70			15.35	15.35	
10		38.98	389.76			13.91	13.91	
AMP_20200828B_SD		2	178.64	357.29	353.29	9.61	4.42	4.42
	2	180.12	360.25	5.28			5.28	
	2	171.16	342.32	0.04			0.04	
	5	71.77	358.86	353.39	13.78	4.88	4.88	3.90
	5	72.72	363.60			6.26	6.26	
	5	67.54	337.72			-1.30	1.30	
	10	31.55	315.48	327.50	13.75	-7.80	7.80	4.20
	10	34.25	342.49			0.09	0.09	
	10	32.45	324.53			-5.15	5.15	
	AMP_20200901B_CLR	2	148.20	296.40	330.84	34.40	-13.38	13.38
2		182.60	365.19	6.73			6.73	
2		165.46	330.92	-3.29			3.29	
5		70.43	352.15	369.65	15.35	2.92	2.92	4.15
5		75.19	375.97			9.88	9.88	
5		76.17	380.83			11.30	11.30	
10		35.70	357.01	350.04	6.16	4.34	4.34	1.76
10		34.54	345.37			0.93	0.93	
10		34.77	347.73			1.63	1.63	
AMP_20200901B_PK		2	190.84	381.68	393.43	10.35	11.54	11.54
	2	200.59	401.17	17.24			17.24	
	2	198.72	397.44	16.15			16.15	
	5	71.09	355.44	383.83	25.11	3.88	3.88	6.54
	5	76.57	392.87			14.82	14.82	
	5	80.63	403.17			17.83	17.83	
	10	39.58	395.79	377.29	25.93	15.67	15.67	6.87
	10	34.77	347.65			1.60	1.60	
	10	38.84	388.41			13.51	13.51	

Dilution Factor	2
Avg %Bias	8.38
Max Within-Run Precision (%CV)	10.40

Dilution Factor	5
Avg %Bias	7.97
Max Within-Run Precision (%CV)	6.54

Dilution Factor	10
Avg %Bias	6.90
Max Within-Run Precision (%CV)	9.72

Results: It is suitable to dilute blood samples 2x, 5x, or 10x prior to analysis.
Comments: N/A

Acceptance Criteria:

Avg %Bias must be less than 20%

Validation Study 8Analyte: MDEA
Units: ng/mL
Instrument: LCMS-1**PROCESSED SAMPLE STABILITY**Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

Run Date	Sample	Time Zero Response	24H Response	48H Response	72H Response	Acceptable Range		Pass/Unstable
						Low	High	
	LQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			
	HQC	N/A	N/A	N/A	N/A	#DIV/0!	#DIV/0!	N/A
		N/A	N/A	N/A	N/A			
		N/A	N/A	N/A	N/A			

Results:**Comments:** Study not performed. Sample preparation will be completed once started without prolonged interruptions.**Acceptance Criteria:**

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 8

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

AUTOSAMPLER STABILITY

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero (Response Ratio)
AMP_20200923B_SD & AMP_20200924B_SD	Calibrator 1	9.89	0.29	9.79	9.86	0.29	-1%
	Calibrator 2	19.85	0.55	19.75	19.46	0.55	0%
	Calibrator 3	53.88	1.42	53.90	52.37	1.42	0%
	Calibrator 4	101.92	2.64	102.29	99.01	2.65	0%
	Calibrator 5	240.68	6.18	243.40	235.00	6.25	1%
	Calibrator 6	479.71	12.27	479.52	462.54	12.27	0%

Stability of Punctured Calibrators

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero)	Response Ratio	Concentration (Time Zero)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero	Concentration (Time Zero)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	Calibrator 1	10.06	0.30	N/A	N/A	N/A	N/A	10.18	10.14	0.31	1%
	Calibrator 2	20.06	0.56	N/A	N/A	N/A	N/A	20.18	19.72	0.56	1%
	Calibrator 3	47.26	1.25	N/A	N/A	N/A	N/A	47.05	45.48	1.25	0%
	Calibrator 4	102.43	2.67	N/A	N/A	N/A	N/A	103.02	99.14	2.68	1%
	Calibrator 5	254.18	6.55	N/A	N/A	N/A	N/A	252.42	242.36	6.50	-1%
	Calibrator 6	502.21	12.90	N/A	N/A	N/A	N/A	500.99	480.65	12.86	0%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	Response Ratio
AMP_20200923B_SD & AMP_20200924B_SD	LQC	25.34	0.69	24.79	24.32	0.67	-2%
		25.03	0.68	25.38	24.89	0.69	1%
		24.76	0.67	25.22	24.74	0.69	2%
	HQC	358.65	9.19	360.16	347.51	9.23	0%
		356.12	9.12	360.40	347.75	9.23	1%
		356.62	9.14	359.21	346.60	9.20	1%

Stability of Punctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero)	Response Ratio	Concentration (Time Zero)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero	Concentration (Time Zero)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	LQC	25.45	0.70	N/A	N/A	N/A	N/A	25.56	24.88	0.70	0%
		25.36	0.69	N/A	N/A	N/A	N/A	25.45	24.78	0.70	0%
		25.63	0.70	N/A	N/A	N/A	N/A	25.20	24.53	0.69	-2%
	HQC	363.52	9.35	N/A	N/A	N/A	N/A	359.53	345.04	9.24	-1%
		360.43	9.27	N/A	N/A	N/A	N/A	360.98	346.43	9.28	0%
		363.25	9.34	N/A	N/A	N/A	N/A	362.14	347.55	9.31	0%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours			
		Concentration (Time Zero Curve)	Response Ratio	Concentration (Time Zero Curve)	Concentration (24 h Curve)	Response Ratio	Response Ratio
AMP_20200923B_SD & AMP_20200924B_SD	LQC	25.34	0.69	25.04	24.56	0.68	-1%
		25.03	0.68	25.35	24.87	0.69	1%
		24.76	0.67	24.98	24.50	0.68	1%
	HQC	358.65	9.19	360.25	347.60	9.23	0%
		356.12	9.12	358.85	346.25	9.19	1%
		356.62	9.14	361.01	348.33	9.25	1%

Stability of Unpunctured Controls

Run Date	Sample	Time Zero		24 Hours				48 Hours			
		Concentration (Time Zero)	Response Ratio	Concentration (Time Zero)	Concentration (24 h Curve)	Response Ratio	% Difference from Time Zero	Concentration (Time Zero)	Concentration (48 h Curve)	Response Ratio	% Difference from Time
AMP_20201019B_SD & AMP_20201021B_SD	LQC	25.45	0.70	N/A	N/A	N/A	N/A	25.34	24.66	0.69	0%
		25.36	0.69	N/A	N/A	N/A	N/A	25.42	24.74	0.70	0%
		25.63	0.70	N/A	N/A	N/A	N/A	25.27	24.60	0.69	-1%
	HQC	363.52	9.35	N/A	N/A	N/A	N/A	360.98	346.44	9.28	-1%
		360.43	9.27	N/A	N/A	N/A	N/A	361.63	347.06	9.30	0%
		363.25	9.34	N/A	N/A	N/A	N/A	363.15	348.52	9.34	0%

Results: Punctured and unpunctured samples were shown to be stable for 48 hours.
Comments: AMP_20200904B_SD showed the punctured and unpunctured samples for methamphetamine at 72 hr were not stable; therefore the 72 hr data for MDEA were not included. These data can be found in a supplemental Excel located in the Toxicology section shared electronic storage.

Acceptance Criteria:

Average signal (peak area, or ratio of peak area analyte/IS) compared to time 0 is within 20%

Validation Study 9

Analyte: MDEA
 Units: ng/mL
 Instrument: LCMS-1

MATRIX EFFECTS

Analyst: SD, PK, and CLR
 Study Dates: 8/14/2020 to 9/24/2020
 Matrix: Blood

Neat Response at Analyte RT		
	LQC	HQC
1	407305	3900146
2	382858	3874153
3	397271	3911042
4	398607	3914083
5	384454	3729381
6	382650	3935315
Mean	392191	3877353
SD	10328	75192
%CV	3%	2%

Neat Response at IS RT		
	LQC	HQC
	586466	448869
	547858	451012
	629070	452954
	610671	437647
	570033	471021
	577715	425476
Mean	586969	447830
SD	29109	15354
%CV	5%	3%

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at Analyte RT			
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	332993	-15%	3520615	-9%	
5	276253	-30%	3645120	-6%	
7	284383	-27%	3546662	-9%	
8	321140	-18%	3358029	-13%	
9	291710	-26%	3403063	-12%	
10	321126	-18%	2874818	-26%	
11	332387	-15%	3543726	-9%	
12	353234	-10%	3613103	-7%	
13	278037	-29%	3579524	-8%	
14	331703	-15%	3729843	-4%	
Mean	312297		3481450		
SD	27318		239091		
%CV	9%		7%		
Average % suppression/enhancement	-20%		-10%		

Matrix Effect		Study Date: AMP_20200904B_SD			
Matrix Source	Peak Response at LQC	Peak at IS RT			
		% suppression/enhancement	Peak Response at HQC	% suppression/enhancement	
4	546608	-7%	418968	-6%	
5	457871	-22%	427257	-5%	
7	472156	-20%	426408	-5%	
8	525896	-10%	401327	-10%	
9	480321	-18%	406874	-9%	
10	535505	-9%	336457	-25%	
11	536677	-9%	417987	-7%	
12	582285	-1%	433751	-3%	
13	448073	-24%	420820	-6%	
14	547079	-7%	444020	-1%	
Mean	513247		413387		
SD	45111		29681		
%CV	9%		7%		
Average % suppression/enhancement	-13%		-8%		

Results: Average percent suppression values for LQC and HQC were -20% and -10%, respectively.
Comments: N/A

Acceptance Criteria: Average suppression/enhancement ≤25% or the ≤%CV of the suppression/enhancement 20%

SUMMARY OF VALIDATION PERFORMANCE

Analyte: MDEA
Units: ng/mL
Instrument: LCMS-1

Analyst: SD, PK, and CLR
Study Dates: 8/14/2020 to 9/24/2020
Matrix: Blood

The intent of this summary is to capture and document important information about the performance of this method outside the required measurements for validation.

Failed Runs (include dates/reasons):

Date	Reason
N/A	N/A

Deviations from SOP: N/A

Other Observations: Working Standards Verified in Validation:
Calibrators: 200805C-C-10, 200805C-C-2.5, 200805C-C-0.2, 200923C-C-10, 200923C-C-2.5, 200923C-C-0.2
Controls: 200807L-Q-Mix, 200813L-Q-0.5, 200813L-St-5
Internal Standard: 200805C-IS-1, 200903C-IS-1, 200915C-IS-1

Sample Preparation Steps: Refer to Toxicology Analytical Manual v3.5, "Amphetamines Confirmation by Liquid Chromatography-Tandem Mass Spectrometry" section titled "Extraction Procedure".

Location of Raw Data: Toxicology section shared electronic storage.

Recommended Maximum Run Length (Unknown Samples): 30

Conclusion: This method is fit for use on casework for MDEA confirmation analysis in blood.