

## Verification of Immunoassay Using Commercial Kits

Purpose	To demonstrate the newly made calibrators and controls are valid for use in casework
Analytes	Amphetamines, Barbiturates, Benzodiazepines, Cocaine and metabolites, Cannabinoids, Carisoprodol, Fentanyl Methadone, Methamphetamine, Opiates, Phencyclidine, and Oxycodone
Units of Measure	ng/mL
Analysts Performing Validation Studies	A. Gooden & V. Coronado
Responsible Manager	Dayong Lee, Ph.D.
Start Date	August 10, 2022
Completion Date	August 18, 2022
Primary Matrix	Blood
Calibrator Concentration	d-Amphetamine (20 ng/mL); Secobarbital (150 ng/mL); Oxazepam (20 ng/mL); Buprenorphine (1 ng/mL); Benzoylcegonine (50 ng/mL); 11-nor-9-carboxy-THC (10 ng/mL); Carisoprodol (500 ng/mL); Fentanyl (1 ng/mL); d-Methamphetamine (20 ng/mL); Morphine (10 ng/mL); Phencyclidine (10 ng/mL); and Oxycodone (10 ng/mL)
Equipment/Instrument	Tecan-2
Instrument Serial Number	1010008386
Method	ELISA

### Validation Approval

Analyst: **Andrea Gooden** Digitally signed by Andrea Gooden  
Date: 2022.08.26 13:55:42 -05'00' 08/26/2022  
Date

Analyst: **Valerie L. Coronado** Digitally signed by Valerie L. Coronado  
Date: 2022.08.26 13:53:58 -05'00' 08/26/2022  
Date

Responsible Manager: **Dayong Lee** Digitally signed by Dayong Lee  
Date: 2022.08.26 14:07:22 -05'00' 08/26/2022  
Date

**Validation Study**

**ELISA CONTROL CRITERIA**

Method **ELISA**  
 Units: **ng/mL**  
 Instrument: **Tecan-2**

Analyst: **A. Gooden & V. Coronado**  
 Study Dates: **8/10/22 to 8/18/22**  
 Matrix: **Blood**

Batch Name	Analyte	Blank Avg. Abs	Blank % CV	Cal % Binding	Absorbance of Blank>Neg>Cal>Pos
Day 1	<i>Acceptance Criteria</i>	>1.0	<20		Yes
EIA_20220810B_VC	Amphetamines	2.77	0.66	62.60	Yes
	Barbiturates	1.77	2.67	36.41	Yes
	Benzodiazepines	1.78	3.94	39.33	Yes
	Buprenorphine	2.86	0.59	41.74	Yes
	Cocaine and metabolites	2.65	0.83	44.11	Yes
	Cannabinoids	2.99	0.99	46.31	Yes
	Carisoprodol	2.78	0.25	35.81	Yes
	Fentanyl	2.36	0.54	32.04	Yes
	Methamphetamine	2.54	2.51	51.16	Yes
	Opiates	2.37	3.05	54.76	Yes
	Opiates*	2.34	0.88	54.45	Yes
	Phencyclidine	2.47	2.89	41.95	Yes
Oxycodone	2.75	0.05	14.66	Yes	

Batch Name	Analyte	Blank Avg. Abs	Blank % CV	Cal % Binding	Absorbance of Blank>Neg>Cal>Pos
Day 2	<i>Acceptance Criteria</i>	>1.0	<20		Yes
EIA_20220811B_ASG	Amphetamines	2.70	0.39	63.01	Yes
	Barbiturates	1.83	10.38	37.37	Yes
	Benzodiazepines	1.69	1.80	42.19	Yes
	Buprenorphine	2.84	1.85	44.57	Yes
	Cocaine and metabolites*	2.81	0.93	49.67	Yes
	Cannabinoids	3.08	0.64	47.30	Yes
	Carisoprodol	2.80	0.78	37.39	Yes
	Fentanyl	2.47	2.17	32.91	Yes
	Methamphetamine	2.54	0.67	54.94	Yes
	Opiates	2.07	2.32	59.83	Yes
	Phencyclidine	2.39	0.59	47.85	Yes
	Oxycodone	2.60	3.32	15.48	Yes

Batch Name	Analyte	Blank Avg. Abs	Blank % CV	Cal % Binding	Absorbance of Blank>Neg>Cal>Pos
Day 3	<i>Acceptance Criteria</i>	>1.0	<20		Yes
EIA_20220815_VC	Amphetamines	2.81	1.36	66.32	Yes
	Barbiturates	1.98	2.14	38.66	Yes
	Benzodiazepines	1.69	2.01	39.20	Yes
	Buprenorphine	2.79	1.14	41.53	Yes
	Cocaine and metabolites	2.78	4.86	42.51	Yes
	Cannabinoids	2.82	4.87	49.04	Yes
	Carisoprodol	2.81	1.43	37.39	Yes
	Fentanyl	2.34	1.36	30.78	Yes
	Methamphetamine	2.49	1.28	51.25	Yes
	Opiates	2.28	0.34	51.90	Yes
	Phencyclidine	2.33	2.46	45.15	Yes
	Oxycodone	2.51	5.65	12.85	Yes

Batch Name	Analyte	Blank Avg. Abs	Blank % CV	Cal % Binding	Absorbance of Blank>Neg>Cal>Pos
Day 4	<i>Acceptance Criteria</i>	>1.0	<20		Yes
EIA_20220818_ASG	Benzodiazepines*	2.31	0.28	41.92	Yes

Comments: Refer to Page 18 for all abbreviations. \*Samples were reinjected or reanalyzed.

Acceptance Criteria: Refer to the Enzyme-Linked Immunosorbent Assay (ELISA) section 12.9 of the Analytical Manual Standard Operating Procedures Version 3.8.

**Analyte: Amphetamines**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

**Analyte: Amphetamines**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Target Analyte: d-Amphetamine  
Cal Conc: 20 ng/mL  
Neg QC Conc: 10 ng/mL  
Pos QC Conc: 60 ng/mL  
Sample Volume: 10 µL

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	2.023	1.739	0.901	73.138	62.871	32.574
2	2.201	1.724	0.884	79.573	62.328	31.960
3	2.237	1.629	0.913	80.875	58.894	33.008
4	2.256	1.703	0.972	81.562	61.569	35.141
5	2.184	1.624	0.946	78.959	58.713	34.201

	Mean	Stdev	%CV
Mean	2.180	0.923	33.377
Stdev	0.092	0.035	1.283
%CV	4.237	3.843	3.843

Mean -(neg) +(pos) 2SD  
Acceptable? PASS

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	2.062	1.767	0.900	76.243	65.336	33.278
2	2.086	1.641	0.980	77.131	60.677	36.236
3	2.175	1.732	0.941	80.422	64.041	34.794
4	2.065	1.700	0.903	76.354	62.858	33.369
5	2.115	1.516	0.954	78.203	56.055	35.275

	Mean	Stdev	%CV
Mean	2.101	1.671	61.793
Stdev	0.047	0.098	3.637
%CV	2.222	3.652	5.886

Mean -(neg) +(pos) 2SD  
Acceptable? PASS


	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	2.284	1.806	0.910	81.310	64.293	32.396
2	2.283	1.920	0.911	81.274	68.352	32.431
3	2.315	1.786	0.850	82.414	63.581	30.260
4	2.206	1.814	0.877	78.533	64.578	31.221
5	2.173	1.664	0.967	77.536	59.238	34.425

	Mean	Stdev	%CV
Mean	2.252	1.798	64.008
Stdev	0.060	0.091	3.249
%CV	2.656	5.075	5.075


Mean -(neg) +(pos) 2SD  
Acceptable? PASS

	Mean	Stdev	%CV
Mean	2.178	1.718	62.226
Stdev	0.090	0.097	3.120
%CV	4.156	5.662	5.013


Mean Callibrator and Control Absorbances Over Three Days



Mean Callibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days



Analyte: Barbiturates									
Instrument: Tecan-2 Analyst: A. Gooden & V. Coronado					Study Dates: 8/10/22 to 8/18/22 Matrix: Blood				
Target Analyte: Secobarbital	Sample Volume:	20 µL							
Cal Conc: 150 ng/ml	Calibrator (Abs)	0.651	Calibrator (Abs)	0.380	Negative % Binding	Calibrator % Binding	Positive % Binding		
Neg QC Conc: 75 ng/ml	0.639	44.266	36.749	48.401	44.551	36.071	22.749		
Pos QC Conc: 450 ng/ml	0.701	45.837	39.571	48.890	45.837	38.837	23.822		
	0.889	48.998	38.724	48.998	48.998	38.724	21.394		
	0.668	46.796	37.990	46.796	46.796	37.990	22.377		
Mean	0.673	0.396	0.188	1.498	1.007	0.400	0.200		
Stdev	0.027	0.018	0.018	4.501	4.501	0.200	0.200		
%CV	4.172	4.501	4.172	0.432	0.432	0.432	0.432		
Mean -(neg) +(pos) ZSD	0.760	0.432	0.432	PASS	PASS	PASS	PASS		
Acceptable?									
EIA_20220810B_VC									
1	0.861	0.432	0.432	47.178	37.534	23.671	23.671		
2	0.894	0.393	0.393	48.986	37.205	21.534	21.534		
3	0.727	0.400	0.400	39.836	34.849	21.918	21.918		
4	0.804	0.367	0.367	44.055	40.822	20.110	20.110		
5	0.774	0.362	0.362	42.411	37.370	19.836	19.836		
Mean	0.812	0.391	0.391	44.493	37.556	21.414	21.414		
Stdev	0.067	0.028	0.028	3.661	2.130	1.545	1.545		
%CV	8.229	7.218	7.218	8.228	5.672	7.217	7.217		
Mean -(neg) +(pos) ZSD	0.678	0.447	0.447	37.171	24.505	24.505	24.505		
Acceptable?									
EIA_20220815_VC									
1	0.941	0.451	0.451	47.525	37.374	22.778	22.778		
2	0.949	0.411	0.411	47.929	39.949	20.758	20.758		
3	0.867	0.470	0.470	43.778	36.768	23.737	23.737		
4	0.925	0.462	0.462	46.717	39.192	23.333	23.333		
5	0.897	0.401	0.401	45.303	37.525	20.253	20.253		
Mean	0.916	0.439	0.439	46.250	38.162	22.172	22.172		
Stdev	0.034	0.031	0.031	1.708	1.344	1.569	1.569		
%CV	3.686	7.078	7.078	3.693	3.522	7.076	7.076		
Mean -(neg) +(pos) ZSD	0.848	0.501	0.501	42.834	25.31	25.31	25.31		
Acceptable?									
Between-Run									
Mean	0.852	0.409	0.409	45.847	37.903	21.987	21.987		
Stdev	0.064	0.033	0.033	2.605	1.589	1.364	1.364		
%CV	7.565	8.079	8.079	5.682	4.191	6.202	6.202		



For EIA\_20220811B\_ASG, the mean absorbance (Abs) and %binding of the Negative QC minus two standard deviations were less than those of the mean calibrator. However, the mean values of the calibrator and controls of EIA\_20220811B\_ASG and between-run mean values met the requirement of average Abs and %binding of Neg>Cal>Pos. While in-run separation between calibrator and controls was monitored, this parameter is not among the criteria for acceptance or rejection of the immunoassay verification.

**Analyte: Benzodiazepines**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

**Analyte: Benzodiazepines**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Target Analyte: Oxazepam  
Cal Conc: 20 ng/mL Sample Volume: 30 µL  
Neg QC Conc: 10 ng/mL  
Pos QC Conc: 60 ng/mL

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	0.909	0.714	0.437	51.139	40.169	24.585
2	0.912	0.684	0.470	51.308	38.481	26.442
3	0.888	0.699	0.365	49.958	39.325	20.534
4	0.874	0.650	0.484	49.170	36.568	27.229
5	0.864	0.636	0.531	48.608	35.781	29.873

	Mean	Stdev	%CV
Mean	0.889	0.077	50.037
Stdev	0.021	0.033	1.186
%CV	2.371	13.494	3.472
Mean -(neg) +(pos) 2SD	0.847	4.852	4.852
Acceptable?	PASS	PASS	PASS

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	0.895	0.699	0.420	53.068	41.447	24.904
2	0.767	0.724	0.457	45.479	42.929	27.098
3	0.889	0.771	0.430	52.713	45.716	25.497
4	0.897	0.654	0.510	53.187	38.779	30.240
5	0.863	0.626	0.537	57.101	37.118	31.841


	Mean	Stdev	%CV
Mean	0.882	0.095	52.310
Stdev	0.071	0.057	4.217
%CV	8.061	10.807	8.061
Mean -(neg) +(pos) 2SD	0.740	8.229	8.229
Acceptable?	PASS	PASS	PASS

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	0.908	0.667	0.432	53.887	39.585	25.638
2	0.840	0.654	0.424	49.852	38.813	25.163
3	1.006	0.721	0.386	59.703	42.789	22.908
4	1.052	0.647	0.413	62.433	38.398	24.510
5	0.847	0.586	0.596	50.267	34.777	25.784

	Mean	Stdev	%CV
Mean	0.931	0.055	38.872
Stdev	0.095	0.048	2.866
%CV	10.213	18.511	10.213
Mean -(neg) +(pos) 2SD	0.741	7.372	7.372
Acceptable?	PASS	PASS	PASS


	Mean	Stdev	%CV
Mean	0.901	0.068	26.150
Stdev	0.047	0.062	2.874
%CV	7.562	13.575	7.401

Concentration	Day 1	Day 2	Day 3	Grand Mean
10	0.889	0.882	0.931	0.901
20	0.677	0.695	0.655	0.675
60	0.457	0.471	0.450	0.459
Slope	-0.008	-0.007	-0.008	-0.008



**Mean Callibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days**

Day	Negative	Callibrator	Positive
1	0.847	0.611	0.334
1	0.932	0.742	0.581
2	0.740	0.580	0.369
2	1.024	0.809	0.573



**Analyte: Buprenorphine**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

**Target Analyte: Buprenorphine**

Cal Conc:	1	ng/mL	Sample Volume:	50 µL
Neg QC Conc:	0.5	ng/mL		
Pos QC Conc:	3	ng/mL		

	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	1.764	1.198	0.623	61.721	41.917	21.798
2	1.855	1.188	0.726	62.596	41.568	25.297
3	1.999	1.366	0.700	69.944	47.796	24.493
4	1.852	1.327	0.786	64.801	46.431	27.502
5	1.872	1.263	0.800	65.500	44.892	27.992

Mean	1.855	1.272	0.726	64.912	44.521	25.416
Stdev	0.092	0.078	0.071	3.210	2.739	2.497
%CV	4.945	6.152	9.824	4.946	6.153	9.825
Mean -(neg) +(pos) 2SD	1.672		0.869	58.492		30.411
Acceptable?	PASS		PASS	PASS		PASS

1	1.612	1.277	0.635	56.861	45.044	22.399
2	1.602	1.250	0.730	56.508	44.092	25.750
3	1.835	1.239	0.707	64.727	43.704	24.938
4	1.957	1.339	0.747	69.030	47.231	26.349
5	1.868	1.352	0.678	65.891	47.690	23.915

Mean	1.775	1.291	0.699	62.803	45.552	24.870
Stdev	0.160	0.051	0.044	5.629	1.816	1.564
%CV	8.992	3.987	6.340	8.992	3.987	6.340
Mean -(neg) +(pos) 2SD	1.456		0.788	51.345		27.798
Acceptable?	PASS		PASS	PASS		PASS

1	1.749	1.199	0.634	62.722	42.988	22.736
2	1.664	1.117	0.658	59.674	40.057	23.597
3	1.871	1.314	0.550	67.097	47.122	19.724
4	1.743	1.242	0.679	62.507	44.540	24.350
5	1.810	1.226	0.788	64.909	43.966	28.259

Mean	1.767	1.220	0.662	63.382	43.737	23.733
Stdev	0.078	0.071	0.086	2.788	2.561	3.080
%CV	4.398	5.856	12.978	4.398	5.857	12.978
Mean -(neg) +(pos) 2SD	1.612		0.834	57.807		29.89
Acceptable?	PASS		PASS	PASS		PASS

Between-Run						
Mean	1.799	1.261	0.696	63.633	44.603	24.607
Stdev	0.114	0.070	0.070	3.899	2.356	2.387
%CV	6.362	5.581	10.035	6.128	5.283	9.702


**Analyte: Buprenorphine**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood


**Mean Calibrator and Control Absorbances Over Three Days**

Concentration	Day 1	Day 2	Day 3	Grand Mean
0.5	1.855	1.775	1.767	1.799
1	1.272	1.291	1.220	1.261
3	0.726	0.699	0.662	0.696
Slope	-0.401	-0.392	-0.396	-0.396



**Mean Calibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days**

Day	Negative	Calibrator	Positive
1	1.672	1.116	0.984
1	2.039	1.429	0.869
2	1.456	1.188	0.611
2	2.094	1.394	0.788
3	1.612	1.077	0.490
3	1.923	1.362	0.834



**Analyte: Cocaine and Metabolites**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Target Analyte: Benzoylscgonine  
Cal Conc: 50 ng/mL Sample Volume: 75 µL  
Neg QC Conc: 25 ng/mL  
Pos QC Conc: 150 ng/mL

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	1.371	1.127	0.715	51.824	42.601	27.027
2	1.337	1.207	0.740	50.539	45.625	27.972
3	1.327	1.157	0.718	50.161	43.735	27.140
4	1.361	1.103	0.827	51.446	41.693	31.261
5	1.376	1.118	0.761	52.013	42.260	28.766
Mean	1.354	1.142	0.752	51.197	43.183	28.433
Stdev	0.021	0.041	0.046	0.810	1.556	1.730
%CV	1.583	3.602	6.083	1.583	3.602	6.084
Mean -(neg) +(pos) 2SD	1.312	0.844	0.844	49.576	31.893	31.893
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	1.728	1.379	0.718	61.440	49.031	25.529
2	1.756	1.415	0.841	62.436	50.311	29.902
3	1.627	1.282	0.896	57.849	45.582	31.858
4	1.556	1.258	0.831	55.324	44.729	29.547
5	1.577	1.284	1.122	56.071	45.653	39.893
Mean	1.649	1.324	0.882	58.624	47.061	31.346
Stdev	0.089	0.069	0.149	3.181	2.452	5.303
%CV	5.425	5.210	16.917	5.426	5.210	16.917
Mean -(neg) +(pos) 2SD	1.470	1.180	1.180	52.262	41.951	41.951
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	1.461	1.178	0.582	52.545	42.366	20.931
2	1.535	1.186	0.781	55.206	42.654	28.088
3	1.310	1.198	0.633	47.114	43.086	22.766
4	1.333	1.109	0.813	47.941	39.885	29.239
5	1.335	1.135	0.917	48.013	40.820	31.763
Mean	1.395	1.161	0.745	50.164	41.762	26.557
Stdev	0.098	0.038	0.137	3.534	1.353	4.546
%CV	7.045	3.240	18.328	7.045	3.240	17.117
Mean -(neg) +(pos) 2SD	1.198	1.018	1.018	43.095	35.65	35.65
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS


Between-Run	Mean	Stdev	%CV
Mean	1.209	0.097	8.117
Stdev	0.153	0.097	6.333
%CV	10.428	7.992	76.600

**Analyte: Cocaine and Metabolites**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado


Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Concentration	Day 1	Day 2	Day 3	Grand Mean
25	1.354	1.649	1.395	1.466
50	1.142	1.324	1.161	1.209
150	0.752	0.882	0.745	0.793
Slope	-0.005	-0.006	-0.005	-0.005



**Mean Callibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days**

Day	Negative	Callibrator	Positive
1	1.312	1.060	0.661
1	1.397	1.225	0.844
2	1.470	1.186	0.583
2	1.828	1.462	1.180
3	1.198	1.086	0.472
3	1.591	1.236	1.018



**Analyte: Cannabinoids**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Target Analyte: 11-nor-9-carboxy-THC  
Cal Conc: 10 ng/mL  
Neg QC Conc: 5 ng/mL  
Pos QC Conc: 30 ng/mL  
Sample Volume: 50 µL

	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	1.837	1.341	0.777	61.397	44.820	25.969
2	1.866	1.430	0.814	62.366	47.794	27.206
3	2.037	1.510	0.671	68.082	50.468	22.426
4	2.032	1.470	0.605	67.914	49.131	20.221
5	1.989	1.549	0.760	66.477	51.771	25.401

	Mean	Stdev	%CV	Mean -(neg) +(pos) 2SD	Acceptable?
EIA_20220810B_VC	1.952	0.094	4.833	11.771	PASS
	0.094	0.085	3.154	2.673	2.854
	4.833	4.834	5.477	11.771	11.771
	1.763	0.896	58.940	29.952	29.952
	PASS	PASS	PASS	PASS	PASS

	Mean	Stdev	%CV	Mean -(neg) +(pos) 2SD	Acceptable?
EIA_20220811B_ASSG	2.005	0.774	65.140	50.260	25.146
	1.916	0.819	62.248	44.347	26.608
	2.203	0.728	71.572	47.986	23.652
	2.298	0.818	74.659	53.281	26.576
	2.222	0.799	72.190	54.743	25.958
	2.129	0.788	69.162	50.123	25.588
	0.161	0.038	5.224	4.161	1.235
	7.554	4.825	7.554	8.302	4.825
	1.807	0.864	58.713	28.057	28.057
	PASS	PASS	PASS	PASS	PASS

	Mean	Stdev	%CV	Mean -(neg) +(pos) 2SD	Acceptable?
EIA_20220815_VC	1.870	0.699	66.383	48.101	24.814
	1.874	0.703	66.525	49.982	24.956
	2.052	0.682	72.843	54.171	24.210
	2.003	0.685	71.104	53.319	24.317
	2.030	0.701	72.062	52.929	24.885
	1.966	0.694	69.783	51.700	24.636
	0.087	0.072	3.101	2.556	0.346
	4.445	4.945	4.444	4.945	1.405
	1.791	0.713	63.580	25.33	25.33
	PASS	PASS	PASS	PASS	PASS

	Mean	Stdev	%CV	Mean -(neg) +(pos) 2SD	Acceptable?
Between-Run	2.016	1.486	68.064	50.207	24.823
Mean	0.138	0.099	4.208	3.219	1.771
Stdev	6.841	6.627	6.183	6.412	7.136
%CV					


**Analyte: Cannabinoids**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood


Mean Callibrator and Control Absorbances Over Three Days

Concentration	Day 1	Day 2	Day 3	Grand Mean
5	1.952	2.129	1.966	2.016
10	1.460	1.543	1.456	1.486
30	0.725	0.788	0.694	0.736
Slope	-0.046	-0.049	-0.047	-0.047



Mean Callibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days


Day	Negative	Callibrator	Positive
1	1.763	1.300	0.555
1	2.141	1.620	0.896
2	1.807	1.287	0.712
2	2.450	1.799	0.864
3	1.791	1.312	0.675
3	2.141	1.600	0.713






Target Analyte: Carisoprodol  
 Sample Volume: 10 µL

Concentration	Day 1	Day 2	Day 3	Grand Mean
200	1.348	1.424	1.379	1.384
500	0.912	0.997	0.997	0.969
1500	0.541	0.650	0.587	0.592
Slope	-0.001	-0.001	-0.001	-0.001



Mean Calibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days

Day	Negative	Calibrator	Positive
1	1.190	0.720	0.468
1	1.506	1.103	0.613
2	1.130	0.830	0.436
2	1.717	1.164	0.863
3	1.311	0.880	0.549
3	1.447	1.114	0.624



Mean Calibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days

Day	Negative	Calibrator	Positive
1	1.190	0.720	0.468
1	1.506	1.103	0.613
2	1.130	0.830	0.436
2	1.717	1.164	0.863
3	1.311	0.880	0.549
3	1.447	1.114	0.624

Cal Conc:	500	200	1500
Neg QC Conc:	500	200	1500
Pos QC Conc:	500	200	1500

Concentration	Day 1	Day 2	Day 3	Grand Mean
200	1.348	1.424	1.379	1.384
500	0.912	0.997	0.997	0.969
1500	0.541	0.650	0.587	0.592

Concentration	Day 1	Day 2	Day 3	Grand Mean
200	1.348	1.424	1.379	1.384
500	0.912	0.997	0.997	0.969
1500	0.541	0.650	0.587	0.592

Concentration	Day 1	Day 2	Day 3	Grand Mean
200	1.348	1.424	1.379	1.384
500	0.912	0.997	0.997	0.969
1500	0.541	0.650	0.587	0.592

**Analyte: Fentanyl**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

**Analyte: Fentanyl**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Target Analyte: Fentanyl  
Cal Conc: 1 ng/mL  
Neg QC Conc: 0.5 ng/mL  
Pos QC Conc: 3 ng/mL  
Sample Volume: 75 µL


	Negative QC (Abs)	Callibrator (Abs)	Positive QC (Abs)	Negative % Binding	Callibrator % Binding	Positive % Binding
1	1.012	0.776	0.538	42.918	32.909	22.816
2	1.115	0.735	0.667	47.286	31.170	28.287
3	1.056	0.917	0.527	44.784	38.889	22.349
4	1.035	0.745	0.521	43.893	31.595	22.095
5	1.095	0.807	0.521	46.438	34.224	22.095
Mean	1.063	0.796	0.555	45.064	33.757	23.528
Stdev	0.042	0.073	0.063	1.794	3.108	2.676
%CV	3.982	9.207	11.374	3.982	9.207	11.375
Mean -(neg) +(pos) 2SD	0.978	0.681	0.681	41.475	28.881	20.040
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	1.080	0.844	0.581	43.689	34.142	23.503
1	1.080	0.844	0.581	43.689	34.142	23.503
2	1.084	0.763	0.560	43.851	31.675	22.654
3	1.107	0.805	0.541	44.782	32.565	21.885
4	1.103	0.742	0.271	44.620	30.016	10.963
5	1.085	0.792	0.524	43.892	32.039	21.197
Mean	1.092	0.793	0.495	44.167	32.087	20.040
Stdev	0.012	0.037	0.127	0.497	1.493	5.147
%CV	1.124	4.652	25.683	1.125	4.652	25.682
Mean -(neg) +(pos) 2SD	1.067	0.750	0.750	43.173	30.334	20.040
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS


	1.004	0.712	0.491	42.915	30.434	20.987
1	1.004	0.712	0.491	42.915	30.434	20.987
2	0.966	0.728	0.457	41.291	31.118	19.534
3	1.007	0.815	0.493	43.043	34.837	21.073
4	1.020	0.810	0.528	43.599	34.623	22.569
5	1.068	0.797	0.469	45.651	34.067	20.047
Mean	1.013	0.772	0.488	43.300	33.016	20.842
Stdev	0.037	0.049	0.027	1.571	2.078	1.162
%CV	3.627	6.294	5.574	3.627	6.294	5.575
Mean -(neg) +(pos) 2SD	0.940	0.542	0.542	40.159	23.17	16.391
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	1.056	0.787	0.513	44.177	32.960	21.470
1	1.056	0.787	0.513	44.177	32.960	21.470
2	0.046	0.052	0.083	1.500	2.394	3.519
3	4.313	6.622	16.251	3.396	7.264	16.391
Mean	1.056	0.787	0.513	44.177	32.960	21.470
Stdev	0.046	0.052	0.083	1.500	2.394	3.519
%CV	4.313	6.622	16.251	3.396	7.264	16.391

Mean Callibrator and Control Absorbances Over Three Days



Mean Callibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days



The % CV for the absolute absorbances (Abs) and % binding of Positive QCs from EIA\_20220811B\_ASG were greater than 20%. However, the between-run % CV of the Abs and % binding were acceptable. Section considers % CV >20% as large variability. However, this parameter is not among the criteria for acceptance or rejection of the immunoassay verification.

**Analyte: Methamphetamine**

Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Target Analyte: d-Methamphetamine  
Sample Volume: 25 µL

	Cal Conc:	ng/mL	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	20	1.886	1.886	1.298	0.833	74.639	51.183	32.847
2	10	1.610	1.610	1.297	0.865	63.486	51.144	34.109
3	60	1.293	1.293	1.524	0.992	60.095	50.986	39.117
4		1.517	1.517	1.257	0.506	59.819	49.566	19.953
5		1.926	1.926	1.208	0.944	75.946	47.634	37.224
Mean		1.693	1.693	1.271	0.828	66.797	50.103	32.650
Stdev		0.199	0.199	0.039	0.191	7.902	1.534	0.800
%CV		11.741	11.741	3.062	23.029	11.830	3.062	23.028
Mean -(neg) *(pos) 2SD		1.295	1.295	1.209	0.969	50.993	PASS	47.687
Acceptable?		PASS	PASS	PASS	PASS	PASS	PASS	% CV Too High
1		1.690	1.690	1.408	0.928	66.457	55.368	36.492
2		1.708	1.708	1.386	0.832	67.165	54.503	32.717
3		1.650	1.650	1.248	0.767	64.884	49.076	30.161
4		1.611	1.611	1.268	0.766	63.350	49.862	30.122
5		1.631	1.631	1.234	0.864	64.137	48.525	33.976
Mean		1.658	1.658	1.309	0.831	65.199	51.467	32.694
Stdev		0.040	0.040	0.082	0.069	1.589	3.216	2.697
%CV		2.436	2.436	6.249	8.250	2.436	6.250	8.250
Mean -(neg) *(pos) 2SD		1.577	1.577	0.969	0.969	62.022	PASS	38.088
Acceptable?		PASS	PASS	PASS	PASS	PASS	PASS	PASS
1		1.537	1.537	1.281	0.841	61.616	51.353	33.714
2		1.574	1.574	1.276	0.807	63.099	51.153	32.351
3		1.489	1.489	1.218	0.815	59.691	48.827	32.672
4		1.537	1.537	1.214	0.631	61.616	48.667	25.296
5		1.573	1.573	1.212	0.865	63.059	48.587	34.676
Mean		1.542	1.542	1.240	0.792	61.816	49.717	31.742
Stdev		0.035	0.035	0.035	0.093	1.395	1.406	3.718
%CV		2.257	2.257	2.828	11.713	2.257	2.828	11.712
Mean -(neg) *(pos) 2SD		1.472	1.472	0.977	0.977	59.026	PASS	39.18
Acceptable?		PASS	PASS	PASS	PASS	PASS	PASS	PASS
Between-Run		1.631	1.631	1.273	0.817	64.604	50.429	32.362
Mean		0.129	0.129	0.059	0.121	4.872	2.190	4.731
Stdev		7.886	7.886	4.672	14.755	7.542	4.343	14.620
%CV								


The % CV for the absolute absorbances (Abs) and % binding of Positive QCs from EIA\_20220810B\_VC were greater than 20%. However, the between-run % CV of the Abs and % binding were acceptable. Section considers % CV >20% as large variability. However, this parameter is not among the criteria for acceptance or rejection of the immunoassay verification.

**Analyte: Methamphetamine**


Instrument: Tecan-2  
Analyst: A. Gooden & V. Coronado

Study Dates: 8/10/22 to 8/18/22  
Matrix: Blood

Mean Callibrator and Control Absorbances Over Three Days



Mean Callibrator and Control Absorbances (+/- 2 Standard Deviations) Over Three Days



**Analyte: Opiates**  
 Instrument: Tecan-2  
 Analyst: A. Gooden & V. Coronado  
 Study Dates: 8/10/22 to 8/18/22  
 Matrix: Blood

**Target Analyte:** Morphine  
**Cal Conc:** 10 ng/mL  
**Neg QC Conc:** 5 ng/mL  
**Pos QC Conc:** 30 ng/mL

**Sample Volume:** 10 µL

	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	1.534	1.320	0.818	64.863	55.814	34.588
2	1.598	1.270	0.878	67.569	53.700	37.125
3	1.533	1.365	0.814	64.820	54.419	34.419
4	1.661	1.191	0.762	70.877	50.359	32.220
5	1.752	1.212	0.790	74.080	51.247	33.404
Mean	1.616	1.272	0.812	68.442	54.351	34.351
Stdev	0.093	0.073	0.043	4.013	3.073	1.816
%CV	5.745	5.287	5.287	5.863	5.287	5.287
Mean (-neg) + (pos) ZSD	1.430	5.714	0.898	60.416	37.964	37.964
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	1.425	1.205	0.886	68.841	58.213	42.802
2	1.554	1.272	0.827	75.072	61.449	39.952
3	1.402	1.101	0.643	67.729	53.188	31.063
4	1.537	1.281	0.551	74.251	54.493	26.618
5	1.418	1.245	0.730	68.502	60.145	35.266
Mean	1.467	1.190	0.727	70.879	57.488	35.140
Stdev	0.072	0.074	0.135	3.488	3.561	6.541
%CV	4.922	6.194	18.613	4.922	6.194	18.613
Mean (-neg) + (pos) ZSD	1.323	6.194	0.998	63.902	48.222	48.222
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	1.446	1.284	0.824	63.379	56.279	36.117
2	1.452	1.094	0.744	63.642	47.513	32.610
3	1.542	1.303	0.567	67.887	57.112	24.852
4	1.494	1.176	0.815	65.483	51.545	35.722
5	1.447	1.227	0.654	63.423	53.780	28.665
6	1.206	1.215	0.721	52.860	53.246	31.583
Mean	1.431	1.215	0.721	62.729	53.246	31.583
Stdev	0.116	0.089	0.110	5.105	3.879	4.811
%CV	8.139	7.285	15.226	8.139	0.729	15.226
Mean (-neg) + (pos) ZSD	1.198	7.285	0.640	52.918	41.21	41.21
Acceptable?	Low control range fails below cutoff	Low control range fails below cutoff	PASS	Low control range fails below cutoff	PASS	PASS

	Mean	Stdev	%CV
Mean	1.520	1.226	0.754
Stdev	0.097	0.081	0.105
%CV	6.378	6.601	13.963

For EIA\_20220815\_VC, the mean absorbance (Abs) and %binding of the Negative QC minus two standard deviations were less than those of the mean calibrator. However, the mean values of the calibrator and controls of EIA\_20220815\_VC and between-run mean values met the requirement of average Abs and %binding of Neg-Cal-P-Pos. While in-run separation between calibrator and controls was monitored, this parameter is not among the criteria for acceptance or rejection of the immunoassay verification.

**Analyte: Opiates**  
 Instrument: Tecan-2  
 Analyst: A. Gooden & V. Coronado  
 Study Dates: 8/10/22 to 8/18/22  
 Matrix: Blood

**Target Analyte:** Morphine  
**Cal Conc:** 10 ng/mL  
**Neg QC Conc:** 5 ng/mL  
**Pos QC Conc:** 30 ng/mL

**Sample Volume:** 10 µL

	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	1.534	1.320	0.818	64.863	55.814	34.588
2	1.598	1.270	0.878	67.569	53.700	37.125
3	1.533	1.365	0.814	64.820	54.419	34.419
4	1.661	1.191	0.762	70.877	50.359	32.220
5	1.752	1.212	0.790	74.080	51.247	33.404
Mean	1.616	1.272	0.812	68.442	54.351	34.351
Stdev	0.093	0.073	0.043	4.013	3.073	1.816
%CV	5.745	5.287	5.287	5.863	5.287	5.287
Mean (-neg) + (pos) ZSD	1.430	5.714	0.898	60.416	37.964	37.964
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	1.425	1.205	0.886	68.841	58.213	42.802
2	1.554	1.272	0.827	75.072	61.449	39.952
3	1.402	1.101	0.643	67.729	53.188	31.063
4	1.537	1.281	0.551	74.251	54.493	26.618
5	1.418	1.245	0.730	68.502	60.145	35.266
Mean	1.467	1.190	0.727	70.879	57.488	35.140
Stdev	0.072	0.074	0.135	3.488	3.561	6.541
%CV	4.922	6.194	18.613	4.922	6.194	18.613
Mean (-neg) + (pos) ZSD	1.323	6.194	0.998	63.902	48.222	48.222
Acceptable?	PASS	PASS	PASS	PASS	PASS	PASS

	Negative QC (Abs)	Calibrator (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
1	1.446	1.284	0.824	63.379	56.279	36.117
2	1.452	1.094	0.744	63.642	47.513	32.610
3	1.542	1.303	0.567	67.887	57.112	24.852
4	1.494	1.176	0.815	65.483	51.545	35.722
5	1.447	1.227	0.654	63.423	53.780	28.665
6	1.206	1.215	0.721	52.860	53.246	31.583
Mean	1.431	1.215	0.721	62.729	53.246	31.583
Stdev	0.116	0.089	0.110	5.105	3.879	4.811
%CV	8.139	7.285	15.226	8.139	0.729	15.226
Mean (-neg) + (pos) ZSD	1.198	7.285	0.640	52.918	41.21	41.21
Acceptable?	Low control range fails below cutoff	Low control range fails below cutoff	PASS	Low control range fails below cutoff	PASS	PASS


  

	Mean	Stdev	%CV
Mean	1.520	1.226	0.754
Stdev	0.097	0.081	0.105
%CV	6.378	6.601	13.963

For EIA\_20220815\_VC, the mean absorbance (Abs) and %binding of the Negative QC minus two standard deviations were less than those of the mean calibrator. However, the mean values of the calibrator and controls of EIA\_20220815\_VC and between-run mean values met the requirement of average Abs and %binding of Neg-Cal-P-Pos. While in-run separation between calibrator and controls was monitored, this parameter is not among the criteria for acceptance or rejection of the immunoassay verification.



Analyte: Phencyclidine		Study Dates: 8/10/22 to 8/18/22		Matrix: Blood	
Instrument: Tecan-2		Analyst: A. Gooden & V. Coronado		Study Dates: 8/10/22 to 8/18/22	
Instrument: Tecan-2		Analyst: A. Gooden & V. Coronado		Matrix: Blood	
Target Analyte: Phencyclidine		Sample Volume:	10 µL		
Cal Conc:	10 ng/mL				
Neg QC Conc:	5 ng/mL				
Pos QC Conc:	30 ng/mL				
		Negative QC (Abs)	Positive QC (Abs)	Calibrator (Abs)	Sample Volume
1	1.293	1.088	0.637	44.075	10 µL
2	1.434	0.983	0.621	39.822	
3	1.407	0.970	0.638	39.295	
4	1.388	1.233	0.641	49.949	
5	1.308	1.071	0.599	43.387	
Mean	1.368	1.069	0.627	43.306	
Stdev	0.063	0.105	0.018	4.271	
%CV	4.590	9.862	2.804	9.862	
Mean (-neg) + (pos) ZSD	1.240	9.862	0.662	26.853	
Acceptable?	PASS	PASS	PASS	PASS	
		Negative % Binding	Calibrator % Binding	Positive % Binding	
1	1.401	58.546	47.179	25.805	
2	1.502	62.766	48.517	25.157	
3	1.405	58.713	33.055	25.846	
4	1.417	59.214	46.302	25.967	
5	1.467	61.304	44.254	24.266	
Mean	1.438	60.109	43.861	25.408	
Stdev	0.044	1.849	6.236	0.712	
%CV	3.077	3.077	14.218	2.803	
Mean (-neg) + (pos) ZSD	1.350	56.410	14.218	26.853	
Acceptable?	PASS	PASS	PASS	PASS	
		Calibrator	Negative	Positive	
1	1.311	1.016	1.240	1.240	
2	1.324	1.090	0.858	0.858	
3	1.239	1.010	1.491	1.491	
4	1.073	0.991	1.350	1.350	
5	1.351	0.972	1.527	1.527	
Mean	1.260	1.036	1.348	1.348	
Stdev	0.112	0.053	0.930	0.930	
%CV	8.910	5.089	1.141	1.141	
Mean (-neg) + (pos) ZSD	1.035	5.089	1.141	1.141	
Acceptable?	PASS	PASS	PASS	PASS	
		Day	Negative	Positive	
Between-Run		1	1.240	0.592	
Mean	1.355	2	1.491	0.662	
Stdev	0.105	3	1.350	0.526	
%CV	7.763		1.527	0.798	

Analyte: Phencyclidine		Study Dates: 8/10/22 to 8/18/22		Matrix: Blood	
Instrument: Tecan-2		Analyst: A. Gooden & V. Coronado		Study Dates: 8/10/22 to 8/18/22	
Instrument: Tecan-2		Analyst: A. Gooden & V. Coronado		Matrix: Blood	
Target Analyte: Phencyclidine		Sample Volume:	10 µL		
Cal Conc:	10 ng/mL				
Neg QC Conc:	5 ng/mL				
Pos QC Conc:	30 ng/mL				
		Negative QC (Abs)	Positive QC (Abs)	Calibrator (Abs)	Sample Volume
1	1.293	1.088	0.637	44.075	10 µL
2	1.434	0.983	0.621	39.822	
3	1.407	0.970	0.638	39.295	
4	1.388	1.233	0.641	49.949	
5	1.308	1.071	0.599	43.387	
Mean	1.368	1.069	0.627	43.306	
Stdev	0.063	0.105	0.018	4.271	
%CV	4.590	9.862	2.804	9.862	
Mean (-neg) + (pos) ZSD	1.240	9.862	0.662	26.853	
Acceptable?	PASS	PASS	PASS	PASS	
		Negative % Binding	Calibrator % Binding	Positive % Binding	
1	1.401	58.546	47.179	25.805	
2	1.502	62.766	48.517	25.157	
3	1.405	58.713	33.055	25.846	
4	1.417	59.214	46.302	25.967	
5	1.467	61.304	44.254	24.266	
Mean	1.438	60.109	43.861	25.408	
Stdev	0.044	1.849	6.236	0.712	
%CV	3.077	3.077	14.218	2.803	
Mean (-neg) + (pos) ZSD	1.350	56.410	14.218	26.853	
Acceptable?	PASS	PASS	PASS	PASS	
		Calibrator	Negative	Positive	
1	1.311	1.016	1.240	1.240	
2	1.324	1.090	0.858	0.858	
3	1.239	1.010	1.491	1.491	
4	1.073	0.991	1.350	1.350	
5	1.351	0.972	1.527	1.527	
Mean	1.260	1.036	1.348	1.348	
Stdev	0.112	0.053	0.930	0.930	
%CV	8.910	5.089	1.141	1.141	
Mean (-neg) + (pos) ZSD	1.035	5.089	1.141	1.141	
Acceptable?	PASS	PASS	PASS	PASS	
		Day	Negative	Positive	
Between-Run		1	1.240	0.592	
Mean	1.355	2	1.491	0.662	
Stdev	0.105	3	1.350	0.526	
%CV	7.763		1.527	0.798	



For EIA\_20220815\_VC, the mean absorbance (Abs) and %binding of the Negative QC minus two standard deviations were less than those of the mean calibrator. However, the mean values of the calibrator and controls of EIA\_20220815\_VC and between-run mean values met the requirement of average Abs and %binding of Neg-Cal-Pos. While in-run separation between calibrator and controls was monitored, this parameter is not among the criteria for acceptance or rejection of the immunoassay verification.

Target Analyte:	Oxycodone	ng/mL	Sample Volume:	10 µL	Negative QC (Abs)	Positive QC (Abs)	Negative % Binding	Calibrator % Binding	Positive % Binding
Cal Conc:	10	0.848	0.437	0.279	30.848	10.149	15.897	10.149	10.149
Neg QC Conc:	5	0.369	0.252	0.648	23.572	9.167	13.423	9.167	9.167
Pos QC Conc:	30	0.980	0.468	0.265	35.649	9.640	17.024	9.640	9.640
		0.935	0.459	0.310	34.012	11.277	16.697	11.277	11.277
		1.015	0.444	0.285	36.923	10.731	16.151	10.731	10.731
Mean		0.885	0.435	0.280	32.201	10.193	15.838	10.193	10.193
Stdev		0.147	0.039	0.023	5.333	0.840	1.421	0.840	0.840
%CV		16.561	8.245	8.245	16.561	8.245	8.973	8.245	8.245
Mean -(neg) +(pos) ZSD		0.592	0.326	0.326	21.535	11.874	PASS	11.874	PASS
Acceptable?		PASS	PASS	PASS	PASS	PASS	PASS	PASS	PASS
		0.819	0.407	0.265	31.536	10.204	15.672	10.204	10.204
		0.908	0.397	0.317	34.963	12.206	15.287	12.206	12.206
		0.677	0.420	0.260	26.069	10.012	16.173	10.012	10.012
		0.786	0.327	0.282	30.266	10.859	12.591	10.859	10.859
		1.004	0.409	0.283	38.660	10.897	15.749	10.897	10.897
Mean		0.839	0.392	0.281	32.299	10.836	15.094	10.836	10.836
Stdev		0.124	0.037	0.022	4.772	0.860	1.434	0.860	0.860
%CV		14.775	9.501	7.941	14.774	7.938	9.503	7.938	7.938
Mean -(neg) +(pos) ZSD		0.591	0.326	0.326	22.755	12.556	PASS	12.556	PASS
Acceptable?		PASS	PASS	PASS	PASS	PASS	PASS	PASS	PASS
		0.529	0.313	0.223	21.118	8.902	12.495	8.902	8.902
		0.331	0.331	0.228	21.118	9.102	13.214	9.102	9.102
		0.555	0.349	0.211	22.156	8.423	13.932	8.423	8.423
		0.635	0.282	0.250	25.349	9.980	11.257	9.980	9.980
		0.513	0.342	0.257	20.479	10.259	13.653	10.259	10.259
Mean		0.552	0.323	0.234	22.044	9.333	12.910	9.333	9.333
Stdev		0.049	0.027	0.019	1.943	0.765	1.072	0.765	0.765
%CV		8.815	8.299	8.202	8.814	8.200	8.301	8.200	8.200
Mean -(neg) +(pos) ZSD		0.455	0.272	0.272	18.158	10.86	PASS	10.86	PASS
Acceptable?		PASS	PASS	PASS	PASS	PASS	PASS	PASS	PASS
Between-Run		0.759	0.384	0.285	28.848	10.121	14.614	10.121	10.121
Mean		0.186	0.058	0.030	6.365	0.993	1.774	0.993	0.993
Stdev		24.460	15.012	11.480	22.064	9.813	12.140	9.813	9.813
%CV									



The between-run % CV for the absolute absorbances (Abs) and % binding of negative QCs were greater than 20%. The between-run % CV of the Abs and % binding were evaluated to determine the variability likely to be encountered across immunoassay kits during routine use. Immunoassay kits showing large between-run % CV may show a large variability in Abs or %binding from day to day. Section considers % CV >20% as large variability. However, this parameter for Abs and % binding is not among the criteria for acceptance or rejection of the immunoassay verification.

## SUMMARY OF VALIDATION PERFORMANCE

Analyte: **Multiple**  
Units: **ng/mL**  
Instrument: **Tecan-2**

Analyst: **A. Gooden & V. Coronado**  
Dates of Validation: **8/10/22 to 8/18/22**  
Matrix: **Blood**

The intent of this summary is to capture and document important information about the performance of this method outside the required measurements for validation.

**Failed Runs (include dates/reasons):**

N/A

**Deviations from SOP:**

N/A

**Limitations of the Method:**

The mean absorbance (Abs) and %binding of the Negative QC minus two standard deviations were less than those of the mean calibrator of the phencyclidine, opiates, and barbiturates assays. However, the mean values of the calibrator and controls of EIA\_20220811B\_ASG and between-run mean values met the requirement of average Abs and %binding of Neg>Cal>Pos. While in-run separation between calibrator and controls was monitored, this parameter is not among the criteria for acceptance or rejection of the immunoassay verification. The between-run % CV for the absolute absorbances (Abs) and % binding of negative QCs were greater than 20% for the oxycodone assay. The between-run % CV of the Abs and % binding were evaluated to determine the variability likely to be encountered across immunoassay kits during routine use. Immunoassay kits showing large between-run % CV may show a large variability in Abs or %binding from day to day. Section considers % CV >20% as large variability. However, this parameter for Abs is not among the criteria for acceptance or rejection of the immunoassay verification.

**Other Observations:**

Abbreviations:  
ELISA - Enzyme-Linked Immunosorbent Assay  
% Binding - Percent Binding  
% RSD - Percent Relative Standard Deviation  
Avg. Abs. - Average Absorbance  
Abs - Absolute Absorbance  
Cal - Calibrator  
Neg - Negative Control  
Pos - Positive Control  
% CV - Percent Coefficient of Variation  
QC - Quality Control  
Outlier - Data point outside the 1.5 interquartile range (IQR) of the replicate values (i.e., <Q1 - 1.5\*IQR or >Q3 + 1.5\*IQR).  
PBS: Phosphate Buffer Saline

The absorption and % binding of POS-4 for Fentanyl (EIA\_20220811B\_ASG) and POS-4 for Methamphetamine (EIA\_20220810B\_VC) are outliers but were included in data evaluation for this verification.

**Lot Numbers:**

**Calibrators:** 220809B-MXC & 220809B-C-10  
**Controls:** 220809B-MXN, 220809B-MXP, 220809B-Q-5 & 220809B-Q-30  
**Blank Matrices:** E47246, E46905 & E45536  
**PBS:** 220808-PBS7

**Recommended Maximum Run Length (# unknown samples):**

30

**Conclusion:**

Method is fit for purpose for analysis of blood samples using the amphetamines, barbiturates, benzodiazepines, buprenorphine, cocaine and metabolites, cannabinoids, carisoprodol, fentanyl, methamphetamine, opiates, phencyclidine, and oxycodone assays.